

American  
National  
Standard



ANSI/AAMI  
ST65:2008/  
(R)2018

Processing of reusable  
Surgical textiles for use in  
health care facilities

# Objectives and uses of AAMI standards and recommended practices

It is most important that the objectives and potential uses of an AAMI product standard or recommended practice are clearly understood. The objectives of AAMI's technical development program derive from AAMI's overall mission: the advancement of medical instrumentation. Essential to such advancement are (1) a continued increase in the safe and effective application of current technologies to patient care, and (2) the encouragement of new technologies. It is AAMI's view that standards and recommended practices can contribute significantly to the advancement of medical instrumentation, provided that they are drafted with attention to these objectives and provided that arbitrary and restrictive uses are avoided.

A voluntary *standard* for a *medical device* recommends to the manufacturer the information that should be provided with or on the product, basic safety and performance criteria that should be considered in qualifying the device for clinical use, and the measurement techniques that can be used to determine whether the device conforms with the safety and performance criteria and/or to compare the performance characteristics of different products. Some standards emphasize the information that should be provided with the device, including performance characteristics, instructions for use, warnings and precautions, and other data considered important in ensuring the safe and effective use of the device in the clinical environment. Recommending the disclosure of performance characteristics often necessitates the development of specialized test methods to facilitate uniformity in reporting; reaching consensus on these tests can represent a considerable part of committee work. When a drafting committee determines that clinical concerns warrant the establishment of *minimum* safety and performance criteria, referee tests must be provided and the reasons for establishing the criteria must be documented in the rationale.

A *recommended practice* provides guidelines for the use, care, and/or processing of a medical device or system. A recommended practice does not address device performance *per se*, but rather procedures and practices that will help ensure that a device is used safely and effectively and that its performance will be maintained.

Although a device standard is primarily directed to the manufacturer, it may also be of value to the potential purchaser or user of the device as a frame of reference for device evaluation. Similarly, even though a recommended practice is usually oriented towards healthcare professionals, it may be useful to the manufacturer in better understanding the environment in which a medical device will be used. Also, some recommended practices, while not addressing device performance criteria, provide guidelines to industrial personnel on such subjects as sterilization processing, methods of collecting data to establish safety and efficacy, human engineering, and other processing or evaluation techniques; such guidelines may be useful to health care professionals in understanding industrial practices.

In determining whether an AAMI standard or recommended practice is relevant to the specific needs of a potential user of the document, several important concepts must be recognized:

All AAMI standards and recommended practices are *voluntary* (unless, of course, they are adopted by government regulatory or procurement authorities). The application of a standard or recommended practice is solely within the discretion and professional judgment of the user of the document.

Each AAMI standard or recommended practice reflects the collective expertise of a committee of health care professionals and industrial representatives, whose work has been reviewed nationally (and sometimes internationally). As such, the consensus recommendations embodied in a standard or recommended practice are intended to respond to clinical needs and, ultimately, to help ensure patient safety. A standard or recommended practice is limited, however, in the sense that it responds generally to perceived risks and conditions that may not always be relevant to specific situations. A standard or recommended practice is an important *reference* in responsible decision-making, but it should never *replace* responsible decision-making.

Despite periodic review and revision (at least once every five years), a standard or recommended practice is necessarily a static document applied to a dynamic technology. Therefore, a standards user must carefully review the reasons why the document was initially developed and the specific rationale for each of its provisions. This review will reveal whether the document remains relevant to the specific needs of the user.

Particular care should be taken in applying a product standard to existing devices and equipment, and in applying a recommended practice to current procedures and practices. While observed or potential risks with existing equipment typically form the basis for the safety and performance criteria defined in a standard, professional judgment must be used in applying these criteria to existing equipment. No single source of information will serve to identify a particular product as "unsafe". A voluntary standard can be used as one resource, but the ultimate decision as to product safety and efficacy must take into account the specifics of its utilization and, of course, cost-benefit considerations. Similarly, a recommended practice should be analyzed in the context of the specific needs and resources of the individual institution or firm. Again, the rationale accompanying each AAMI standard and recommended practice is an excellent guide to the reasoning and data underlying its provision.

In summary, a standard or recommended practice is truly useful only when it is used in conjunction with other sources of information and policy guidance and in the context of professional experience and judgment.

## INTERPRETATIONS OF AAMI STANDARDS AND RECOMMENDED PRACTICES

Requests for interpretations of AAMI standards and recommended practices must be made in writing, to the AAMI Vice President, Standards Policy and Programs. An official interpretation must be approved by letter ballot of the originating committee and subsequently reviewed and approved by the AAMI Standards Board. The interpretation will become official and representation of the Association only upon exhaustion of any appeals and upon publication of notice of interpretation in the "Standards Monitor" section of the *AAMI News*. The Association for the Advancement of Medical Instrumentation disclaims responsibility for any characterization or explanation of a standard or recommended practice which has not been developed and communicated in accordance with this procedure and which is not published, by appropriate notice, as an *official interpretation* in the *AAMI News*.

# Processing of reusable surgical textiles for use in health care facilities

Developed by  
**AAMI**

Approved 4 December 2008 and reaffirmed 2 December 2013 and 6 September 2018 by  
**American National Standards Institute Inc.**

**Abstract:** This recommended practice provides guidelines for the proper handling, processing, and preparation of reusable surgical textiles either on-site or off-site for use in health care facilities. This recommended practice specifically addresses design criteria for functional work areas; staff qualifications, education, training, dress codes, and other personnel considerations; receiving and handling of soiled surgical textiles; laundry processing considerations; transport of both soiled and clean surgical textiles; installation, care, and maintenance of laundry equipment; quality control; and regulatory considerations. Definitions of terms and a bibliography are also provided.

**Keywords:** laundry, surgical drapes, surgical gowns, wrappers

## AAMI Recommended Practice

This Association for the Advancement of Medical Instrumentation (AAMI) recommended practice implies a consensus of those substantially concerned with its scope and provisions. The existence of an AAMI recommended practice does not in any respect preclude anyone, whether they have approved the recommended practice or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the recommended practice. AAMI recommended practices are subject to periodic review, and users are cautioned to obtain the latest editions.

**CAUTION NOTICE:** This AAMI recommended practice may be revised or withdrawn at any time. AAMI procedures require that action be taken to reaffirm, revise, or withdraw this recommended practice no later than five years from the date of publication. Interested parties may obtain current information on all AAMI documents by calling or writing AAMI.

All AAMI standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

*Published by*

AAMI  
4301 N. Fairfax Drive, Suite 301  
Arlington, VA 22203-1633  
[www.aami.org](http://www.aami.org)

© 2009 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, contact AAMI at 4301 N. Fairfax Drive, Suite 301, Arlington, VA 22203-1633. Phone: (703) 525-4890; Fax: (703) 525-1067.

Printed in the United States of America

**ISBN 978-1-57020-336-7**

# Contents

	Page
Glossary of equivalent standards .....	vii
Committee representation .....	ix
Foreword .....	xi
Introduction: Need for the recommended practice .....	xii
<b>1</b> Scope .....	<b>1</b>
<b>1.1</b> General .....	<b>1</b>
<b>1.2</b> Inclusions .....	<b>1</b>
<b>1.3</b> Exclusions .....	<b>1</b>
<b>2</b> Definitions, symbols, and abbreviations .....	<b>2</b>
<b>3</b> Design considerations .....	<b>5</b>
<b>3.1</b> General rationale .....	<b>5</b>
<b>3.2</b> Work area design and functional work flow .....	<b>6</b>
<b>3.2.1</b> Definitions of work areas .....	<b>6</b>
<b>3.2.2</b> Design criteria for work areas .....	<b>7</b>
<b>3.2.3</b> Functional work flow patterns .....	<b>8</b>
<b>3.2.4</b> Traffic control .....	<b>8</b>
<b>3.3</b> Physical facilities: laundry area .....	<b>8</b>
<b>3.3.1</b> Space requirements .....	<b>8</b>
<b>3.3.2</b> Mechanical systems .....	<b>8</b>
<b>3.3.3</b> Floors, walls, ceilings, and vents .....	<b>8</b>
<b>3.3.4</b> Ventilation .....	<b>9</b>
<b>3.3.5</b> Temperature and humidity control .....	<b>9</b>
<b>3.3.6</b> Lighting systems .....	<b>9</b>
<b>3.3.7</b> Handwashing facilities .....	<b>9</b>
<b>3.3.8</b> Emergency eyewash/shower equipment .....	<b>9</b>
<b>3.3.9</b> Soil-sort area .....	<b>10</b>
<b>3.3.10</b> Chemical storage area .....	<b>10</b>
<b>3.3.11</b> Storage area for clean textile packs .....	<b>10</b>
<b>3.3.12</b> Housekeeping .....	<b>10</b>
<b>3.4</b> Physical facilities: surgical pack assembly area .....	<b>11</b>
<b>3.4.1</b> Space requirements .....	<b>11</b>
<b>3.4.2</b> Mechanical systems .....	<b>11</b>
<b>3.4.3</b> Floors, walls, ceilings, and vents .....	<b>11</b>
<b>3.4.4</b> Ventilation .....	<b>11</b>
<b>3.4.5</b> Temperature and humidity control .....	<b>11</b>
<b>3.4.6</b> Lighting systems .....	<b>12</b>
<b>3.4.7</b> Hand hygiene facilities .....	<b>12</b>
<b>3.4.8</b> Storage area for clean textile packs .....	<b>12</b>
<b>3.4.9</b> Surgical pack sterilization area .....	<b>12</b>
<b>3.4.10</b> Sterile storage area .....	<b>12</b>
<b>3.4.11</b> Housekeeping .....	<b>12</b>
<b>4</b> Personnel considerations .....	<b>12</b>
<b>4.1</b> General rationale .....	<b>12</b>
<b>4.2</b> Qualifications .....	<b>13</b>
<b>4.2.1</b> Supervisors/managers .....	<b>13</b>
<b>4.2.2</b> Personnel .....	<b>13</b>
<b>4.3</b> Training and education .....	<b>13</b>
<b>4.4</b> Health and personal hygiene .....	<b>14</b>
<b>4.5</b> Attire .....	<b>15</b>
<b>4.5.1</b> General .....	<b>15</b>

4.5.2	Personal protective equipment (PPE) .....	15
5	Receiving and handling of reusable surgical textiles .....	15
5.1	General rationale .....	15
5.2	Newly purchased items .....	16
5.2.1	Identification and handling .....	16
5.2.2	Washing .....	16
5.3	Collecting and transporting soiled surgical textiles .....	16
5.3.1	Collecting soiled surgical textiles at the point of use .....	16
5.3.2	Transporting soiled surgical textiles .....	16
5.4	Sorting of soiled textiles .....	17
5.4.1	General considerations .....	17
5.4.2	Pre-sort systems .....	17
5.4.3	Post-sort systems .....	17
6	Laundry processing recommendations .....	18
6.1	General rationale .....	18
6.2	Washing .....	18
6.2.1	Procedures .....	18
6.2.2	Loading of washing equipment .....	18
6.2.3	Steps in the washing process: laundry formulas .....	18
6.3	Drying .....	20
6.3.1	Procedures .....	20
6.3.2	Equipment loading .....	20
6.3.3	Steps in the drying process: drying formulas .....	21
6.4	Process monitoring .....	21
6.4.1	Rationale for process monitoring .....	21
6.4.2	Process monitoring: supplies .....	21
6.4.3	Process monitoring: equipment operation .....	23
6.4.4	Process monitoring: finished products .....	23
7	Inspection, testing, and maintenance of laundered textiles .....	24
7.1	General rationale .....	24
7.2	Visual inspection .....	25
7.2.1	Quality standards .....	25
7.2.2	Stains .....	25
7.2.3	Physical defects .....	27
7.2.4	Chemical or thermal damage .....	27
7.2.5	Foreign debris .....	27
7.2.6	Labeling .....	28
7.2.7	Tracking system .....	28
7.3	Testing .....	28
7.3.1	Test procedures .....	28
7.3.2	Microbiological cleanliness .....	29
7.3.3	Effective life .....	29
7.3.4	Important functional attributes .....	29
7.4	Maintenance .....	31
7.4.1	Patching .....	31
7.4.2	Mending .....	31
7.4.3	Rewash .....	31
7.4.4	Rejuvenation .....	32
7.4.5	Retirement or alternate use .....	32
8	Preparation and packaging .....	32
8.1	General rationale .....	32
8.2	Procedures .....	32
8.3	Folding .....	32
8.3.1	General considerations .....	32
8.3.2	Guidelines for folding gowns .....	32
8.3.3	Guidelines for folding drapes .....	33
8.4	Pack assembly .....	34

8.5	Wrapping.....	34
8.6	Labeling/identification of packs.....	34
<b>9</b>	<b>Handling, transport, and storage of laundered textiles .....</b>	<b>35</b>
9.1	General rationale.....	35
9.2	Procedures.....	35
9.3	Personnel attire and hygiene.....	35
9.4	Handling clean/sterile textiles.....	35
9.5	Transport.....	35
9.5.1	General considerations .....	35
9.5.2	Method of transport.....	35
9.5.3	Separation of clean/sterile and soiled textiles .....	36
9.5.4	Laundry cart cleaning, disinfection, and loading.....	36
9.5.5	Truck cleaning and loading .....	36
9.6	Storage.....	37
9.6.1	Storage conditions .....	37
9.6.2	Storage shelving .....	37
9.6.3	Stock rotation .....	37
<b>10</b>	<b>Installation, operation, care, and maintenance of laundry equipment .....</b>	<b>37</b>
10.1	General rationale.....	37
10.2	Documentation .....	38
10.2.1	Identification.....	38
10.2.2	Safety.....	38
10.2.3	Manuals and installation/operating instructions.....	38
10.3	Installation .....	38
10.3.1	General considerations .....	38
10.3.2	Utilities .....	39
10.4	Operation .....	40
10.4.1	General considerations .....	40
10.4.2	Washing, extraction, and drying equipment .....	40
10.4.3	Support systems .....	40
10.5	Routine care and maintenance.....	42
10.5.1	Routine care.....	42
10.5.2	Scheduled (preventive) maintenance.....	42
10.5.3	Unscheduled maintenance (repairs) .....	42
10.5.4	Calibration.....	42
10.5.5	Recordkeeping.....	43
<b>11</b>	<b>Quality control.....</b>	<b>43</b>
11.1	General rationale.....	43
11.2	General quality control criteria.....	43
11.2.1	Functional performance criteria.....	43
11.2.2	Verification of laundry processes .....	44
11.3	Policies and procedures.....	44
11.4	Barrier efficacy .....	44
11.5	Tracking uses of reusable surgical textile products .....	45
11.6	Process performance .....	45
11.6.1	Quality assessment.....	45
11.6.2	Quality process .....	46
<b>12</b>	<b>Medical device regulatory considerations.....</b>	<b>47</b>
<b>Annexes</b>		
<b>A</b>	<b>Examples of folding procedures.....</b>	<b>48</b>
<b>B</b>	<b>Bibliography .....</b>	<b>51</b>
<b>Tables</b>		
<b>1</b>	<b>An example of inspection criteria for stains .....</b>	<b>26</b>

**Figures**

**A.1** Example of a folding procedure for gowns .....48

**A.2** Example of a folding procedure for standard laparotomy drapes .....49

**A.3** Example of a folding procedure for mayo stand covers .....50

**A.4** Example of a folding procedure for towels .....50

**A.5** Example of a folding procedure for sheets .....50

## Glossary of equivalent standards

International Standards adopted in the United States may include normative references to other International Standards. For each International Standard that has been adopted by AAMI (and ANSI), the table below gives the corresponding U.S. designation and level of equivalency to the International Standard. NOTE: Documents are sorted by international designation.

Other normatively referenced International Standards may be under consideration for U.S. adoption by AAMI; therefore, this list should not be considered exhaustive.

International designation	U.S. designation	Equivalency
IEC 60601-1:2005	ANSI/AAMI ES60601-1:2005	Major technical variations
IEC 60601-1-2:2007	ANSI/AAMI/IEC 60601-1-2:2007	Identical
IEC 60601-2-2:2006	ANSI/AAMI/IEC 60601-2-2:2006	Identical
IEC 60601-2-4:2002	ANSI/AAMI DF80:2003	Major technical variations
IEC 60601-2-19:1990 and A1:1996	ANSI/AAMI II36:2004	Major technical variations
IEC 60601-2-20:1990 and A1:1996	ANSI/AAMI II51:2004	Major technical variations
IEC 60601-2-21:1994 and Amendment 1:1996	ANSI/AAMI/IEC 60601-2-21 and Amendment 1:2000 (consolidated texts)	Identical
IEC 60601-2-24:1998	ANSI/AAMI ID26:2004	Major technical variations
IEC 60601-2-47:2001	ANSI/AAMI EC38:2007	Major technical variations
IEC 60601-2-50:2001	ANSI/AAMI/IEC 60601-2-50:2006	Identical
IEC 80601-2-58:2008	ANSI/AAMI/IEC 80601-2-58:2008	Identical
IEC/TR 60878:2003	ANSI/AAMI/IEC TIR60878:2003	Identical
IEC/TR 62296:2003	ANSI/AAMI/IEC TIR62296:2003	Identical
IEC 62304:2006	ANSI/AAMI/IEC 62304:2006	Identical
IEC/TR 62348:2006	ANSI/AAMI/IEC TIR62348:2006	Identical
ISO 5840:2005	ANSI/AAMI/ISO 5840:2005	Identical
ISO 7198:1998	ANSI/AAMI/ISO 7198:1998/2001/(R)2004	Identical
ISO 7199:1996	ANSI/AAMI/ISO 7199:1996/(R)2002	Identical
ISO 8637:2004	ANSI/AAMI RD16:2007	Major technical variations
ISO 8638:2004	ANSI/AAMI RD17:2007	Major technical variations
ISO 10993-1:2003	ANSI/AAMI/ISO 10993-1:2003	Identical
ISO 10993-2:2006	ANSI/AAMI/ISO 10993-2:2006	Identical
ISO 10993-3:2003	ANSI/AAMI/ISO 10993-3:2003	Identical
ISO 10993-4:2002 and A1:2006	ANSI/AAMI/ISO 10993-4:2002 and A1:2006	Identical
ISO 10993-5:1999	ANSI/AAMI/ISO 10993-5:1999	Identical
ISO 10993-6:2007	ANSI/AAMI/ISO 10993-6:2007	Identical
ISO 10993-7:2008	ANSI/AAMI/ISO 10993-7:2008	Identical
ISO 10993-9:1999	ANSI/AAMI/ISO 10993-9:1999/(R)2005	Identical
ISO 10993-10:2002 and Amendment 1:2006	ANSI/AAMI BE78:2002/(R)2008 ANSI/AAMI BE78:2002/A1:2006/(R)2008	Minor technical variations Identical
ISO 10993-11:2006	ANSI/AAMI/ISO 10993-11:2006	Identical
ISO 10993-12:2007	ANSI/AAMI/ISO 10993-12:2007	Identical
ISO 10993-13:1998	ANSI/AAMI/ISO 10993-13:1999/(R)2004	Identical
ISO 10993-14:2001	ANSI/AAMI/ISO 10993-14:2001/(R)2006	Identical
ISO 10993-15:2000	ANSI/AAMI/ISO 10993-15:2000/(R)2006	Identical
ISO 10993-16:1997	ANSI/AAMI/ISO 10993-16:1997/(R)2003	Identical
ISO 10993-17:2002	ANSI/AAMI/ISO 10993-17:2002/(R)2008	Identical
ISO 10993-18:2005	ANSI/AAMI BE83:2006	Major technical variations
ISO/TS 10993-19:2006	ANSI/AAMI/ISO TIR10993-19:2006	Identical
ISO/TS 10993-20:2006	ANSI/AAMI/ISO TIR10993-20:2006	Identical
ISO 11135-1:2007	ANSI/AAMI/ISO 11135-1:2007	Identical

<b>International designation</b>	<b>U.S. designation</b>	<b>Equivalency</b>
ISO/TS 11135-2:2008	ANSI/AAMI/ISO TIR11135-2:2008	Identical
ISO 11137-1:2006	ANSI/AAMI/ISO 11137-1:2006	Identical
ISO 11137-2:2006 (2006-08-01 corrected version)	ANSI/AAMI/ISO 11137-2:2006	Identical
ISO 11137-3:2006	ANSI/AAMI/ISO 11137-3:2006	Identical
ISO 11138-1: 2006	ANSI/AAMI/ISO 11138-1:2006	Identical
ISO 11138-2: 2006	ANSI/AAMI/ISO 11138-2:2006	Identical
ISO 11138-3: 2006	ANSI/AAMI/ISO 11138-3:2006	Identical
ISO 11138-4: 2006	ANSI/AAMI/ISO 11138-4:2006	Identical
ISO 11138-5: 2006	ANSI/AAMI/ISO 11138-5:2006	Identical
ISO/TS 11139:2006	ANSI/AAMI/ISO 11139:2006	Identical
ISO 11140-1:2005	ANSI/AAMI/ISO 11140-1:2005	Identical
ISO 11140-3:2007	ANSI/AAMI/ISO 11140-3:2007	Identical
ISO 11140-4:2007	ANSI/AAMI/ISO 11140-4:2007	Identical
ISO 11140-5:2007	ANSI/AAMI/ISO 11140-5:2007	Identical
ISO 11607-1:2006	ANSI/AAMI/ISO 11607-1:2006	Identical
ISO 11607-2:2006	ANSI/AAMI/ISO 11607-2:2006	Identical
ISO 11737-1: 2006	ANSI/AAMI/ISO 11737-1:2006	Identical
ISO 11737-2:1998	ANSI/AAMI/ISO 11737-2:1998	Identical
ISO 13408-1:2008	ANSI/AAMI/ISO 13408-1:2008	Identical
ISO 13408-2:2003	ANSI/AAMI/ISO 13408-2:2003	Identical
ISO 13408-3:2006	ANSI/AAMI/ISO 13408-3:2006	Identical
ISO 13408-4:2005	ANSI/AAMI/ISO 13408-4:2005	Identical
ISO 13408-5:2006	ANSI/AAMI/ISO 13408-5:2006	Identical
ISO 13408-6:2006	ANSI/AAMI/ISO 13408-6:2006	Identical
ISO 13485:2003	ANSI/AAMI/ISO 13485:2003	Identical
ISO 14155-1:2003	ANSI/AAMI/ISO 14155-1:2003/(R)2008	Identical
ISO 14155-2:2003	ANSI/AAMI/ISO 14155-2:2003/(R)2008	Identical
ISO 14160:1998	ANSI/AAMI/ISO 14160:1998/(R)2008	Identical
ISO 14161:2000	ANSI/AAMI/ISO 14161:2000	Identical
ISO 14937:2000	ANSI/AAMI/ISO 14937:2000	Identical
ISO/TR 14969:2004	ANSI/AAMI/ISO TIR14969:2004	Identical
ISO 14971:2007	ANSI/AAMI/ISO 14971:2007	Identical
ISO 15223-1:2007 and A1:2008	ANSI/AAMI/ISO 15223-1:2007 and A1:2008	Identical
ISO 15225:2000 and A1:2004	ANSI/AAMI/ISO 15225:2000/(R)2006 and A1:2004/(R)2006	Identical
ISO 15674:2001	ANSI/AAMI/ISO 15674:2001	Identical
ISO 15675:2001	ANSI/AAMI/ISO 15675:2001	Identical
ISO 15882:2008	ANSI/AAMI/ISO 15882:2008	Identical
ISO/TR 16142:2006	ANSI/AAMI/ISO TIR16142:2005	Identical
ISO 17664:2004	ANSI/AAMI ST81:2004	Major technical variations
ISO 17665-1:2006	ANSI/AAMI/ISO 17665-1:2006	Identical
ISO 18472:2006	ANSI/AAMI/ISO 18472:2006	Identical
ISO/TS 19218:2005	ANSI/AAMI/ISO 19218:2005	Identical
ISO 22442-1:2007	ANSI/AAMI/ISO 22442-1:2007	Identical
ISO 22442-2:2007	ANSI/AAMI/ISO 22442-2:2007	Identical
ISO 22442-3:2007	ANSI/AAMI/ISO 22442-3:2007	Identical
ISO 25539-1:2003 and A1:2005	ANSI/AAMI/ISO 25539-1:2003 and A1:2005	Identical
ISO 25539-2:2008	ANSI/AAMI/ISO 25539-2:2008	Identical
ISO 81060-1:2007	ANSI/AAMI/ISO 81060-1:2007	Identical

## Committee representation

### Association for the Advancement of Medical Instrumentation Reusable Surgical Textile Processing Working Group

This recommended practice was developed by the AAMI Reusable Surgical Textile Processing Working Group under the auspices of the AAMI Sterilization Standards Committee. Approval of the recommended practice does not necessarily mean that all working group members voted for its approval.

At the time this document was published, the **AAMI Reusable Surgical Textile Processing Working Group** had the following members:

<i>Cochairs:</i>	John Hamilton Linda A. Slone, RN, CNOR
<i>Members:</i>	Nola Bayes, MBA, Wentworth-Douglass Hospital, Dover, NH Nathan L. Belkin, PhD, Clearwater, FL Peter L. Brown, W.L. Gore & Associates Bradley J. Bushman, Standard Textile Co., Inc. Nancy Chobin, RN, CSPDM, St. Barnabas Health Care System, West Orange, NJ Linda Clement, CRCST, STERIS Corporation Ramona Conner, RN, MSN, CNOR, Association of PeriOperative Registered Nurses Betty D. Edge, North Shore University Hospital, Manhasset, NY Linda Fairbanks, Association for Linen Management John Hamilton, SRI Surgical Charles O. Hancock, H&W Technology LLC Jeffrey Hills, RM, Nelson Laboratories Inc. Nyla Skee Japp, RN, PhD, CSPDM, IMS, Phoenix, AZ David W. Johnson, Kimberly-Clark Corporation Susan G. Klacik, CCSMC, FCS, ACE, International Association of Healthcare Central Service Materiel Management Colleen Patricia Landers, RN, Canadian Standards Association Mark E. Lanning, Hill-Rom Company Sharon K. Lappalainen, U.S. Food and Drug Administration Teckla A. Maresca, LPN, CSPDM, St. Clare's Health System, Denville, NJ Joseph A. Palomo, Cardinal Health K. John Pournoor, PhD, 3M Health Care Frank Sizemore, Wake Forest University Baptist Medical Center, Winston-Salem, NC Linda A. Slone, RN, CNOR, Sibley Memorial Hospital, Washington, DC Donna Swenson, CSPDM, West Suburban Hospital Medical Center, Oak Park, IL Nora E. Wikander, RN, CSPDM, St. Joseph's Wayne Hospital, Wayne, NJ
<i>Alternates:</i>	Pamela Carter, RN, CNOR, STERIS Corporation Todd Hillam, MBA, Nelson Laboratories Inc. Natalie Lind, International Association of Healthcare Central Service Materiel Management Sheila A. Murphey, MD, U.S. Food and Drug Administration

---

NOTE—Participation by federal agency representatives in the development of this recommended practice does not constitute endorsement by the federal government or any of its agencies.

---

At the time this document was published, the **AAMI Sterilization Standards Committee** had the following members.

*Chairs:* Victoria M. Hitchins, PhD  
William E. Young

*Members:* Trabue D. Bryans, WuXi AppTec Inc.  
Peter A. Burke, PhD, STERIS Corporation  
Nancy Chobin, RN, CSPDM, Saint Barnabas Health Care System, West Orange, NJ  
Charles Cogdill, Boston Scientific Corporation  
Ramona Conner, RN, MSN, CNOR, Association of periOperative Registered Nurses  
Jacqueline Daley, Association for Professionals in Infection Control and Epidemiology  
Kimbrell Darnell, CR Bard  
Lisa Foster, Sterigenics International  
Joel R. Gorski, PhD, NAMSA  
Deborah A. Havlik, Hospira Worldwide Inc.  
Victoria M. Hitchins, PhD, U.S. Food and Drug Administration  
Danny Hutson, Cardinal Health  
Lois Atkinson Jones, MS, Cary, IN  
Susan G. Klacik, CCSMC, FCS, ACE, International Association of Healthcare Central Service Materiel Management  
Byron J. Lambert, PhD, Abbott Laboratories  
Colleen Patricia Landers, RN, Canadian Standards Association  
Lisa N. Macdonald, Becton Dickinson & Company  
Jeff Martin, Alcon Laboratories Inc.  
Patrick J. McCormick, PhD, Bausch & Lomb Inc.  
Rainer Newman, Johnson & Johnson  
Janet Prust, 3M Health Care  
Nancy J. Rakiewicz, Ethox International Inc.  
Michael H. Scholla, MS, PhD, DuPont Nonwovens  
Mark Seybold, Baxter Healthcare Corporation  
Andrew Sharavara, Propper Manufacturing Co. Inc.  
Mark N. Smith, Getinge USA  
William N. Thompson, Covidien  
James L. Whitby, MA, MB, FRCP, London, Ontario  
Martell Kress Winters, SM, Nelson Laboratories Inc.

*Alternates:* Lloyd Brown, Covidien  
Glenn W. Calvert, Becton Dickinson & Company  
David Dion, Cardinal Health  
Steven J. Elliott, WuXi AppTec Inc.  
Thomas J. Frazar, Johnson & Johnson  
Kathy Hoffman, Sterigenics International  
Jim Kaiser, Bausch & Lomb Inc.  
Joseph J. Lasich, Alcon Laboratories Inc.  
Chiu Lin, PhD, U.S. Food and Drug Administration  
Natalie Lind, International Association of Healthcare Central Service Materiel Management  
Ralph Makinen, Boston Scientific Corporation  
Mary S. Mayo, CR Bard  
David Ford McGoldrick, Abbott Laboratories  
Jerry R. Nelson, MS, PhD, Nelson Laboratories Inc.  
Karen Polkinghorne, DuPont Nonwovens  
Michael Sadowski, Baxter Healthcare Corporation  
John R. Scoville, Jr., STERIS Corporation  
Jason Voisinet, Ethox International Inc.  
Craig A. Wallace, 3M Health Care  
Valerie Welter, Hospira Worldwide Inc.  
William E. Young, Boston Scientific Corporation

---

NOTE—Participation by federal agency representatives in the development of this recommended practice does not constitute endorsement by the federal government or any of its agencies.

---

## Foreword

This recommended practice was developed by the AAMI Reusable Surgical Textile Processing Working Group under the auspices of the AAMI Sterilization Standards Committee. The objective of this recommended practice is to provide guidance in the handling and processing of reusable surgical textiles.

This recommended practice is the second edition of ANSI/AAMI ST65:2000, *Processing of Reusable Surgical Textiles for Use in Health Care Facilities*. In addition to a general updating of the glossary and bibliography, the new edition reflects the incorporation of new provisions regarding sharps precautions, stains and discolorations, ANSI/AAMI PB70, and folding, as well as the revision of the recommendations regarding bleaching, pack identification, the transport and storage of textiles, strike-through, and medical device regulatory considerations.

The provisions of this recommended practice should be reviewed by various department managers, as applicable, and adapted to the needs of their particular institutions. Written policies and procedures should be developed and implemented in consultation with representatives of end users.

This recommended practice reflects the conscientious efforts of health care professionals, in cooperation with representatives of hospital-owned laundries and outsourcing services, to develop recommendations for optimum performance levels in the processing of reusable surgical textiles. It is not intended that these recommendations be construed as universally applicable in all circumstances. Also, it is recognized that in many cases these recommendations might not be immediately achievable. Therefore, the document should be used to guide personnel toward desirable performance objectives, and all of its provisions should be considered and applied in the light of professional judgment and experience.

The concepts incorporated in this recommended practice should be considered flexible and dynamic. The recommendations set forth in this document are reviewed and updated periodically to assimilate progressive technological developments. AAMI policies and procedures require that AAMI standards and recommended practices be reviewed and, if necessary, revised at least once every 5 years.

As used within the context of this document, “shall” indicates requirements strictly to be followed in order to conform to the recommended practice; “should” indicates that among several possibilities one is recommended as particularly suitable, without mentioning or excluding others, or that a certain course of action is preferred but not necessarily required, or that (in the negative form) a certain possibility or course of action should be avoided but is not prohibited; “may” is used to indicate that a course of action is permissible within the limits of the recommended practice; and “can” is used as a statement of possibility and capability. “Must” is used only to describe “unavoidable” situations, including those mandated by government regulation.

Suggestions for improving this recommended practice are invited. Comments and suggested revisions should be sent to Standards Department, AAMI, 1110 N. Glebe Road, Suite 220, Arlington, VA 22201-4795.

---

NOTE—This foreword does not contain provisions of ANSI/AAMI ST65, *Processing of reusable surgical textiles for use in health care facilities*, but it does provide important information about the development and intended use of the document.

---

## **Introduction: Need for the recommended practice**

Hospital-owned laundries, as well as those providing outsourcing services to health care facilities, are providing clean and disinfected surgical textiles, nonsterile reusable surgical textile packs, and/or sterile reusable surgical textile packs to health care facilities. The ability of service providers to furnish reusable products that meet the performance requirements of end users while providing quality patient care has been enhanced by technological advances in reusable textiles and processing equipment. Current and future advancements in this segment of the industry may necessitate modified or alternative processing techniques.

This recommended practice is intended to provide guidelines that will help materiel managers, laundry managers, central service managers, and other health care professionals implement effective quality assurance systems for the processing of reusable surgical textiles. The guidelines provided here may also be useful to hospitals/users in evaluating the capabilities of facilities being considered for the processing of reusable surgical textiles.

It should be noted that laundry facilities that place surgical textiles (whether sterile or nonsterile) into commercial distribution come under the jurisdiction of Food and Drug Administration (FDA) regulations (see Section 12). In addition, laundry facilities in general may be subject to local, state, and/or federal Environmental Protection Agency (EPA) and Occupational Safety and Health Administration (OSHA) requirements.

# Processing of reusable surgical textiles for use in health care facilities

## 1 Scope

### 1.1 General

This recommended practice provides guidelines for properly handling, processing, and preparing reusable surgical textiles for use in health care facilities. These guidelines describe a quality assurance program for the processing of reusable surgical textiles, including processes and techniques for the preparation of clean bulk items for delivery to user sites and the assembly of textile packs for sterilization prior to end use. These guidelines apply to all facilities that process surgical textiles, whether on-premise laundries (OPLs), hospital-owned cooperatives, or commercial facilities.

NOTE 1—Surgical textiles labeled for single use only should not be reprocessed or reused, because it may not be possible to adequately reprocess them and maintain their performance and safety attributes. In addition, the health care facility's liability may be affected if the manufacturer's written instructions for use are not followed. See also FDA regulations applicable to the reprocessing of single-use medical devices ([www.fda.gov/cdrh/reprocessing/](http://www.fda.gov/cdrh/reprocessing/)).

NOTE 2—For purposes of this recommended practice, "health care facility" means hospitals, nursing homes, extended care facilities, freestanding surgical centers, clinics, and medical, surgical, and dental offices. For convenience, the term "hospital" is sometimes used in this recommended practice; in all instances, the term should be taken to encompass all other health care facilities.

### 1.2 Inclusions

This recommended practice specifically addresses

- a) design criteria for functional work areas involved with the receiving, staging, and handling of soiled surgical textiles; the separation of soiled and clean textiles; the laundering of reusable surgical textiles; and the inspection and preparation of clean bulk items and surgical textile packs;
- b) staff qualifications, education, training, dress codes, and other personnel considerations;
- c) transporting, receiving, and handling of both newly purchased and soiled surgical textiles;
- d) laundry processing (loading, washing, drying) recommendations;
- e) inspection, testing, and maintenance of laundered textiles;
- f) preparation and packaging of laundered textiles;
- g) handling, transport, and storage of laundered textiles;
- h) installation, operation, care, and maintenance of laundry equipment;
- i) quality control measures, procedures, and practices;
- j) medical device regulatory considerations.

Definitions of terms and a bibliography are also provided in this recommended practice.

### 1.3 Exclusions

This recommended practice does *not* cover

- a) design or construction criteria for equipment used to process reusable surgical textiles;

NOTE—Performance requirements for washer–disinfectors are provided in ISO 15883-1, 15883-2, and 15883-3. Performance requirements for hospital steam sterilizers are provided in ANSI/AAMI ST8.