

# Technical Information Report

## AAMI TIR28: 2009/(R)2013

Product adoption  
and process equivalence  
for ethylene oxide  
sterilization



# **Product adoption and process equivalence for ethylene oxide sterilization**

Approved 17 March 2009 and reaffirmed 20 May 2013 by  
**Association for the Advancement of Medical Instrumentation**

**Abstract:** This technical information report provides guidance for the adoption of new or modified products into an existing validated sterilization process and for the determination of equivalence of a sterilization process as conducted with different equipment. Guidance is intended to augment the ANSI/AAMI/ISO 11135 series in the areas of product adoption and process equivalence.

**Keywords:** sterilization, ethylene oxide, product adoption, process equivalence, product family

## AAMI Technical Information Report

A technical information report (TIR) is a publication of the Association for the Advancement of Medical Instrumentation (AAMI) Standards Board that addresses a particular aspect of medical technology.

Although the material presented in a TIR may need further evaluation by experts, releasing the information is valuable because the industry and the professions have an immediate need for it.

A TIR differs markedly from a standard or recommended practice, and readers should understand the differences between these documents.

Standards and recommended practices are subject to a formal process of committee approval, public review, and resolution of all comments. This process of consensus is supervised by the AAMI Standards Board and, in the case of American National Standards, by the American National Standards Institute.

A TIR is not subject to the same formal approval process as a standard. However, a TIR is approved for distribution by a technical committee and the AAMI Standards Board.

Another difference is that, although both standards and TIRs are periodically reviewed, a standard must be acted on—reaffirmed, revised, or withdrawn—and the action formally approved usually every five years but at least every 10 years. For a TIR, AAMI consults with a technical committee about five years after the publication date (and periodically thereafter) for guidance on whether the document is still useful—that is, to check that the information is relevant or of historical value. If the information is not useful, the TIR is removed from circulation.

A TIR may be developed because it is more responsive to underlying safety or performance issues than a standard or recommended practice or because achieving consensus is extremely difficult or unlikely. Unlike a standard, a TIR permits the inclusion of differing viewpoints on technical issues.

**CAUTION NOTICE:** This AAMI TIR may be revised or withdrawn at any time. Because it addresses a rapidly evolving field or technology, readers are cautioned to ensure that they have also considered information that may be more recent than this document.

All standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Comments on this technical information report are invited and should be sent to AAMI, Attn: Standards Department, 1110 N. Glebe Road, Suite 220, Arlington, VA 22201-4795.

*Published by*

Association for the Advancement of Medical Instrumentation  
1110 N. Glebe Road, Suite 220  
Arlington, VA 22201-4795  
[www.aami.org](http://www.aami.org)

© 2009 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, complete the reprint request form at [www.aami.org](http://www.aami.org) or contact AAMI at 1110 N. Glebe Road, Suite 220, Arlington, VA 22201-4795. Phone: (703) 525-4890; Fax: (703) 276-0793.

Printed in the United States of America

**ISBN 1-57020-340-7**

# Contents

	<b>Page</b>
Glossary of equivalent standards .....	iv
Committee representation .....	vi
Foreword .....	ix
1 Scope .....	1
2 Terms and definitions .....	1
3 Product adoption .....	2
3.1 Establishment of an EO product family .....	2
3.2 Establishment of EO processing group .....	3
3.3 Evaluation for product adoption .....	3
3.3.1 Determination of adverse effects .....	3
3.3.2 Determination of product design effects .....	4
3.3.3 Determination of product material and characteristics effects .....	4
3.3.4 Determination of sterile barrier system effects .....	4
3.3.5 Determination of load configuration effects .....	4
3.4 Conclusions of product adoption evaluation .....	4
3.4.1 Comparative resistance study .....	4
3.4.2 Temperature and humidity distribution study .....	5
3.4.3 Residual study .....	5
3.5 Maintenance of EO product adoption .....	5
3.6 Documentation .....	5
4 Process equivalence .....	6
4.1 General .....	6
4.2 Requirements of process equivalence .....	6
4.3 Determination of process equivalence .....	6
4.3.1 Process analysis and evaluation .....	6
4.3.2 Microbiological evaluation .....	7
4.3.3 Results evaluation .....	8
4.4 Process requalification and maintenance of equivalence .....	8
4.5 Documentation .....	8
Annex A Guide for evaluation of a product for adoption into an EO product family or EO processing group .....	10
Bibliography .....	12

## Glossary of equivalent standards

International Standards adopted in the United States may include normative references to other International Standards. For each International Standard that has been adopted by AAMI (and ANSI), the table below gives the corresponding U.S. designation and level of equivalency to the International Standard. NOTE: Documents are sorted by international designation.

Other normatively referenced International Standards may be under consideration for U.S. adoption by AAMI; therefore, this list should not be considered exhaustive.

International designation	U.S. designation	Equivalency
IEC 60601-1:2005	ANSI/AAMI ES60601-1:2005	Major technical variations
IEC 60601-1-2:2007	ANSI/AAMI/IEC 60601-1-2:2007	Identical
IEC 60601-2-2:2009	ANSI/AAMI/IEC 60601-2-2:2009	Identical
IEC 60601-2-4:2002	ANSI/AAMI DF80:2003	Major technical variations
IEC 60601-2-19:2009	ANSI/AAMI/IEC 60601-2-19:2009	Identical
IEC 60601-2-20:2009	ANSI/AAMI/IEC 60601-2-20:2009	Identical
IEC 60601-2-21:2009	ANSI/AAMI/IEC 60601-2-21:2009	Identical
IEC 60601-2-24:1998	ANSI/AAMI ID26:2004	Major technical variations
IEC 60601-2-47:2001	ANSI/AAMI EC38:2007	Major technical variations
IEC 60601-2-50:2009	ANSI/AAMI/IEC 60601-2-50:2009	Identical
IEC 80601-2-58:2008	ANSI/AAMI/IEC 80601-2-58:2008	Identical
IEC/TR 60878:2009	ANSI/AAMI/IEC TIR60878:2003	Identical
IEC/TR 62296:2009	ANSI/AAMI/IEC TIR62296:2009	Identical
IEC 62304:2006	ANSI/AAMI/IEC 62304:2006	Identical
IEC/TR 62348:2006	ANSI/AAMI/IEC TIR62348:2006	Identical
ISO 5840:2005	ANSI/AAMI/ISO 5840:2005	Identical
ISO 7198:1998	ANSI/AAMI/ISO 7198:1998/2001/(R)2004	Identical
ISO 7199:2009	ANSI/AAMI/ISO 7199:2009	Identical
ISO 8637:2004	ANSI/AAMI RD16:2007	Major technical variations
ISO 8638:2004	ANSI/AAMI RD17:2007	Major technical variations
ISO 10993-1:2003	ANSI/AAMI/ISO 10993-1:2003	Identical
ISO 10993-2:2006	ANSI/AAMI/ISO 10993-2:2006	Identical
ISO 10993-3:2003	ANSI/AAMI/ISO 10993-3:2003	Identical
ISO 10993-4:2002 and Amendment 1:2006	ANSI/AAMI/ISO 10993-4:2002/(R)2009 and Amendment 1:2006/(R)2009	Identical
ISO 10993-5:1999	ANSI/AAMI/ISO 10993-5:1999	Identical
ISO 10993-6:2007	ANSI/AAMI/ISO 10993-6:2007	Identical
ISO 10993-7:2008	ANSI/AAMI/ISO 10993-7:2008	Identical
ISO 10993-9:1999	ANSI/AAMI/ISO 10993-9:1999/(R)2005	Identical
ISO 10993-10:2002 and Amendment 1:2006	ANSI/AAMI BE78:2002/(R)2008 ANSI/AAMI BE78:2002/A1:2006/(R)2008	Minor technical variations Identical
ISO 10993-11:2006	ANSI/AAMI/ISO 10993-11:2006	Identical
ISO 10993-12:2007	ANSI/AAMI/ISO 10993-12:2007	Identical
ISO 10993-13:1998	ANSI/AAMI/ISO 10993-13:1999/(R)2004	Identical
ISO 10993-14:2001	ANSI/AAMI/ISO 10993-14:2001/(R)2006	Identical
ISO 10993-15:2000	ANSI/AAMI/ISO 10993-15:2000/(R)2006	Identical
ISO 10993-16:1997	ANSI/AAMI/ISO 10993-16:1997/(R)2009	Identical
ISO 10993-17:2002	ANSI/AAMI/ISO 10993-17:2002/(R)2008	Identical
ISO 10993-18:2005	ANSI/AAMI BE83:2006	Major technical variations
ISO/TS 10993-19:2006	ANSI/AAMI/ISO TIR10993-19:2006	Identical
ISO/TS 10993-20:2006	ANSI/AAMI/ISO TIR10993-20:2006	Identical

<b>International designation</b>	<b>U.S. designation</b>	<b>Equivalency</b>
ISO 11135-1:2007	ANSI/AAMI/ISO 11135-1:2007	Identical
ISO/TS 11135-2:2008	ANSI/AAMI/ISO TIR11135-2:2008	Identical
ISO 11137-1:2006	ANSI/AAMI/ISO 11137-1:2006	Identical
ISO 11137-2:2006 (2006-08-01 corrected version)	ANSI/AAMI/ISO 11137-2:2006	Identical
ISO 11137-3:2006	ANSI/AAMI/ISO 11137-3:2006	Identical
ISO 11138-1: 2006	ANSI/AAMI/ISO 11138-1:2006	Identical
ISO 11138-2: 2006	ANSI/AAMI/ISO 11138-2:2006	Identical
ISO 11138-3: 2006	ANSI/AAMI/ISO 11138-3:2006	Identical
ISO 11138-4: 2006	ANSI/AAMI/ISO 11138-4:2006	Identical
ISO 11138-5: 2006	ANSI/AAMI/ISO 11138-5:2006	Identical
ISO/TS 11139:2006	ANSI/AAMI/ISO 11139:2006	Identical
ISO 11140-1:2005	ANSI/AAMI/ISO 11140-1:2005	Identical
ISO 11140-3:2007	ANSI/AAMI/ISO 11140-3:2007	Identical
ISO 11140-4:2007	ANSI/AAMI/ISO 11140-4:2007	Identical
ISO 11140-5:2007	ANSI/AAMI/ISO 11140-5:2007	Identical
ISO 11607-1:2006	ANSI/AAMI/ISO 11607-1:2006	Identical
ISO 11607-2:2006	ANSI/AAMI/ISO 11607-2:2006	Identical
ISO 11737-1: 2006	ANSI/AAMI/ISO 11737-1:2006	Identical
ISO 11737-2:1998	ANSI/AAMI/ISO 11737-2:1998	Identical
ISO 13408-1:2008	ANSI/AAMI/ISO 13408-1:2008	Identical
ISO 13408-2:2003	ANSI/AAMI/ISO 13408-2:2003	Identical
ISO 13408-3:2006	ANSI/AAMI/ISO 13408-3:2006	Identical
ISO 13408-4:2005	ANSI/AAMI/ISO 13408-4:2005	Identical
ISO 13408-5:2006	ANSI/AAMI/ISO 13408-5:2006	Identical
ISO 13408-6:2006	ANSI/AAMI/ISO 13408-6:2006	Identical
ISO 13485:2003	ANSI/AAMI/ISO 13485:2003	Identical
ISO 14155-1:2003	ANSI/AAMI/ISO 14155-1:2003/(R)2008	Identical
ISO 14155-2:2003	ANSI/AAMI/ISO 14155-2:2003/(R)2008	Identical
ISO 14160:1998	ANSI/AAMI/ISO 14160:1998/(R)2008	Identical
ISO 14161:2000	ANSI/AAMI/ISO 14161:2000	Identical
ISO 14937:2000	ANSI/AAMI/ISO 14937:2000	Identical
ISO/TR 14969:2004	ANSI/AAMI/ISO TIR14969:2004	Identical
ISO 14971:2007	ANSI/AAMI/ISO 14971:2007	Identical
ISO 15223-1:2007 and A1:2008	ANSI/AAMI/ISO 15223-1:2007 and A1:2008	Identical
ISO 15225:2000 and A1:2004	ANSI/AAMI/ISO 15225:2000/(R)2006 and A1:2004/(R)2006	Identical
ISO 15674:2009	ANSI/AAMI/ISO 15674:2009	Identical
ISO 15675:2009	ANSI/AAMI/ISO 15675:2009	Identical
ISO 15882:2008	ANSI/AAMI/ISO 15882:2008	Identical
ISO/TR 16142:2006	ANSI/AAMI/ISO TIR16142:2005	Identical
ISO 17664:2004	ANSI/AAMI ST81:2004	Major technical variations
ISO 17665-1:2006	ANSI/AAMI/ISO 17665-1:2006	Identical
ISO 18472:2006	ANSI/AAMI/ISO 18472:2006	Identical
ISO/TS 19218:2005	ANSI/AAMI/ISO 19218:2005	Identical
ISO 22442-1:2007	ANSI/AAMI/ISO 22442-1:2007	Identical
ISO 22442-2:2007	ANSI/AAMI/ISO 22442-2:2007	Identical
ISO 22442-3:2007	ANSI/AAMI/ISO 22442-3:2007	Identical
ISO 25539-1:2003 and A1:2005	ANSI/AAMI/ISO 25539-1:2003 and A1:2005	Identical
ISO 25539-2:2008	ANSI/AAMI/ISO 25539-2:2008	Identical
ISO 81060-1:2007	ANSI/AAMI/ISO 81060-1:2007	Identical

## Committee representation

### Association for the Advancement of Medical Instrumentation Industrial Ethylene Oxide Sterilization Working Group

This technical information report (TIR) was developed by the AAMI Industrial Ethylene Oxide Sterilization Working Group under the auspices of the AAMI Sterilization Standards Committee. Working Group approval of the TIR does not necessarily imply that all committee members voted for its approval.

At the time this document was published, the **AAMI Industrial Ethylene Oxide Sterilization Working Group** had the following members:

<i>Cochairs</i>	Charles Cogdill, Boston Scientific Corporation Gerry A. O'Dell, MS, Gerry O'Dell Consulting
<i>Members</i>	Curt Andrew Bogue, BAS, Cook Inc. Anne F. Booth, MS, Conmed Corporation Lloyd Brown, Covidien Bradley J. Catalone, PhD, Olympus America Inc. Dennis E. Christensen, BS, Process Challenge Devices Charlie Christianson, St. Jude Medical Inc. Charles Cogdill, Boston Scientific Corporation Gary N. Cranston, Consulting & Technical Services/PCS Douglas D. Davie, Sterilization Validation Services Shawn A. Doyle, Sterilator Company Inc. Brian R. Drumheller, CR Bard William Facemire, Independent Expert William F. FitzGerald, PE, FitzGerald & Associates Ltd. Dan B. Floyd, RM, Nelson Laboratories Inc. Zory R. Glaser, PhD MPH CSPDM, (Independent Expert) Arthur C. Harris, Cook Inc. Deborah A. Havlik, Hospira Worldwide Inc. Danny Hutson, Cardinal Health (MP&S) Jim Kaiser, Bausch & Lomb Inc. Bert Kingsbury, Terumo Medical Corporation Carolyn L. Kinsley, LexaMed Karen A. Kowalczyk, Centurion Sterilization Services Christine Loshbaugh, Edwards Lifesciences Ted May, Andersen Products Inc. David Ford McGoldrick, BS, Abbott Laboratories Craig A. Meadows, Medtronic Inc. Joseph M. Mello, Ethide Laboratories Inc. Russell D. Mills, Zimmer Inc. Gary Mitchel, PE, Johnson & Johnson Sarah A. Mowitt, (Independent Expert) Gerry A. O'Dell, MS, Gerry O'Dell Consulting Ken Paddock, Baxter Healthcare Corporation Dave Parente, NAMSA Manuel Saavedra, Jr., Kimberly-Clark Corporation Zenius V. Seliokas, Stericon Inc. Jon Seulean, CaridianBCT Sterilization Services Inc. Barb Smith, Getinge USA Bill South, Steris Corporation Ralph Stick, WuXi AppTec Wayne Swallow, Becton Dickinson & Company Radhakrishna S. Tirumalai, US Pharmacopeia Convention Inc. Steven E. Turtill, FDA/CDRH Jason Voisinet, Ethox International Inc. Craig A. Wallace, 3M Healthcare Richard L. Weisman, Fresenius Medical Care Renal Therapies Group Casimir John Woss, PhD, Alcon Laboratories Inc. William T. Young, Sterigenics International

*Alternates*

August Baur, Centurion Sterilization Services  
John Broad, NAMSA  
Delores Bruce, Steris Corporation  
Trabue D. Bryans, WuXi AppTec  
Susan Bullis, Johnson & Johnson  
Tim Carlson, Becton Dickinson & Company  
John DiCaro, Covidien  
Joseph R. Durbin, Hospira Worldwide Inc.  
Mark Fischer, Nelson Laboratories Inc.  
David Michael Gasparik, Cardinal Health (MP&S)  
Thomas L. Hansen, Terumo Medical Corporation  
Donna Horner, Abbott Laboratories  
Brent Huberty, St. Jude Medical Inc.  
Ezra Koski, A, CaridianBCT Sterilization Services Inc  
John M. Kuchinski, FDA/CDRH  
James P. Kulla, BS MS, LexaMed  
Richard Lenz, Medtronic Inc.  
John Lindley, Andersen Products Inc.  
Ralph Makinen, Boston Scientific Corporation  
Jeff Martin, Alcon Laboratories Inc.  
Janette Martinez, BS, Edwards Lifesciences  
Michael E. Mattison, Getinge USA  
Mary S. Mayo, CR Bard  
Michael A. Padilla, BSME, Sterigenics International  
Frank Peacock, Jr., Bausch & Lomb Inc.  
Jeff Peltier, Boston Scientific Corporation  
Nancy Rakiewicz, Ethox International Inc.  
Tyrone S. Rouse, Kimberly-Clark Corporation  
David Silor, Zimmer Inc.  
Larry Talapa, 3M Healthcare

AAMI also acknowledges the **TIR28 Editing Task Group**, comprised of the following members, for its special contribution in the development of this document:

Task Group Leader: Gerry A. O'Dell, MS, Gerry O'Dell Consulting

Task Group Members: David Michael Gasparik, Cardinal Health (MP&S)  
Ralph Makinen, Boston Scientific Corporation  
Michael A. Padilla, BSME, Sterigenics International  
Dave Parente, NAMSA  
Jeff Peltier, Boston Scientific Corporation

At the time this document was published, the **AAMI Sterilization Standards Committee** had the following members:

*Chairs*

Victoria M. Hitchins, PhD, FDA/CDRH  
William E. Young, Boston Scientific Corporation

*Members*

Trabue D. Bryans, WuXi AppTec  
Peter A. Burke, PhD, Steris Corporation  
Nancy Chobin, RN CSPDM, St. Barnabas Healthcare System (Independent Expert)  
Charles Cogdill, Boston Scientific Corporation  
Ramona Conner, RN MSN CNOR, Association of periOperative Registered Nurses  
Jacqueline Daley, Association for Professionals in Infection Control and Epidemiology  
Kimbrell Darnell, CR Bard  
Lisa Foster, Sterigenics International  
Joel R. Gorski, PhD, NAMSA  
Deborah A. Havlik, Hospira Worldwide Inc.  
Victoria M. Hitchins, PhD, FDA/CDRH  
Danny Hutson, Cardinal Health (MP&S)  
Lois Atkinson Jones, MS, (Independent Expert)  
Susan G. Klacik, CCSMC FCS ACE, IAHCSSM  
Byron J. Lambert, PhD, Abbott Laboratories  
Colleen Patricia Landers, RN, Canadian Standards Association

Lisa N. Macdonald, Becton Dickinson & Company  
Jeff Martin, Alcon Laboratories Inc.  
Patrick J. McCormick, PhD, Bausch & Lomb Inc.  
Rainer Newman, Johnson & Johnson  
Janet M. Prust, 3M Healthcare  
Nancy Rakiewicz, Ethox International Inc.  
Michael H. Scholla, Dupont Nonwovens  
Mark Seybold, Baxter Healthcare Corporation  
Andrew Sharavara, PhD, Propper Manufacturing Co. Inc.  
Mark N. Smith, Getinge USA  
William N. Thompson, Covidien  
James L. Whitby, MA MB FRCP, (Independent Expert)  
Martell Kress Winters, BS SM, Nelson Laboratories Inc.

*Alternates*

Lloyd Brown, Covidien  
Glenn W. Calvert, Becton Dickinson & Company  
Dave Dion, Cardinal Health (MP&S)  
Thomas J. Frazar, Johnson & Johnson  
Kathy Hoffman, Sterigenics International  
Jim Kaiser, Bausch & Lomb Inc.  
Joseph J. Lasich, BS, Alcon Laboratories Inc.  
Chiu S. Lin, PhD, FDA/CDRH  
Natalie Lind, IAHCSSM  
Ralph Makinen, Boston Scientific Corporation  
Mary S. Mayo, CR Bard  
David Ford McGoldrick, BS, Abbott Laboratories  
Jerry R. Nelson, PhD, Nelson Laboratories Inc.  
Karen Polkinghorne, Dupont Nonwovens  
Mike Sadowski, Baxter Healthcare Corporation  
John R. Scoville, Jr., Steris Corporation  
Ralph Stick, WuXi AppTec  
Jason Voisinet, Ethox International Inc.  
Craig A. Wallace, 3M Healthcare  
Valerie Welter, Hospira Worldwide Inc.  
William E. Young, Boston Scientific Corporation

---

NOTE—Participation by federal agency representatives in the development of this technical information report does not constitute endorsement by the federal government or any of its agencies.

---

## Foreword

This document is part of a series of technical information reports (TIRs) intended for use in conjunction with ANSI/AAMI/ISO 11135-1:2007. The other reports in the series are

- AAMI TIR14:2009, Contract sterilization using ethylene oxide;
- AAMI TIR15:1997, Ethylene oxide sterilization equipment, process considerations, and pertinent calculations (currently under revision);
- AAMI TIR16:2000, Process development and performance qualification for ethylene oxide sterilization— Microbiological aspects (currently under revision); and
- ANSI/AAMI/ISO TIR11135-2:2008, Sterilization of health care products - Ethylene oxide - Part 2: Guidance on the application of ANSI/AAMI/ISO 11135-1

The original TIR28, along with other AAMI TIRs, provided additional guidance to the 1994 edition of the industrial EO sterilization standard 11135, which was revised in 2007 under a new designation, ANSI/AAMI/ISO 11135-1:2007, *Sterilization of health care products - Ethylene oxide - Part 1: Requirements for the development, validation and routine control of a sterilization process for medical devices*. In 2008, ISO published its own guidance document for the 11135 standard, ISO/TR 11135-2:2008, *Sterilization of health care products - Ethylene oxide - Part 2: Guidance on the application of ISO 11135-1*, which was based to a great extent on the earlier AAMI technical information reports. Correspondingly, the AAMI Industrial EO sterilization working group is updating its TIRs to take into account changes to the 11135 standard as well as to avoid redundancy with ANSI/AAMI/ISO TIR11135-2:2008.

This TIR provides guidance for the adoption of new or modified products into an existing validated sterilization process and for the determination of equivalence of a sterilization process as conducted with different equipment. These areas are not addressed in depth by ANSI/AAMI/ISO 11135-1:2007 or ANSI/AAMI/ISO 11135-2:2008, but they are important industry practices that are used to reduce the expense and time associated with the validation or requalification of an ethylene oxide sterilization process and are based on accumulated process knowledge.

The adoption of a new or modified product into an existing validated sterilization process involves the determination that the product is no more of a challenge than the product (i.e., master product or representative product) or process challenge device that was used in the performance qualification for the ethylene oxide sterilization process.

The process equivalence section of this TIR will provide guidance on how to establish equivalence between processes performed in separate equipment or sets of equipment and guidance on the level of microbiological performance qualification testing required. It also includes guidance on the maintenance of equivalence and requalification when equivalence has been established.

This TIR contains guidelines that are not intended to be absolute or to apply in all circumstances. One should use judgment in applying the information in this TIR.

As used within the context of this document, “should” indicates that among several possibilities, one is recommended as particularly suitable, without mentioning or excluding others, or that a certain course of action is preferred but not necessarily required, or that (in the negative form) a certain possibility or course of action should be avoided but is not prohibited. “May” is used to indicate that a course of action is permissible within the limits of the recommended practice. “Can” is used as a statement of possibility and capability. Finally, “must” is used only to describe “unavoidable” situations, including those mandated by government regulation. See also the NOTE on Page 1.

Suggestions for improving this technical information report are invited. Comments and suggested revisions should be sent to AAMI, 1110 N. Glebe Road, Suite 220, Arlington, VA 22201-4795.

---

NOTE—This foreword does not contain provisions of the AAMI TIR28:2009 titled *Product adoption and process equivalence for ethylene oxide sterilization*, but it does provide important information about the development and intended use of the document.

---



# Product adoption and process equivalence for ethylene oxide sterilization

NOTE—This technical information report is not a standard and the material contained herein is not normative in nature. The committee has in a few places used the term "shall" based on their knowledge of requirements contained in relevant standards and/or regulatory requirements.

## 1 Scope

This TIR addresses medical devices that are processed by ethylene oxide (EO) sterilization using conventional or parametric product release. The document applies to the following situations for the sterilization of medical devices:

- a) a new product is being added to the previously validated process,
- b) changes to validated products are being evaluated,
- c) a previously validated process is being moved to a different facility or to different equipment, and
- d) equivalency of a sterilization process is being evaluated.

Although the information presented was developed for application to medical devices, the content of this guideline may also be applied to other relevant products or materials.

## 2 Terms and definitions

For the purposes of this AAMI TIR, the following terms and definitions apply:

**2.1 candidate equipment:** New or modified piece of equipment intended to deliver an existing validated process.

**2.2 candidate product:** New or modified product, including the packaging system, proposed for inclusion in the existing validated sterilization process.

**2.3 EO processing group:** Collection of products or product families that can be sterilized in the same EO sterilization process.

NOTE—All products within the processing group have been determined to present an equal or lesser challenge to the sterilization process than the challenge device for that group.

[ANSI/AAMI/ISO 11135-2:2008]

**2.4 EO product family:** Collection of products determined to be similar or equivalent for validation purposes.

[ANSI/AAMI/ISO 11135-2:2008]

**2.5 installation qualification (IQ):** Process of obtaining and documenting evidence that equipment has been provided and installed in accordance with its specification.

[ANSI/AAMI/ISO TIR11139:2006]

**2.6 load configuration:** Totality of attributes defining the presentation of the product to the sterilization process. This configuration includes (a) the orientation of the product within the sterile barrier system (primary package); (b) the quantity and orientation of the primary package within the protective packaging (secondary or tertiary package); (c) the quantity, orientation, and placement of the product in the protective packaging on the sterilizer pallets (or within the carriers); and (d) the quantity and placement of the pallets (or carriers) within the vessel or area.

**2.7 operational qualification (OQ):** Process of obtaining and documenting evidence that installed equipment operates within predetermined limits when used in accordance with its operational procedures.

[ANSI/AAMI/ISO TIR11139:2006]