

American National Standard



ANSI/AAMI
HA60601-1-
11:2015

(IEC 60601-1-11:2015, MOD)

MEDICAL ELECTRICAL EQUIPMENT — Part 1-11:
General requirements for basic safety and essential
performance – Collateral Standard: Requirements for
medical electrical equipment and medical electrical
systems used in the home healthcare environment

Objectives and uses of AAMI standards and recommended practices

It is most important that the objectives and potential uses of an AAMI product standard or recommended practice are clearly understood. The objectives of AAMI's technical development program derive from AAMI's overall mission: the advancement of medical instrumentation. Essential to such advancement are (1) a continued increase in the safe and effective application of current technologies to patient care, and (2) the encouragement of new technologies. It is AAMI's view that standards and recommended practices can contribute significantly to the advancement of medical instrumentation, provided that they are drafted with attention to these objectives and provided that arbitrary and restrictive uses are avoided.

A voluntary *standard* for a *medical device* recommends to the manufacturer the information that should be provided with or on the product, basic safety and performance criteria that should be considered in qualifying the device for clinical use, and the measurement techniques that can be used to determine whether the device conforms with the safety and performance criteria and/or to compare the performance characteristics of different products. Some standards emphasize the information that should be provided with the device, including performance characteristics, instructions for use, warnings and precautions, and other data considered important in ensuring the safe and effective use of the device in the clinical environment. Recommending the disclosure of performance characteristics often necessitates the development of specialized test methods to facilitate uniformity in reporting; reaching consensus on these tests can represent a considerable part of committee work. When a drafting committee determines that clinical concerns warrant the establishment of *minimum* safety and performance criteria, referee tests must be provided and the reasons for establishing the criteria must be documented in the rationale.

A *recommended practice* provides guidelines for the use, care, and/or processing of a medical device or system. A recommended practice does not address device performance *per se*, but rather procedures and practices that will help ensure that a device is used safely and effectively and that its performance will be maintained.

Although a device standard is primarily directed to the manufacturer, it may also be of value to the potential purchaser or user of the device as a frame of reference for device evaluation. Similarly, even though a recommended practice is usually oriented towards healthcare professionals, it may be useful to the manufacturer in better understanding the environment in which a medical device will be used. Also, some recommended practices, while not addressing device performance criteria, provide guidelines to industrial personnel on such subjects as sterilization processing, methods of collecting data to establish safety and efficacy, human engineering, and other processing or evaluation techniques; such guidelines may be useful to health care professionals in understanding industrial practices.

In determining whether an AAMI standard or recommended practice is relevant to the specific needs of a potential user of the document, several important concepts must be recognized:

All AAMI standards and recommended practices are *voluntary* (unless, of course, they are adopted by government regulatory or procurement authorities). The application of a standard or recommended practice is solely within the discretion and professional judgment of the user of the document.

Each AAMI standard or recommended practice reflects the collective expertise of a committee of health care professionals and industrial representatives, whose work has been reviewed nationally (and sometimes internationally). As such, the consensus recommendations embodied in a standard or recommended practice are intended to respond to clinical needs and, ultimately, to help ensure patient safety. A standard or recommended practice is limited, however, in the sense that it responds generally to perceived risks and conditions that may not always be relevant to specific situations. A standard or recommended practice is an important *reference* in responsible decision-making, but it should never *replace* responsible decision-making.

Despite periodic review and revision (at least once every five years), a standard or recommended practice is necessarily a static document applied to a dynamic technology. Therefore, a standards user must carefully review the reasons why the document was initially developed and the specific rationale for each of its provisions. This review will reveal whether the document remains relevant to the specific needs of the user.

Particular care should be taken in applying a product standard to existing devices and equipment, and in applying a recommended practice to current procedures and practices. While observed or potential risks with existing equipment typically form the basis for the safety and performance criteria defined in a standard, professional judgment must be used in applying these criteria to existing equipment. No single source of information will serve to identify a particular product as "unsafe". A voluntary standard can be used as one resource, but the ultimate decision as to product safety and efficacy must take into account the specifics of its utilization and, of course, cost-benefit considerations. Similarly, a recommended practice should be analyzed in the context of the specific needs and resources of the individual institution or firm. Again, the rationale accompanying each AAMI standard and recommended practice is an excellent guide to the reasoning and data underlying its provision.

In summary, a standard or recommended practice is truly useful only when it is used in conjunction with other sources of information and policy guidance and in the context of professional experience and judgment.

INTERPRETATIONS OF AAMI STANDARDS AND RECOMMENDED PRACTICES

Requests for interpretations of AAMI standards and recommended practices must be made in writing, to the AAMI Vice President, Standards Policy and Programs. An official interpretation must be approved by letter ballot of the originating committee and subsequently reviewed and approved by the AAMI Standards Board. The interpretation will become official and representation of the Association only upon exhaustion of any appeals and upon publication of notice of interpretation in the "Standards Monitor" section of the *AAMI News*. The Association for the Advancement of Medical Instrumentation disclaims responsibility for any characterization or explanation of a standard or recommended practice which has not been developed and communicated in accordance with this procedure and which is not published, by appropriate notice, as an *official interpretation* in the *AAMI News*.

MEDICAL ELECTRICAL EQUIPMENT — Part 1-11: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment

Approved 7 August 2015 by
Association for the Advancement of Medical Instrumentation

Approved 25 August 2015 by
American National Standards Institute, Inc.

Abstract: This standard applies to the safety and essential performance of medical electrical equipment and medical electrical systems, which are intended by the manufacturer for use in home care applications usually without continual professional supervision and temporarily in the clinical environment. These medical electrical equipment and medical electrical systems will frequently be used in locations where driving power and safety means of the electrical installation is not reliable. These medical electrical equipment and medical electrical systems will often be supervised by non-healthcare personnel with different levels of training.

Keywords: medical electrical equipment, safety, home care, home use

Published by

Association for the Advancement of Medical Instrumentation
4301 N Fairfax Drive, Suite 301
Arlington, VA 22203-1633
www.aami.org

© 2015 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, contact AAMI at 4301 N. Fairfax Drive, Suite 301, Arlington, VA 22203-1633. Phone: (703) 525-4890; Fax: (703) 525-1067.

Printed in the United States of America

ISBN 1-57020-593-0

Contents

Glossary of equivalent standards.....	vii
Committee representation	viii
Background of ANSI/AAMI adoption of IEC 60601-1-11:2015	x
Foreword	xi
U.S. Deviations to IEC 60601-1-11:2015	xiv
Introduction.....	xv
1 Scope, object and related standards.....	1
1.1 * Scope	1
1.2 Object	1
1.3 Related standards.....	1
1.3.1 IEC 60601-1	1
1.3.2 Particular standards	2
2 Normative references	2
3 Terms and definitions	3
4 General requirements.....	4
4.1 * Additional requirements for SUPPLY MAINS for ME EQUIPMENT and ME SYSTEMS	4
4.2 * Environmental conditions for ME EQUIPMENT	4
4.2.1 General	4
4.2.2 * Environmental conditions of transport and storage between uses	4
4.2.3 * Environmental operating conditions.....	5
5 * General requirements for testing ME EQUIPMENT	7
6 * Classification of ME EQUIPMENT and ME SYSTEMS	8
7 ME EQUIPMENT identification, marking and documents.....	9
7.1 * USABILITY of the ACCOMPANYING DOCUMENTS.....	9
7.2 * Additional requirements for marking of IP classification	9
7.3 ACCOMPANYING DOCUMENTS	9
7.3.1 Contact information.....	9
7.3.2 LAY OPERATOR briefing information	9
7.4 Instructions for use	10
7.4.1 Additional requirements for warning and safety notices	10
7.4.2 * Additional requirements for an electrical power source	10
7.4.3 Additional requirements for ME EQUIPMENT description	10
7.4.4 Additional requirements for ME EQUIPMENT start-up PROCEDURE	11
7.4.5 Additional requirements for operating instructions.....	11
7.4.6 Additional requirements for ME EQUIPMENT messages	11
7.4.7 * Additional requirements for cleaning, disinfection and sterilization	11
7.4.8 Additional requirements for maintenance	12

7.4.9	Additional requirements for environmental protection	12
7.4.10	Additional requirements for ME EQUIPMENT and ME SYSTEMS	12
7.5	Technical description	12
7.5.1	PERMANENTLY INSTALLED CLASS I ME EQUIPMENT	12
7.5.2	Additional requirements for professional hygienic maintenance.....	13
8	Protection against excessive temperatures and other HAZARDS	13
8.1	* Additional requirements for cleaning, disinfection of ME EQUIPMENT and ME SYSTEMS	13
8.2	* Additional requirements for sterilization of ME EQUIPMENT and ME SYSTEMS.....	13
8.3	Additional requirements for ingress of water or particulate matter into ME EQUIPMENT and ME SYSTEMS.....	13
8.3.1	* Ingress of water or particulate matter into ME EQUIPMENT.....	13
8.3.2	* Ingress of water or particulate matter into ME SYSTEMS.....	13
8.4	Additional requirements for interruption of the power supply/SUPPLY MAINS to ME EQUIPMENT and ME SYSTEM	14
8.5	Additional requirements for an INTERNAL ELECTRICAL POWER SOURCE	14
8.5.1	* Indication of state	14
8.5.2	Accessibility of small INTERNAL ELECTRICAL POWER SOURCES	15
9	Accuracy of controls and instruments and protection against hazardous outputs.....	15
10	Construction of ME EQUIPMENT	15
10.1	* Additional requirements for mechanical strength	15
10.1.1	General requirements for mechanical strength.....	15
10.1.2	* Requirements for mechanical strength for non-TRANSIT-OPERABLE ME EQUIPMENT	17
10.1.3	* Requirements for mechanical strength for TRANSIT-OPERABLE ME EQUIPMENT.....	18
10.2	Additional requirements for actuating parts of controls of ME EQUIPMENT	19
11	* Protection against strangulation or asphyxiation	20
12	Additional requirements for ELECTROMAGNETIC EMISSIONS of ME EQUIPMENT and ME SYSTEMS.....	20
13	Additional requirements for ALARM SYSTEMS of ME EQUIPMENT and ME SYSTEMS.....	20
13.1	* Additional requirement for generation of ALARM SIGNALS	20
13.2	* Additional requirement for ALARM SIGNAL volume	20
Annex A (informative)	General guidance and rationale	21
A.1	General guidance	21
A.2	Rationale for particular clauses and subclauses	22
Annex B (informative)	Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS	39
B.1	Marking on the outside of ME EQUIPMENT, ME SYSTEMS or their parts	39
B.2	ACCOMPANYING DOCUMENTS, general	39
B.3	ACCOMPANYING DOCUMENTS, instructions for use	40
B.4	ACCOMPANYING DOCUMENTS, technical description	41
Annex C (informative)	Symbols on marking.....	42
Bibliography	44
Index of defined terms used in this collateral standard		46

Figure 1 – Small finger probe Ø 5.6	8
Figure A.1 – Saturation water vapor pressure as function of temperature	25
Table 1 – Mechanical strength test applicability, non-TRANSIT-OPERABLE	16
Table 2 – Mechanical strength test applicability, TRANSIT-OPERABLE	17
Table A.1 – Saturation water vapor pressure as function of temperature	26
Table A.2 – Summary by use of HOME HEALTHCARE ENVIRONMENT ME EQUIPMENT ENCLOSURE ingress of water and particulate matter requirements	34
Table A.3 – Qualitative assessment of HOME HEALTHCARE ENVIRONMENT ME EQUIPMENT subjected to shock and vibration	35
Table B.1 – Marking on the outside of ME EQUIPMENT, ME SYSTEMS or their parts	39
Table B.2 – ACCOMPANYING DOCUMENTS, general.....	39
Table B.3 – ACCOMPANYING DOCUMENTS, instructions for use	40
Table B.4 – ACCOMPANYING DOCUMENTS, technical description	41
Table C.1 – General symbols	42

Glossary of equivalent standards

International Standards adopted in the United States may include normative references to other International Standards. AAMI maintains a current list of each International Standard that has been adopted by AAMI (and ANSI). Available on the AAMI website at the address below, this list gives the corresponding U.S. designation and level of equivalency to the International Standard.

www.aami.org/standards/glossary.pdf

Committee representation

Association for the Advancement of Medical Instrumentation Home Care Environment Committee

The adoption of IEC 60601-1-11:2015 as an American National Standard was initiated by the AAMI Home Care Environment Committee. The AAMI Home Care Environment Committee also functions as a U.S. Technical Advisory Group to the relevant work in the International Electrotechnical Commission (IEC). U.S. representatives from the AAMI Home Care Environment Committee (U.S. Sub-TAG for IEC/SC62A/JWG 6) played an active part in developing the IEC standard.

At the time this document was published, the **AAMI Home Care Environment Committee** had the following members:

Cochair	Denny Treu, BSME, NxStage Medical Inc
Members	Daniel Aghassian, Boston Scientific Corporation
	Steve Bernard, Nestle HealthCare Nutrition Inc
	Prashant Bhadri, CareFusion
	Gail Bynum, Smiths Medical
	Ron Charnock, Kwikpoint/Gaia LLC
	Anthony Ciccarello, Philips Electronics North America
	Danny Concepcion, St Joseph Hospital Renal Center
	Todd Cooper, Center for Medical Interoperability
	Conor Curtin, Fresenius Medical Care
	Reggie Cyrus,
	Susan Dorsch
	LeonardEisner, Eisner Safety Consultants
	Ila Elson, Abbott Laboratories
	Michael Friedman, Amgen Inc
	Daryle Gardner-Bonneau, Bonneau and Associates
	Donald Gillespie
	Pamela Gwynn, UL LLC
	Edward Halpern, Abbvie
	John Hedley-Whyte, Harvard University
	Dean Hooper, HE Consulting
	Byron Jacobs, Sanford Healthcare
	Mike Jaffe, Cardiorespiratory Consulting LLC
	Kevin Jensen
	Carolynn Johnson, Daedalus
	Margherita Labson, JCAHO
	Aaron Macan, Covidien
	Jim Milostan, Medical Specialties Distributors LLC
	Jan Frank Nielsen, Novo Nordisk
	Pat Patterson, Agilis Consulting Group LLC
	Yossi Pri-Paz, Laniado Hospital
	Heather Rakauskas, GFK
	Stefan Robert, Cyberonics Inc
	Adam Shames, Core Human Factors Inc
	Tom Shanks, MDVentures
	Nancy Stark, Clinical Device Group Inc
	Bob Sugarman, RCS Performance Systems Inc
	Denny Treu, NxStage Medical Inc

Bob Virag, TRIFID Medical Group LLC
Daniel Walter
Sara Wegener, Cardinal Health (MP&S)
Stephen Wilcox, Design Science Consulting
David Zinkus, Baxter Healthcare Corporation

Alternates

Andy Doering, Covidien
Kesley Gallagher, Amgen Inc
Piyatida Haerr, GFK
Eric Johnson, Nestle HealthCare Nutrition Inc
Melissa Lemke, Agilis Consulting Group LLC
Debra Milamed, Harvard University
Dave Osborn, Philips Electronics North America
Robert Parks, Daedalus
Kulwinder Plahey, Fresenius Medical Care
Diana Rivi, FDA/CDRH
Allison Stochlic, UL LLC
Chandresh Thakur, CareFusion
Steven White, NxStage Medical Inc
Joyce Young-Stewart, Baxter Healthcare Corporation

NOTE—Participation by federal agency representatives in the development of this document does not constitute endorsement by the federal government or any of its agencies.

Background of ANSI/AAMI adoption of IEC 60601-1-11:2015

As indicated in the foreword to the main body of this document (page xi), the International Electrotechnical Commission (IEC) is a worldwide organization of national electrotechnical committees. The United States is one of the IEC members that took an active role in the development of this standard, which was developed by IEC Technical Subcommittee 62A, Common aspects of electrical equipment used in medical practice.

U.S. participation in this IEC SC is organized through the U.S. National Committee (USNC) and the U.S. Technical Advisory Group for IEC/SC 62A, administered by the AAMI.

The AAMI Home Care Environment Committee initiated the U.S. adoption of IEC 60601-1-11:2015.

The major differences between ANSI/AAMI HA60601-1-11 and IEC 60601-1-11:2015 is the removal of elder care facilities as a home environment; the clarification of alternative life- supporting methods as a PRIMARY OPERATING FUNCTION when applying the USABILITY ENGINEERING PROCESS; the modification of a compliance statement to include inspection of the usability engineering file; and the inclusion of PRIMARY OPERATING FUNCTION in the glossary.

ANSI/AAMI HA60601-1-11:2015 was approved by the American National Standards Institute (ANSI) on 25 August 2015.

AAMI and ANSI procedures require that standards be reviewed every five years and, if necessary, revised to reflect technological advances that may have occurred since publication.

AAMI (and ANSI) have adopted other international standards. See the Glossary of Equivalent Standards for a list of international standards adopted by AAMI, which gives the corresponding U.S. designation and the level of equivalency with the international standard.

As used within the context of this document, “shall” indicates requirements strictly to be followed to conform to the recommended practice. “Should” indicates that among several possibilities, one is recommended as particularly suitable, without mentioning or excluding others, or that a certain course of action is preferred but not necessarily required, or that (in the negative form) a certain possibility or course of action should be avoided but is not prohibited.

“May” is used to indicate that a course of action is permissible within the limits of the recommended practice. “Can” is used as a statement of possibility and capability. Finally, “must” is used only to describe “unavoidable” situations, including those mandated by government regulation.

The concepts incorporated in this standard should not be considered inflexible or static. This standard, like any other, must be reviewed and updated periodically to assimilate progressive technological developments. To remain relevant, it must be modified as technological advances are made and as new data come to light.

Suggestions for improving this standard are invited. Comments and suggested revisions should be sent to Standards Department, AAMI, 4301 N Fairfax Drive, Suite 301, Arlington, VA 22203-1633.

Foreword

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-1-11 has been prepared by a joint working group of IEC subcommittee 62A: Common aspects of electrical equipment used in medical practice of IEC technical committee 62: Electrical equipment in medical practice and ISO subcommittee SC3: Lung ventilators and related devices, of ISO technical committee 121: Anaesthetic and respiratory equipment.

It is published as a double logo standard.

This second edition constitutes a collateral standard to IEC 60601-1 (third edition, including Amendment 1): *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*, hereafter referred to as the general standard.

This second edition cancels and replaces the first edition of IEC 60601-1-11, published in 2010, and constitutes a technical revision.

The most significant changes with respect to the previous edition include the following modifications:

- correction of test method for relative humidity control at temperatures above 35 °C;
- redrafting of subclauses that altered instead of adding to the general standard or other collateral standards; and

- harmonizing with the changes to the amendments to the general standard and other collateral standards.

The text of this collateral standard is based on the following documents:

FDIS	Report on voting
62A/959/FDIS	62A/978/RVD

Full information on the voting for the approval of this collateral standard can be found in the report on voting indicated in the above table. In ISO, the standard has been approved by 17 P-members out of 17 having cast a vote.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In the IEC 60601 series of publications, collateral standards specify general requirements for safety applicable to:

- a subgroup of MEDICAL ELECTRICAL EQUIPMENT (e.g. radiological equipment); or
- a specific characteristic of all MEDICAL ELECTRICAL EQUIPMENT, not fully addressed in the general standard (e.g. ALARM SYSTEMS).

In this collateral standard, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type*;
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS COLLATERAL STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this collateral standard, the term

- “clause” means one of the numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.3.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this collateral standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

Clauses, subclauses and definitions for which a rationale is provided in informative Annex A are marked with an asterisk (*).

A list of all parts of the IEC 60601 series, published under the general title: *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE—The attention of Member Bodies and National Committees is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised ISO or IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for mandatory implementation nationally not earlier than 3 years from the date of publication.

U.S. Deviations to IEC 60601-1-11:2015

As part of an effort to harmonize standards throughout an increasing global industry, the AAMI Home Use Committee voted in 2015 to adopt IEC 60601-1-11:2015, *Medical electrical equipment – Part 1-11: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment*. The AAMI Home Use Committee also agreed that a number of U.S. deviations to the IEC standard would improve the document.

Deviations are listed below. A rationale for each change has also been provided by the committee. Within the document, deletions are indicated by ~~strike through~~ and additions are indicated by underline.

Terms and definitions

1. NOTE 2 in 3.1 modified.

Rationale: Elder care facilities in the U.S. are not considered a 'home environment'.

8.4

2. Additional sentence added to paragraph preceding NOTE 2.

Rationale: Text added to reinforce that the usability engineering process is necessary for validation of the instructions for use.

10.2

3. Compliance statement modified.

Rationale: Inspection of the usability engineering file added to reinforce that the usability engineering process is necessary for validation of the instructions for use.

Index of defined terms

4. PRIMARY OPERATING FUNCTION added to glossary list

Rationale: The additional text in 8.4 necessitates adding the term to the index.

Introduction

Medical practice is increasingly using MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS for monitoring, treatment or diagnosis of PATIENTS in the HOME HEALTHCARE ENVIRONMENT (see 3.1). The safety of MEDICAL ELECTRICAL EQUIPMENT in this uncontrolled environment with regard to the electrical installation and its related safety and protection means is a cause for concern.

The potential lack of training of the LAY OPERATOR and possibly of those supervising the use of the MEDICAL ELECTRICAL EQUIPMENT OR MEDICAL ELECTRICAL SYSTEM and their level of education need to be addressed in the development of the ACCOMPANYING DOCUMENTS and in the relevant marking on the equipment itself so that this material can be understood. This collateral standard gives special guidance on how this should be addressed in the instructions for use.

This collateral standard was developed with contributions from clinicians, engineers and regulators. The terminology, requirements, general recommendations and guidance of this collateral standard are intended to be useful for MANUFACTURERS of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS and for technical committees responsible for the development of particular standards.

MEDICAL ELECTRICAL EQUIPMENT — Part 1-11: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment

1 Scope, object and related standards

1.1 * Scope

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS for use in the HOME HEALTHCARE ENVIRONMENT, as defined in 3.1, and specified by the MANUFACTURER in the instructions for use. This International Standard applies regardless of whether the ME EQUIPMENT or ME SYSTEM is intended for use by a LAY OPERATOR or by trained healthcare personnel.

The HOME HEALTHCARE ENVIRONMENT includes:

- the dwelling place in which a PATIENT lives;
- other places where PATIENTS are present both indoors and outdoors, excluding professional healthcare facility environments where OPERATORS with medical training are continually available when PATIENTS are present.

This International Standard does not apply to ME EQUIPMENT and ME SYSTEMS intended solely for use in the EMERGENCY MEDICAL SERVICES ENVIRONMENT, covered by IEC 60601-1-12 or solely for use in professional healthcare facilities covered by IEC 60601-1 without the additions of IEC 60601-1-12 or this collateral standard. Nonetheless, ME EQUIPMENT or ME SYSTEMS can be intended for multiple use environments, and as such, if also intended for use in the HOME HEALTHCARE ENVIRONMENT, are within the scope of this standard.

EXAMPLE ME EQUIPMENT or ME SYSTEMS intended for both the HOME HEALTHCARE ENVIRONMENT and the professional healthcare facility environment.

NOTE HOME HEALTHCARE ENVIRONMENT ME EQUIPMENT and ME SYSTEMS can frequently be used in locations with unreliable electrical sources and poor electrical grounding.

1.2 Object

The object of this collateral standard is to specify general requirements that are in addition to those of the general standard and to serve as the basis for particular standards.

1.3 Related standards

1.3.1 IEC 60601-1

For ME EQUIPMENT and ME SYSTEMS, this collateral standard complements IEC 60601-1.

When referring to IEC 60601-1 or to this collateral standard, either individually or in combination, the following conventions are used:

- "the general standard" designates IEC 60601-1 alone;
- "this collateral standard" designates IEC 60601-1-11 alone;
- "this standard" designates the combination of the general standard and this collateral standard.