

## Contents

	Page
Foreword .....	ii
<b>1</b> Scope .....	1
<b>2</b> Normative references .....	1
<b>3</b> Definitions .....	2
<b>4</b> Requirements .....	2
4.1 General .....	2
4.2 Type testing (certification) .....	2
4.3 Verification (instrument compliance) .....	3
4.4 Construction and function .....	3
<b>5</b> Test methods .....	4
<b>6</b> Accompanying documents .....	4
<b>7</b> Additional documents .....	4
<b>8</b> Marking .....	5
<b>Table</b>	
<b>1</b> Requirements for paired testing of reference to test tonometers .....	2
<b>Annexes</b>	
<b>A</b> The reference tonometer and method for determining reference IOP .....	6
<b>B</b> Design compliance testing .....	13
<b>C</b> Example of a test certificate .....	17

**Foreword** (This foreword is not part of American National Standard ANSI Z80.10-2018.)

This American National Standard covers tonometers specifically intended for routine clinical use in the estimation of intraocular pressure for the detection, diagnosis, and management of ocular abnormalities and excludes uses such as monitoring induced high intraocular pressure for refractive surgery.

ANSI Z80.10-2014 was adapted by a group of experts within the ANSI Ophthalmic Instruments Subcommittee under the chair of Charles E. Campbell. It is a performance standard.

This standard defines the tolerable range of intraocular pressure (IOP) readings in comparison to measurements made by a reference applanation tonometer specified to be the Goldmann Tonometer. Normative references have been updated. Allowance is made in this standard for a Goldmann-equivalent measurement, taken with an instrument that meets requirements for a test tonometer when compared to the Goldmann reference tonometer. Such an instrument may present intraocular pressure data in formats other than Goldmann. This standard clarifies that a new tonometer that meets all of the specifications and requirements of annex A is qualified as a reference tonometer, and, therefore, it does not require the clinical validation described in annex B and clause 4.2.1. In Annex B, Clause B.2.3 was added to add the requirement that the same examiner shall make the measurements on an eye with the tested and the reference tonometer. The final paragraph of Clause B.5.3 has been removed so that repeated measurements on the same eye for eyes with pressure in excess of 23 mmHg are no longer allowed as means to obtain sufficient measurements in this pressure range group. The minimum number of eyes to be tested specified in B.5.3 was corrected to read 120. It was clarified that additional eyes in each pressure group are only needed for instruments that are claimed to measure eyes with high astigmatism.

Annexes A and B are normative and are considered part of the standard. Annex C is informative and is not considered part of the standard.

Suggestions for improvement of this standard will be welcome. They should be sent to the Vision Council, 225 Reinekers Lane, Suite 700, Alexandria, VA 22314.

This standard was processed and approved for submittal to ANSI by the Accredited Standards Committee on Ophthalmic Optics, Z80. Committee approval of this standard does not necessarily imply that all committee members voted for its approval. At the time of approval of this standard, the Z80 Committee consisted of the following members:

Carl Tubbs, M.D., Chairman  
Neal Roche, Vice-Chairman  
William Benjamin, O.D., Secretary  
Michael Vitale, Secretariat

<i>Organization Represented</i>	<i>Name of Representative</i>
Advanced Medical Technology Association.....	Michael Pflieger
American Academy of Ophthalmology.....	Carl Tubbs
American Academy of Optometry.....	David Loshin
American Ceramic Society.....	Lyle Rubin
American Glaucoma Society.....	Steven Gedde
American Optometric Association.....	William Benjamin
American Society of Cataract and Refractive Surgery.....	Stephen Klyce
Contact Lens Institute.....	Stan Rogaski
Contact Lens Manufacturers Association.....	Martin Dalsing
Cornea Society.....	Michael Belin
Department of Veterans Affairs.....	John Townsend
Food & Drug Administration CDRH/Division.....	Don Calogero
Individual.....	Ralph Stone
Johnson & Johnson Vision.....	Priya Janakiraman
National Association of Optometrists & Opticians.....	Nick Mileti
Optical Laboratory Association.....	Steve Sutherlin
Opticians Association of America.....	Tom Hicks
Sunglass Association of America.....	Tibor Gross
The Vision Council.....	Michael Vitale
ISO TC 172/SC7.....	Michael Vitale

The members of the subcommittee on Ophthalmic Instruments who contributed to this standard are:

Charles E. Campbell, Chair  
William L. Brown, O.D., Ph.D,  
Karl Citek, O.D. Ph.D.  
Bruce Drum, Ph.D.  
Stephen Klyce, Ph.D.  
Dexiu Shi, Ph.D.  
David Sliney, Ph.D.  
Robert Rosenberg, O.D.