

Australian/New Zealand Standard™

Acoustics—Audiometric test methods

Part 3: Speech audiometry



AS/NZS ISO 8253.3:2019

This Joint Australian/New Zealand Standard™ was prepared by Joint Technical Committee AV-003, Acoustics Human Effects. It was approved on behalf of the Council of Standards Australia on 29 January 2019 and by the New Zealand Standards Approval Board on 30 January 2019.

This Standard was published on 11 March 2019.

The following are represented on Committee AV-003:

- Accident Compensation Corporation (New Zealand)
- Association of Australasian Acoustical Consultants
- Australian Acoustical Society
- Australian Chamber of Commerce and Industry
- Australian Council of Trade Unions
- Australian Hearing
- Engineers Australia
- Ministry of Health (NZ)
- New Zealand Audiological Society
- Worksafe Division, Department of Commerce, Western Australia

This Standard was issued in draft form for comment as DR AS ISO 8253.3:2018.

Keeping Standards up-to-date

Ensure you have the latest versions of our publications and keep up-to-date about Amendments, Rulings, Withdrawals, and new projects by visiting:

www.standards.org.au

www.standards.govt.nz

ISBN 978 1 76072 390 3

Australian/New Zealand Standard™

Acoustics—Audiometric test methods

Part 3: Speech audiometry

First published as AS ISO 8253.3—2009.
Jointly revised and redesignated as AS/NZS ISO 8253.3:2019.

COPYRIGHT

© ISO 2019 — All rights reserved

© Standards Australia Limited/the Crown in right of New Zealand, administered by the New Zealand Standards Executive 2019

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher, unless otherwise permitted under the Copyright Act 1968 (Cth) or the Copyright Act 1994 (New Zealand).

Preface

This Standard was prepared by the Joint Standards Australia/Standards New Zealand Committee AV-003, Acoustics Human Effects to supersede AS ISO 8253.3—2009, *Acoustics—Audiometric test methods, Part 3: Speech audiometry*.

The objective of this Standard is to specify basic methods for speech recognition tests for audiological applications.

In order to ensure minimum requirements of precision and comparability between different test procedures including speech recognition tests in different languages, this Standard specifies requirements for the composition, validation and evaluation of speech test materials, and the realization of speech recognition tests. This Standard does not specify the contents of the speech material because of the variety of languages. Furthermore, this Standard also specifies the determination of reference values and fulfilment requirements for the realization and manner of presentation.

This Standard is identical with, and has been reproduced from, ISO 8253-3:2012, *Acoustics — Audiometric test methods — Part 3: Speech audiometry*.

As this document has been reproduced from an International Standard, the following applies:

- (a) In the source text “this part of ISO 8253” should read “this Australian/New Zealand Standard”.
- (b) A full point substitutes for a comma when referring to a decimal marker.

Australian or Australian/New Zealand Standards that are identical adoptions of international normative references may be used interchangeably. Refer to the online catalogue for information on specific Standards.

The term “informative” is used in Standards to define the application of the annexes to which it applies. An “informative” annex is only for information and guidance.

Contents

Preface	ii
Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Requirements for recording of speech material	5
4.1 General requirements	5
4.2 Reference recording	5
4.3 Recording environment	6
4.4 Frequency response of recording equipment	6
4.5 Interval between successive test items	6
4.6 Levels	6
4.6.1 Speech material	6
4.6.2 Calibration signal	6
4.7 Signal-to-noise ratio	6
4.8 Phonemic balance across test lists	6
4.9 Perceptual balance across test lists	7
5 Validation of speech material recordings	7
5.1 General	7
5.2 Determination of reference speech recognition curve	7
5.3 Determination of perceptual equivalence of test lists	8
5.4 Determination of repeatability of results (average test-retest reliability)	8
5.5 Documentation	9
6 Requirements for speech audiometry	9
7 Ambient sound pressure levels in test room for speech audiometry	10
8 Sound field speech audiometry	10
9 Preparation and instruction of test subject	10
9.1 General	10
9.2 Preparation of test subject	10
9.3 Instruction of test subject	10
10 Subject's response mode	11
11 Determination of speech detection threshold level	11
12 Determination of speech recognition threshold level	12
12.1 General	12
12.2 Descending procedure using 5 dB steps	12
12.3 Alternative descending procedure	13
12.4 Adaptive procedure using fixed step sizes	13
12.5 Other adaptive procedure	14
13 Determination of speech recognition scores	14
14 Contralateral masking	15
15 Speech audiometry with competing sound	15
15.1 Type of competing sound	15
15.2 Presentation of competing sound	16
15.3 Speech and competing sound levels	16
15.4 Test procedure	16
15.4.1 General	16

15.4.2	Speech recognition score at fixed speech-to-noise ratio.....	16
15.4.3	Speech recognition threshold with a competing sound.....	16
16	Format of speech audiogram.....	18
17	Measurement uncertainty.....	18
18	Maintenance and calibration of equipment.....	18
18.1	General.....	18
18.2	Intervals between tests.....	19
18.3	Stage A: Routine checking and listening tests.....	19
18.4	Stage B: Periodic electroacoustic tests.....	20
18.5	Stage C: Basic calibration tests.....	20
Annex A	(informative) Example of speech materials.....	21
Annex B	(informative) Examples of competing sound conditions.....	22
Annex C	(informative) Typical results.....	23
Annex D	(informative) Optimization of perceptual balance of test lists.....	25
Annex E	(informative) Measurement uncertainty.....	26
Bibliography	31

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 8253-3 was prepared by Technical Committee ISO/TC 43, *Acoustics*.

This second edition cancels and replaces the first edition (ISO 8253-3:1996), which has been technically revised.

ISO 8253 consists of the following parts, under the general title *Acoustics — Audiometric test methods*:

- *Part 1: Pure-tone air and bone conduction threshold audiometry*
- *Part 2: Sound field audiometry with pure-tone and narrow-band test signals*
- *Part 3: Speech audiometry*

Introduction

Speech audiometry is used for the assessment of hearing in connection with diagnostic evaluation and audiological rehabilitation.

The results of speech audiometry depend on the speech material and test method used. This part of ISO 8253 sets conditions for speech materials in order to assure minimum requirements of precision and comparability between different tests using different speech materials including materials in different languages. It also specifies procedures to be used when testing speech recognition.

Australian/New Zealand Standard

Acoustics—Audiometric test methods

Part 3: Speech audiometry

1 Scope

This part of ISO 8253 specifies basic methods for speech recognition tests for audiological applications.

In order to ensure minimum requirements of precision and comparability between different test procedures including speech recognition tests in different languages, this part of ISO 8253 specifies requirements for the composition, validation and evaluation of speech test materials, and the realization of speech recognition tests. This part of ISO 8253 does not specify the contents of the speech material because of the variety of languages.

Furthermore, this part of ISO 8253 also specifies the determination of reference values and fulfilment requirements for the realization and manner of presentation.

This part of ISO 8253 specifies procedures and requirements for speech audiometry with the recorded test material being presented by air conduction through an earphone, or from a loudspeaker for sound field audiometry. Methods for using noise either for masking the non-test ear or as a competing sound are described.

Some test subjects, for example children, can require amended test procedures not specified in this part of ISO 8253.

Specialized tests such as those used for evaluating directional hearing and dichotic hearing are outside the scope of this part of ISO 8253.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 266, *Acoustics — Preferred frequencies*

ISO 8253-1, *Acoustics — Audiometric test methods — Part 1: Pure-tone air and bone conduction audiometry*

ISO 8253-2, *Acoustics — Audiometric test methods — Part 2: Sound field audiometry with pure-tone and narrow-band test signals*

ISO/IEC Guide 98-3, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

IEC 60645-1, *Electroacoustics — Audiological equipment — Part 1: Pure-tone audiometers*

IEC 60645-2:1993, *Audiometers — Part 2: Equipment for speech audiometry*

IEC 61672-1, *Electroacoustics — Sound level meters — Part 1: Specifications*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 8253-1 and ISO 8253-2 and the following apply.