



## **Medical electrical equipment**

### **Part 2.54: Particular requirements for the basic safety and essential performance of X-ray equipment for radiography and radioscopy (IEC 60601- 2-54:2018 (ED. 1.2), MOD)**

AS 60601.2.54:2018

This Australian Standard® was prepared by the Australian members of the Joint Technical Committee HE-003, Medical Electrical Equipment. It was approved on behalf of the Council of Standards Australia on 11 September 2018.

This Standard was published on 27 September 2018.

The following are represented on Committee HE-003:

- Australasian College of Physical Scientists and Engineers in Medicine
- Australian and New Zealand College of Anaesthetists
- Australian Chamber of Commerce and Industry
- Australian College of Perioperative Nurses
- Australian Radiation Protection and Nuclear Safety Agency
- Australian Society of Anaesthetists
- Certification Body Australia (Certification Interests Australia)
- College of Biomedical Engineering, Engineers Australia
- Department of Defence (Australian Government)
- Engineers Australia
- Medical Technology Association of Australia
- Therapeutic Goods Administration (Australian Government)

This Standard was issued in draft form for comment as DR AS/NZS 60601.2.54:2018.

### **Keeping Standards up-to-date**

Ensure you have the latest versions of our publications and keep up-to-date about Amendments, Rulings, Withdrawals, and new projects by visiting:

[www.standards.org.au](http://www.standards.org.au)

[www.saiglobal.com](http://www.saiglobal.com) (sales and distribution)

ISBN 978 1 76072 185 5



## **Medical electrical equipment**

# **Part 2.54: Particular requirements for the basic safety and essential performance of X-ray equipment for radiography and radioscopy (IEC 60601- 2-54:2018 (ED. 1.2), MOD)**

Originated as AS/NZS 3200.2.7:1994.  
Second edition 1999.  
Revised and redesignated as AS 60601.2.54:2018.

### **COPYRIGHT**

© IEC 2018 — All rights reserved  
© Standards Australia Limited 2018

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher, unless otherwise permitted under the Copyright Act 1968 (Cth).

Published by SAI Global Pty Limited under licence from Standards Australia Limited, GPO Box 476, Sydney, NSW 2001.

## Preface

This Standard was prepared by the Australian members of the Joint Standards Australia/Standards New Zealand Committee HE-003, Medical Electrical Equipment to supersede AS/NZS 3200.2.7:1999, *Medical electrical equipment Part 2.7: Particular requirements for safety—High voltage generators of diagnostic X-ray generators*.

After consultation with stakeholders in both countries, Standards Australia and Standards New Zealand decided to develop this Standard as an Australian Standard rather than an Australian/New Zealand Standard.

The objective of this Standard is to specify particular basic safety and essential performance requirements for medical electrical (ME) equipment and ME systems for radiography and radioscopy. The minimum safety requirements specified in this Standard are considered to provide for a practical degree of safety in the operation of ME equipment for radiography and radioscopy.

This Standard amends and supplements the general Standard IEC 60601-1, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance*.

This Standard is an adoption with national modifications, and has been reproduced from, IEC 60601-2-54:2018 (ED. 1.2), *Medical electrical equipment — Part 2-54: Particular requirements for the basic safety and essential performance of X-ray equipment for radiography and radioscopy*. [Appendix ZZ](#) lists the variations for the application of this Standard in Australia and New Zealand.

As this document has been reproduced from an International Standard, the following applies:

- (a) In the source text “this particular Standard” should read “this Australian Standard”.
- (b) A full point substitutes for a comma when referring to a decimal marker.

Australian or Australian/New Zealand Standards that are identical adoptions of international normative references may be used interchangeably. Refer to the online catalogue for information on specific Standards.

The terms “normative” and “informative” are used in Standards to define the application of the appendices or annexes to which they apply. A “normative” appendix or annex is an integral part of a Standard, whereas an “informative” appendix or annex is only for information and guidance.

## CONTENTS

|  |    |
|--|----|
| FOREWORD.....  | 4  |
| INTRODUCTION.....  | 7  |
| INTRODUCTION TO AMENDMENT 1.....   | 7  |
| INTRODUCTION TO AMENDMENT 2.....   | 7  |
| 201.1 Scope, object and related standards .....  | 8  |
| 201.2 Normative references .....   | 9  |
| 201.3 Terms and definitions .....  | 10 |
| 201.4 General requirements .....   | 12 |
| 201.5 General requirements for testing of ME EQUIPMENT .....   | 13 |
| 201.6 Classification of ME EQUIPMENT and ME SYSTEMS .....  | 13 |
| 201.7 ME EQUIPMENT identification, marking and documents .....   | 13 |
| 201.8 Protection against electrical HAZARDS from ME EQUIPMENT .....  | 16 |
| 201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS.....  | 19 |
| 201.10 Protection against unwanted and excessive radiation HAZARDS .....   | 23 |
| 201.11 Protection against excessive temperatures and other HAZARDS .....   | 23 |
| 201.12 Accuracy of controls and instruments and protection against hazardous<br>outputs .....                                      | 24 |
| 201.13 HAZARDOUS SITUATIONS and fault conditions .....   | 24 |
| 201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....   | 24 |
| 201.15 Construction of ME EQUIPMENT .....  | 24 |
| 201.16 ME SYSTEMS.....   | 24 |
| 201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS .....  | 25 |
| 202 Electromagnetic compatibility – Requirements and tests .....   | 25 |
| 203 Radiation protection in diagnostic X-ray equipment.....  | 25 |
| Annexes .....  | 64 |
| Annex C (informative) Guide to marking and labelling requirements for ME EQUIPMENT<br>and ME SYSTEMS.....                          | 65 |
| Annex AA (informative) Particular guidance and rationale .....   | 67 |
| Bibliography .....   | 72 |
| Index of defined terms used in this particular standard .....  | 74 |
| Figure 203.101 – Zone of EXTRA-FOCAL RADIATION .....   | 46 |
| Figure 203.102 – Discrepancies in covering the IMAGE RECEPTION AREA .....  | 48 |
| Figure 203.103 – Discrepancies in visual indication of the X-RAY FIELD .....   | 52 |
| Figure 203.104 – Testing for STRAY RADIATION (X-RAY BEAM horizontal with X-RAY<br>SOURCE ASSEMBLY below the PATIENT SUPPORT) ..... | 61 |
| Figure 203.105 – Testing for STRAY RADIATION (X-RAY BEAM vertical with X-RAY SOURCE<br>ASSEMBLY below the PATIENT SUPPORT).....    | 61 |
| Figure 203.106 – Testing for STRAY RADIATION (X-RAY BEAM horizontal with X-RAY<br>SOURCE ASSEMBLY above the PATIENT SUPPORT).....  | 62 |
| Figure 203.107 – Testing for STRAY RADIATION (X-RAY BEAM vertical with X-RAY SOURCE<br>ASSEMBLY above the PATIENT SUPPORT) .....   | 63 |

|  |    |
|--|----|
| Table 201.101 – Distributed potential ESSENTIAL PERFORMANCE requirements.....    | 12 |
| Table 203.101 – Tests for verifying reproducibility and linearity.....           | 33 |
| Table 203.102 – Loadings for testing AUTOMATIC EXPOSURE CONTROLS .....           | 35 |
| Table 203.103 – ATTENUATION for the measurement of AIR KERMA.....                | 37 |
| Table 203.104 – ATTENUATION EQUIVALENT of items in the X-RAY BEAM .....          | 54 |
| Table 203.105 – Application categories .....                                     | 56 |
| Table 203.106 – Requirements for PRIMARY PROTECTIVE SHIELDING .....              | 57 |
| Table 203.107 – STRAY RADIATION in SIGNIFICANT ZONES OF OCCUPANCY .....          | 59 |
| Table 201.C.101 – Marking on the outside of ME EQUIPMENT or its parts .....      | 65 |
| Table 201.C.102 – Subclauses requiring statements in ACCOMPANYING DOCUMENTS..... | 65 |

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

### MEDICAL ELECTRICAL EQUIPMENT –

#### **Part 2-54: Particular requirements for the basic safety and essential performance of X-ray equipment for radiography and radioscopy**

#### FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

#### **DISCLAIMER**

**This Consolidated version is not an official IEC Standard and has been prepared for user convenience. Only the current versions of the standard and its amendment(s) are to be considered the official documents.**

**This Consolidated version of IEC 60601-2-54 bears the edition number 1.2. It consists of the first edition (2009-06) [documents 62B/735/FDIS and 62B/750/RVD], its amendment 1 (2015-04) [documents 62B/929/CDV and 62B/956/RVC], its amendment 2 (2018-06) [documents 62B/1089/FDIS and 62B/1097/RVD] and its corrigenda 1 (March 2010) and 2 (June 2011). The technical content is identical to the base edition and its amendments.**

**This Final version does not show where the technical content is modified by amendments 1 and 2. A separate Redline version with all changes highlighted is available in this publication.**

International standard IEC 60601-2-54 has been prepared by IEC subcommittee 62B: Diagnostic imaging equipment, of IEC technical committee 62: Electrical equipment in medical practice.

IEC 60601-2-54 has been developed for use with the third edition of IEC 60601-1 (2005). It replaces and supersedes IEC 60601-2-7 and IEC 60601-2-32, as well as parts of IEC 60601-2-28:1993, all of which were developed to amend earlier editions of IEC 60601-1 and consequently no longer apply to this particular standard.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (\*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of the base publication and its amendments will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

## INTRODUCTION

This particular standard has been prepared to provide, based on IEC 60601-1:2005 (third edition) and its collaterals, a complete set of safety requirements for ME EQUIPMENT for RADIOGRAPHY and RADIOSCOPY. While the previously existing standards for such equipment were dedicated to components and subsystems, this particular standard addresses the system level of X-RAY EQUIPMENT, which consists of a combination of an X-RAY GENERATOR, ASSOCIATED EQUIPMENT and ACCESSORIES. Component functions are addressed as far as necessary.

The minimum safety requirements specified in this particular standard are considered to provide for a practical degree of safety in the operation of ME EQUIPMENT for RADIOGRAPHY and RADIOSCOPY. Requirements for additional provisions for ME EQUIPMENT for interventional applications are covered by IEC 60601-2-43.

## INTRODUCTION TO AMENDMENT 1

The purpose of this first amendment to IEC 60601-2-54:2009 is to introduce changes to reference the first amendment (2012) to IEC 60601-1:2005. As neither IEC 60601-2-54:2009 nor this amendment refers to specific elements of IEC 60601-1-2, the introduction of a dated reference to the latter document has been removed. In addition, a number of technical errors have been corrected.

## INTRODUCTION TO AMENDMENT 2

The purpose of this second amendment to IEC 60601-2-54:2009 is to introduce changes which take the current state of the art into account. Therefore, X-RAY EQUIPMENT specified for DIRECT RADIOSCOPY is no longer in the scope of this document. The normative references were also updated in this amendment, and editorial clarifications and new terms and definitions were added. Provisions for QUALITY CONTROL PROCEDURES to be recommended by the MANUFACTURER are emphasized. Specific attention is paid to EXAMINATION PROTOCOLS in a new subclause which differentiate between adult and paediatric applications, in particular for X-RAY EQUIPMENT without an AUTOMATIC CONTROL SYSTEM. In addition, fixed periods for termination of LOADING after release of the RADIATION control by the OPERATOR are stipulated for RADIOSCOPY.

A new subclause on electronic documentation of EXAMINATION PROTOCOLS is introduced. It recommends providing access to electronic documentation containing relevant parameters of the PRE-PROGRAMMED EXAMINATION PROTOCOL. In another new subclause, the creation of basic documentation of the RADIATION DOSE STRUCTURED REPORT (RDSR) according to IEC 61910-1 is recommended. Furthermore, the subclause describing the LAST IMAGE HOLD RADIOGRAM has been revised and requires that the last image in RADIOSCOPY be displayed rather than provide just a means to display it.

This amendment recommends providing a graphical DISPLAY of the position of the BEAM LIMITING DEVICE blades on the IMAGE DISPLAY DEVICE in the subclause "Indication on the X-RAY EQUIPMENT".

Finally, the requirement for providing means to limit the FOCAL SPOT TO SKIN DISTANCES for radiosopic X-RAY EQUIPMENT differentiates between MOBILE and FIXED EQUIPMENT and extends, in the latter case, the minimum distance in possible clinical applications.

## MEDICAL ELECTRICAL EQUIPMENT –

### Part 2-54: Particular requirements for the basic safety and essential performance of X-ray equipment for radiography and radioscopy

#### 201.1 Scope, object and related standards

Clause 1 of the general standard<sup>1)</sup> applies, except as follows:

##### 201.1.1 Scope

*Replacement:*

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of ME EQUIPMENT and ME SYSTEMS intended to be used for projection RADIOGRAPHY and INDIRECT RADIOSCOPY. IEC 60601-2-43 applies to ME EQUIPMENT and ME SYSTEMS intended to be used for interventional applications and refers to applicable requirements in this particular standard.

ME EQUIPMENT and ME SYSTEMS intended to be used for bone or tissue absorption densitometry, computed tomography, mammography or dental or radiotherapy applications are excluded from the scope of this International Standard. The scope of this International Standard also excludes radiotherapy simulators.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

##### 201.1.2 Object

*Replacement:*

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for ME EQUIPMENT and ME SYSTEMS for RADIOGRAPHY and RADIOSCOPY.

##### 201.1.3 Collateral standards

*Addition:*

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

IEC 60601-1-2 and IEC 60601-1-3 apply as modified in Clauses 202 and 203 respectively. IEC 60601-1-8, IEC 60601-1-9, IEC 60601-1-10, IEC 60601-1-11 and IEC 60601-1-12 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

NOTE OPERATORS of X-RAY EQUIPMENT are used to audible signals as required in this particular standard rather than to the concepts of IEC 60601-1-8. Therefore IEC 60601-1-8 does not apply.

---

<sup>1)</sup> The general standard is IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*