

Australian Standard™

Laboratory glassware—Bottles

Part 1: Screw-neck bottles



This Australian Standard was prepared by Committee CH-001, Laboratory Glassware and Related Apparatus. It was approved on behalf of the Council of Standards Australia on 31 October 2001 and published on 4 January 2002.

The following interests are represented on Committee CH-001:

Australian Chamber of Commerce and Industry
National Association of Testing Authorities, Australia
National Standards Commission
The Royal Australian Chemical Institute
Royal College of Pathologists of Australasia
Science Industry Australia Inc
The University of New South Wales
The University of Sydney

Keeping Standards up-to-date

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about Standards can be found by visiting the Standards Australia web site at www.standards.com.au and looking up the relevant Standard in the on-line catalogue.

Alternatively, the printed Catalogue provides information current at 1 January each year, and the monthly magazine, *The Australian Standard*, has a full listing of revisions and amendments published each month.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at mail@standards.com.au, or write to the Chief Executive, Standards Australia International Ltd, GPO Box 5420, Sydney, NSW 2001.

Australian Standard™

Laboratory glassware—Bottles

Part 1: Screw-neck bottles

First published as AS 4716.1—2002.

COPYRIGHT

© Standards Australia International

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia International Ltd
GPO Box 5420, Sydney, NSW 2001, Australia

ISBN 0 7337 4178 9

PREFACE

This Standard was prepared by the Standards Australia Committee CH-001, Laboratory Glassware and Related Apparatus. This Standard is identical with and has been reproduced from ISO 4796-1:2000, *Laboratory glassware—Bottles, Part 1: Screw-neck bottles*.

ISO 4796 consists of the following parts under the general title, *Laboratory glassware—Bottles*:

Part 1: Screw-neck bottles

Part 2: Conical neck bottles

Part 3: Aspirator bottles

The objective of this Standard is to specify a series of screw-neck bottles suitable for the storage of fluid liquid and solid chemicals and reagents in general laboratory use.

As this Standard is reproduced from an International Standard, the following applies:

- (a) Its number appears on the cover and title page while the International Standard number appears only on the cover.
- (b) In the source text, 'this part of ISO 4796' should read 'this Australian Standard'.
- (c) Figure 2 on page 3 refers to 25 L in the caption. The maximum capacity in the rest of the document is 20 L, so the value should be changed accordingly.
- (d) A full point substitutes for a comma when referring to a decimal marker.
- (e) Substitute 'L' for 'l' and 'mL' for 'ml' wherever it appears.
- (f) The ISO document listed as a normative reference in Clause 2 has not been adopted as an Australian Standard.

CONTENTS

	<i>Page</i>
1 SCOPE.....	1
2 NORMATIVE REFERENCE	1
3 CAPACITIES.....	1
4 DIMENSIONS	1
5 CONSTRUCTION.....	2
6 DESIGNATION	4
7 MARKING.....	4

NOTES

AUSTRALIAN STANDARD

Laboratory glassware — Bottles —

Part 1: Screw-neck bottles

1 Scope

This part of ISO 4796 specifies a series of screw-neck bottles suitable for the storage of fluid liquid and solid chemicals and reagents in general laboratory use. These bottles are also suitable for the preparation and storage of microbiological growth media.

2 Normative reference

The following normative document contains provisions which, through reference in this text, constitute provisions of this part of ISO 4796. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this part of ISO 4796 are encouraged to investigate the possibility of applying the most recent edition of the normative document indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 3585:1998, *Borosilicate glass 3.3 — Properties*.

3 Capacities

3.1 The nominal capacities of screw-neck bottles shall be chosen from the following series:

25 ml — 50 ml — 100 ml — 250 ml and 500 ml;

1 l — 2 l — 5 l — 10 l — 15 l and 20 l.

3.2 The nominal capacity of a bottle indicates the quantity of liquid which a bottle of average wall thickness shall contain when the bottle is filled to the turn of the shoulder.

3.3 The design of the bottle shall be such that the total capacity to the base of the neck shall be approximately 15 % greater than that to the shoulder.

4 Dimensions

The dimensions and tolerances of screw-neck bottles are given in Figures 1 and 2 and in Table 1.