

Australian Standard[®]

**General requirements for
single-use, sterile, plasticized
polyvinyl chloride (PVC) packs
for human blood**

Part 1: Single blood packs

This Australian Standard was prepared by Committee HT/6, Transfusion Equipment for Medical Use. It was approved on behalf of the Council of Standards Australia on 17 January 1997 and published on 5 May 1997.

The following interests are represented on Committee HT/6:

Australian Chamber of Commerce and Industry
Australian Red Cross Society
Australian Society of Anaesthetists
Commonwealth Department of Health and Family Services
Department of Public Works and Services, N.S.W.
Health Department of Western Australia
Medical Industry Association of Australia
N.S.W. Health Department
Queensland Health
Royal College of Pathologists of Australasia
South Australian Health Commission

Review of Australian Standards. To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications; this information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

This Standard was issued in draft form for comment as DR 96046.

Australian Standard[®]

**General requirements for
single-use, sterile, plasticized
polyvinyl chloride (PVC) packs
for human blood**

Part 1: Single blood packs

Originated as AS 3787.1:1990.
Second edition 1997.

PREFACE

This Standard was prepared by the Standards Australia Committee HT/6, Transfusion Equipment for Medical Use to supersede AS 3787.1—1990, *General requirements for single-use, sterile, plasticized polyvinyl chloride (PVC) blood packs for whole blood and blood components*, Part 1: *Single blood packs*.

This Part of AS 3787 deals with single blood packs whereas Part 2 of the Standard deals with multiple blood pack systems.

The committee was aware of the importance for manufacturers to demonstrate the overall system performance of the blood packs in relation to blood product viability, e.g. red cell viability *in vivo* or *in vitro*. This edition does not specifically address this need.

The Standard was originally developed by the Commonwealth Therapeutic Goods Standards Committee's Subcommittee on Blood Bags and submitted to Standards Australia as the basis for the development of an Australia Standard suitable for adoption as an Order under the Therapeutic Goods Act 1966.

The objective of this Standard is to ensure that the quality of blood collection packs is such that blood and blood components are maintained at the highest possible level.

The principal differences between this Standard and the 1990 edition are as follows:

- (a) The design of the outlet port will need to be such that the risk of the closure-piercing device of an infusion (giving) set puncturing the wall of the blood collection pack is minimized.
- (b) Appendices C, F, I, P, M, R and AA have been modified.
- (c) Labelling requirements have been altered.
- (d) Requirements for testing for haemolytic effects have been modified.

The principal difference between this Standard and ISO 3826—1993, *Plastics collapsible containers for human blood and blood components*, are as follows:

- (i) Chemical tests and biological requirements are inadequate in the ISO Standard. For example, chemical testing is carried out on a water extract, which is not appropriate for a blood pack.
- (ii) The labelling requirements in this Standard for single blood pack systems refer to the Therapeutic Goods Order (TGO)—*General requirements for labels for therapeutic devices*, and the Australian Red Cross Blood Service—*Guidelines for blood and blood component labels*.
- (iii) The physical tests are not specific enough in the ISO Standard. Also, there is no test for particulate contamination.

Alternative validated test methods to those given in this Standard may be used, provided that equivalent or comparable results are obtained.

A more modern test for alkylene oxide, alkylene chlorohydrin and alkylene glycol residues (Appendix AC) is under investigation.

The terms 'normative' and 'informative' have been used in this Standard to define the application of the appendix to which they apply. A 'normative' appendix is an integral part of a Standard, whereas an 'informative' appendix is only for information and guidance.

CONTENTS

	<i>Page</i>
1 SCOPE	5
2 REFERENCED DOCUMENTS	5
3 DEFINITIONS	5
4 DESIGN	6
5 BLOOD PACK COMPONENTS	6
6 PERFORMANCE OF BLOOD COLLECTION PACK	8
7 ANTICOAGULANT SOLUTION	11
8 PACKAGING	11
9 MARKING	12
 APPENDICES	
A METHOD OF TESTING STRENGTH OF NEEDLE JOINTS	13
B METHOD OF TESTING NEEDLE PROTECTOR FOR LEAKAGE	17
C METHOD FOR DETERMINING STRENGTH OF CONNECTION AND LEAKAGE BETWEEN COLLECTION TUBE AND SOFT-WALLED CONTAINER (BAG)	19
D METHOD FOR DETERMINING STRENGTH OF CONNECTION BETWEEN CLOSURE-PIERCING DEVICE AND OUTLET PORT	21
E METHOD OF TEST FOR STRENGTH OF SUSPENSION DEVICE(S)	23
F METHOD FOR DETERMINING TRANSPARENCY	24
G METHOD OF TEST FOR LOSS OF CONTENTS	26
H METHOD FOR DETERMINING RESISTANCE TO STRETCH	28
I METHOD FOR DETERMINING LEAKAGE OF THE BLOOD COLLECTION PACK UNDER PRESSURE	29
J METHOD FOR DETERMINING THE TIME FOR FILLING THE BLOOD COLLECTION PACK TO ITS NOMINAL CAPACITY (COLLECTION RATE)	30
K METHOD FOR DETERMINING DELIVERY VOLUME	32
L METHOD FOR DETERMINING AIR CONTENT OF BLOOD COLLECTION PACK	33
M METHOD OF TEST FOR PARTICULATE CONTAMINATION (LIMIT TEST)	34
N METHOD FOR DETERMINING RESISTANCE TO CENTRIFUGATION ...	35
O METHOD OF TEST FOR PYROGENS (BACTERIAL ENDOTOXINS)	36
P METHOD OF TEST FOR CYTOTOXICITY	39
Q METHODS FOR PREPARATION AND IDENTIFICATION OF EXTRACTS FOR CHEMICAL TESTING	45
R METHOD OF TEST FOR EPOXIDIZED OILS	47
S METHOD OF TEST FOR DI(2-ETHYLHEXYL) PHTHALATE (DEHP)	49
T METHOD OF TEST FOR TRI(2-ETHYLHEXYL) TRIMELLITATE (TETM)	51
U METHOD OF TEST FOR N,N'-DIACYLETHYLENEDIAMINES	53
V METHOD OF TEST FOR VINYL CHLORIDE MONOMER (VCM)	54
W METHOD OF TEST FOR POLYVINYL CHLORIDE	57
X METHOD OF TEST FOR TRACE METALS	59
Y METHOD OF TEST FOR CALCIUM	62
Z METHOD OF TEST FOR ZINC	64

	<i>Page</i>
AA METHOD OF TEST FOR REDUCING SUBSTANCES	66
AB METHOD OF TEST FOR RESIDUE ON EVAPORATION	67
AC METHOD OF TEST FOR ALKYLENE OXIDE, ALKYLENE CHLOROHYDRIN AND ALKYLENE GLYCOL RESIDUES	68
AD METHOD FOR DETERMINING VOLUME OF ANTICOAGULANT SOLUTION	71

© Copyright — STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

STANDARDS AUSTRALIA

Australian Standard

General requirements for single-use, sterile, plasticized polyvinyl chloride (PVC) packs for human blood

Part 1: Single blood packs

1 SCOPE This Standard specifies requirements for sterilized, non-vented, collapsible, plasticized polyvinyl chloride (PVC) single blood packs for the collection, storage, transportation and administration of whole blood.

The Standard does not apply to blood bag systems used for the freezing of red cells.

2 REFERENCED DOCUMENTS The following documents are referred to in this Standard:

AS

- 1386 Cleanrooms and clean workstations
- 2103 Dial gauges and dial test indicators (metric series)
- 2134 Recommended practice for chemical analysis of materials by atomic absorption spectrometry
- 2134.1 Part 1: Flame atomic absorption spectrometry
- 2134.2 Part 2: Graphite furnace spectrometry
- 2145 Hypodermic equipment—Hypodermic needle tubing
- 2385 Single-use (sterile) infusion sets for general medical use

ISO

- 10993 Biological evaluation of medical devices
- 10993.4 Part 4: Selection of tests for interactions with blood
- 10993.10 Part 10: Tests for irritation and sensitization

British Pharmacopoeia (BP), Vol. 2

European Pharmacopoeia (EP)

United States Pharmacopoeia (USP), Monograph 85

United States Pharmacopoeia (USP), Monograph 87

United States Pharmacopoeia (USP), Monograph 161

Australian Red Cross Blood Service—Guidelines for blood and blood component labels

Guideline on validation of the *Limulus amoebocyte* lysate test as an end-product endotoxin test for human and animal parenteral drugs, biological products and medical devices, F.D.A., 1987

NHMRC Report, 81st Session, Appendix XV.

Therapeutic Goods Order (TGO)—Standard for sterile therapeutic goods (Commonwealth of Australia)

Therapeutic Goods Order (TGO)—General requirements for labels for therapeutic devices (Commonwealth of Australia)

3 DEFINITIONS For the purposes of this Standard the definitions below apply.

3.1 Collection tube—the tube through which blood is collected from the donor and a means by which blood from the blood collection pack can be sampled.