

Australian Standard<sup>®</sup>

**Medical refrigeration equipment—For  
the storage of blood and blood products**

**Part 1: Manufacturing requirements**



This Australian Standard® was prepared by Committee HE-020, Medical Refrigeration. It was approved on behalf of the Council of Standards Australia on 18 October 2012. This Standard was published on 23 November 2012.

---

The following are represented on Committee HE-020:

- Australia and New Zealand Society of Blood Transfusion
  - Australian College of Nursing
  - Australian Council on Healthcare Standards
  - Australian Institute of Refrigeration, Air-conditioning and Heating
  - Australian Red Cross Blood Service
  - CHOICE
  - Institute of Hospital Engineering Australia
  - National Association of Testing Authorities Australia
  - National Blood Authority
  - Queensland Health
  - Refrigeration & Air-conditioning Contractors Association of Australia
  - Royal College of Pathologists Australasia (RCPA)
  - Therapeutic Goods Administration
- 

This Standard was issued in draft form for comment as DR AS 3864.

Standards Australia wishes to acknowledge the participation of the expert individuals that contributed to the development of this Standard through their representation on the Committee and through the public comment period.

---

#### **Keeping Standards up-to-date**

Australian Standards® are living documents that reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued.

Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments that may have been published since the Standard was published.

Detailed information about Australian Standards, drafts, amendments and new projects can be found by visiting [www.standards.org.au](http://www.standards.org.au)

Standards Australia welcomes suggestions for improvements, and encourages readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at [mail@standards.org.au](mailto:mail@standards.org.au), or write to Standards Australia, GPO Box 476, Sydney, NSW 2001.

---

Australian Standard<sup>®</sup>

**Medical refrigeration equipment—For  
the storage of blood and blood products**

**Part 1: Manufacturing requirements**

Originated as part of AS 3864—1991.  
Previous edition part of AS 3864—1997.  
Revised in part and redesignated as AS 3864.1—2012.

**COPYRIGHT**

© Standards Australia Limited

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher, unless otherwise permitted under the Copyright Act 1968.

Published by SAI Global Limited under licence from Standards Australia Limited, GPO Box 476, Sydney, NSW 2001, Australia

ISBN 978 1 74342 292 2

## PREFACE

This Standard was prepared by the Standards Australia Committee HE-020, Medical Refrigeration, to supersede part of AS 3864—1997, *Medical refrigeration equipment—For the storage of blood and blood products*.

The objective of this Standard is to safeguard recipients of blood transfusions by ensuring that blood and blood products are properly and safely stored at the required temperature in refrigeration equipment or walk-in rooms specifically manufactured for the purpose.

The Standard consists of two parts:

Part 1: Manufacturing requirements (this Standard)

Part 2: User-related requirements for care, maintenance, performance verification and calibration

The principle differences between this edition and the 1997 edition are as follows:

- (a) Separation of the Standard into two parts. Part 1 describes the requirements for manufacturers of medical refrigeration equipment used for storing blood and blood products. Part 2 is intended for the users of this type of equipment and describes the requirements for its care, maintenance, performance verification and calibration.
- (b) Recognition of the variety of data management technologies available for acquisition and storage of data from temperature recording devices.
- (c) Recognition of advancements in refrigeration system technology.

The term, 'normative' has been used in this Standard to define the application of the appendix to which it applies. A 'normative' appendix is an integral part of a Standard.

## CONTENTS

	<i>Page</i>
SECTION 1 SCOPE AND GENERAL	
1.1 SCOPE.....	5
1.2 REFERENCED DOCUMENTS.....	5
1.3 DEFINITIONS.....	6
1.4 ELECTRICAL SAFETY .....	7
1.5 CARE AND MAINTENANCE.....	7
1.6 MARKING .....	7
SECTION 2 SELF-CONTAINED REFRIGERATION EQUIPMENT FOR THE STORAGE OF BLOOD AND BLOOD PRODUCTS IN THE TEMPERATURE RANGE 2°C TO 6°C	
2.1 SCOPE OF SECTION .....	8
2.2 CONSTRUCTION.....	8
2.3 LIGHTING .....	9
2.4 REFRIGERATION SYSTEMS .....	10
2.5 GENERAL CONDITIONS OF TESTING.....	10
2.6 PERFORMANCE REQUIREMENTS .....	10
2.7 INFORMATION TO BE SUPPLIED BY THE MANUFACTURER .....	11
SECTION 3 SELF-CONTAINED REFRIGERATION EQUIPMENT FOR THE STORAGE OF BLOOD PLASMA AND PLASMA PRODUCTS AT OR BELOW -25°C	
3.1 SCOPE OF SECTION .....	12
3.2 CONSTRUCTION.....	12
3.3 LIGHTING .....	14
3.4 REFRIGERATION SYSTEM.....	14
3.5 GENERAL CONDITIONS OF TESTING.....	15
3.6 PERFORMANCE REQUIREMENTS .....	15
3.7 INFORMATION TO BE SUPPLIED BY THE MANUFACTURER .....	16
SECTION 4 WALK-IN TYPE COOLROOMS FOR THE STORAGE OF BLOOD AND BLOOD PRODUCTS IN THE TEMPERATURE RANGE 2°C TO 6°C	
4.1 SCOPE OF SECTION .....	17
4.2 CONSTRUCTION.....	17
4.3 LIGHTING .....	19
4.4 REFRIGERATION SYSTEM.....	20
4.5 GENERAL CONDITIONS OF TESTING.....	20
4.6 PERFORMANCE REQUIREMENTS .....	20
4.7 INFORMATION TO BE SUPPLIED BY THE MANUFACTURER .....	21
SECTION 5 WALK-IN TYPE FREEZER ROOMS FOR THE STORAGE OF BLOOD PLASMA AND PLASMA PRODUCTS AT OR BELOW -25°C	
5.1 SCOPE OF SECTION .....	23
5.2 CONSTRUCTION.....	23
5.3 LIGHTING .....	25
5.4 REFRIGERATION SYSTEM.....	26
5.5 GENERAL CONDITIONS OF TESTING.....	26
5.6 PERFORMANCE REQUIREMENTS .....	27
5.7 INFORMATION TO BE SUPPLIED BY THE MANUFACTURER .....	27

## SECTION 6 TEMPERATURE RECORDING AND ALARM SYSTEMS

6.1	INSTRUMENTS AND TEMPERATURE MEASURING APPARATUS .....	29
6.2	ALARM SYSTEMS .....	30

## APPENDICES

A	GENERAL TEST CONDITIONS.....	33
B	METHOD FOR DETERMINING EXTERNAL SURFACE TEMPERATURE OF A REFRIGERATOR CABINET MAINTAINED IN THE TEMPERATURE RANGE 2°C TO 6°C .....	36
C	OPERATION TEST FOR A REFRIGERATOR CABINET USED FOR THE STORAGE OF BLOOD AND BLOOD PRODUCTS IN THE TEMPERATURE RANGE 2°C TO 6°C.....	38
D	METHOD FOR DETERMINATION OF SERVICE LOAD CAPACITY OF A REFRIGERATOR CABINET.....	41
E	METHOD FOR DETERMINATION OF HEAT INSULATING CAPACITY OF A REFRIGERATOR CABINET.....	45
F	OPERATION TEST FOR A FREEZER CABINET.....	46
G	METHOD FOR DETERMINATION OF EXTERNAL SURFACE TEMPERATURE OF A FREEZER CABINET .....	50
H	METHOD FOR TESTING TEMPERATURE RECORDING AND ALARM SYSTEMS .....	52

# STANDARDS AUSTRALIA

## Australian Standard

### Medical refrigeration equipment—For the storage of blood and blood products

#### Part 1: Manufacturing requirements

## SECTION 1 SCOPE AND GENERAL

### 1.1 SCOPE

This Standard specifies requirements for the manufacture of medical refrigeration equipment used within an ambient temperature range of 10°C to 43°C (see Note 1) for the storage of the following:

- (a) Blood and blood products in the temperature range 2°C to 6°C (see Sections 2 and 4);
- (b) Frozen plasma and plasma products at a temperature of –25°C or lower (see Sections 3 and 5)

This Standard covers both reach-in cabinets and walk-in rooms. This Standard is intended to be read in conjunction with AS 3864.2. Where the term ‘the Standard’ is used it refers to both Parts 1 and 2 of AS 3864.

#### NOTES:

- 1 Where medical refrigeration equipment is intended to be used in an air-conditioned atmosphere, only the ambient temperature range of 10°C to 32°C will be applicable for satisfying the requirements of this Standard, provided there is an adequate back-up electrical supply in case of power failure.
- 2 In some situations equipment meeting the scope of the Standard may also be required to meet additional regulatory requirements. These requirements are outside the scope of the Standard.

### 1.2 REFERENCED DOCUMENTS

The following documents are referred to in this Standard:

#### AS

- |        |   |
|--------|---|
| 1397   | Continuous hot-dip metallic coated steel sheet and strip—Coatings of zinc and zinc alloyed with aluminium and magnesium |
| 1627   | Metal finishing—Preparation and pretreatment of surfaces  |
| 1627.0 | Part 0: Method selection guide  |
| 2070   | Plastics materials for food contact use   |
| 3864   | Medical refrigeration equipment—For the storage of blood and blood products   |
| 3864.2 | Part 2: User-related requirements for care maintenance, performance verification and calibration.                       |
| 60529  | Degrees of protection provided by enclosures (IP Code)  |

#### AS/NZS

- |        |   |
|--------|---|
| 1677   | Refrigeration systems   |
| 1677.1 | Part 1: Refrigerant classification  |
| 1677.2 | Part 2: Safety requirements for fixed applications                            |
| 3100   | Approval and test specification—General requirements for electrical equipment |