

Australian Standard<sup>®</sup>

---

**DENTAL EQUIPMENT—  
SUCTION SYSTEMS**

**Part 2—MOBILE SYSTEMS**

---

This Australian standard was prepared by Committee DN/4, Dental Equipment. It was approved on behalf of the Council of the Standards Association of Australia on 6 February 1985 and published on 7 June 1985.

---

The following interests are represented on Committee DN/4:

- Australian Chamber of Commerce
  - Australian Dental Association
  - Australian Dental Standards Laboratory
  - Australian Dental Trade Association
  - Confederation of Australian Industry
  - Dental Hospitals
  - Department of Defence
  - The Royal Dental Hospital of Melbourne
  - United Dental Hospital of Sydney
  - University of Melbourne, Dental School
  - University of Sydney, Dental School
  - Victorian Employers Federation
- 

**Review of Australian Standards.** *To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.*

*Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications; this information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.*

*Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.*

---

First published . . . . . 1985
--------------------------------

*This standard was issued in draft form for comment as DR 83021.*

## PREFACE

This standard was prepared by the Association's Committee on Dental Equipment under the direction of the Dental Materials and Equipment Standards Committee. In the preparation of this standard, account was taken of Part 1 of this standard, Central Systems.

This standard applies to mobile systems where the pump is integral with the surgery equipment and serves only one chair.

The committee believed that as mobile systems discharge into the room atmosphere, the fitting of an effective bacterial filter is essential.

---

## CONTENTS

	<i>Page</i>
SPECIFICATION	
1 Scope . . . . .	4
2 Application . . . . .	4
3 Referenced Documents . . . . .	4
4 Definitions . . . . .	4
5 Performance Requirements . . . . .	4
6 Maintenance . . . . .	5
7 Noise Restriction . . . . .	5
8 Electrical Requirements . . . . .	5
9 Bacterial Filter . . . . .	5
10 Information to be Supplied by the Manufacturer . . . . .	5
APPENDICES	
A Method for Determining Maximum Suction Pressure . . . . .	7
B Method for Determining Free Air Flow . . . . .	7

© Copyright — STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.