

1982 ED.



**Standards
Association of
Australia**



**Australian Standard[®]
1052—1988**

**ELECTROMAGNETIC INTERFERENCE—
MEASURING APPARATUS AND
MEASUREMENT METHODS**



This Australian Standard was prepared by Committee TE/3, Electromagnetic Interference. It was approved on behalf of the Council of the Standards Association of Australia on 16 November 1987 and published on 4 January 1988.

The following interests are represented on Committee TE/3:

Association of Consulting Engineers Australia
AUSSAT
Australian Broadcasting Corporation
Commonwealth Scientific and Industrial Research Organisation
Confederation of Australian Industry
Consumer Electronics Suppliers Association
Department of Aviation
Department of Transport and Communications
Department of Defence
Electricity Supply Association of Australia
Federation of Australian Radio Broadcasters
Federation of Australian Commercial Television Stations
Institution of Radio and Electronics Engineers Australia
Macquarie University
Metropolitan Transit Authority (Melbourne)
Society of Automotive Engineers Australia
Telecom Australia
Testing interests
Wireless Institute of Australia

Review of Australian Standards. To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up-to-date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all SAA publications will be found in the Catalogue of SAA Publications; this information is supplemented each month by SAA's journal 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of the Association, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

This Standard was issued in draft form for comment as DR 85115.

*Ground planes or Ground Bed see CISPR 22
CRO 9008 28.*

AUSTRALIAN STANDARD

ELECTROMAGNETIC INTERFERENCE— MEASURING APPARATUS AND MEASUREMENT METHODS

AS 1052—1988

AS 1052.2 first published as C348.1—1962 and
C349.1—1962.
Revised, amalgamated and redesignated as
AS 1052.2—1976.
AS 1052.1 first published 1976.
AS 1052.1—1976 and AS 1052.2—1976 revised,
amalgamated and redesignated as AS 1052—1988.

**PUBLISHED BY THE STANDARDS ASSOCIATION OF AUSTRALIA
STANDARDS HOUSE, 80 ARTHUR ST, NORTH SYDNEY, N.S.W.**

ISBN 0 7262 4821 5

PREFACE

This Standard was prepared by the Association's Committee on Electromagnetic Interference and supersedes AS 1052, Electromagnetic Interference Measuring Equipment, Part 1—1976, Equipment for the Frequency Range 10 kHz to 150 kHz, and Part 2—1976, Equipment for the Frequency Range 0.15 MHz to 1000 MHz.

It is identical with and has been reproduced from IEC/CISPR 16 (1987), Specification for Radio Interference Measuring Apparatus and Measurement Methods, except that the title is in Australia's nomenclature.

This Standard retains the generic AS number of AS 1052 since AS 1052, Part 1—1976 and Part 2—1976 are incorporated into legislation (N.S.W. Coal Mines Regulation Act). This Standard is designated AS 1052—1988.

Alignment of this Standard with the IEC/CISPR 16 allows a considerable expansion of the subject matter including measurement techniques from 9 kHz to 18 GHz, new material dealing with average measuring receivers for the measurement of narrowband signals, ferrite clamps, current probes, radiation measurements in the microwave region and requirements for disturbance analysers for the measurement of discontinuous interference.

For the purposes of this Australian Standard, the text of the IEC/CISPR 16 should be modified as follows:

- (a) *Terminology:* The words 'Australian Standard' should replace the words 'IEC Publication' wherever they appear.
- (b) *Cross-references:* The references to IEC Publications should be replaced by references to Australian Standards as follows:

<i>Reference to IEC Publication</i>	<i>Appropriate Australian Standard</i>
CISPR 11 Limits and methods of measurement of radio interference characteristics of industrial, scientific and medical (ISM) radio-frequency equipment (excluding surgical diathermy apparatus)	AS 2064 Limits and methods of measurement of radio interference characteristics of industrial, scientific and medical (ISM) radio-frequency equipment (excluding surgical diathermy apparatus)
CISPR 12 Limits and methods of measurement of radio interference characteristics of vehicles, motor boats and spark-ignited engine-driven devices	No Australian equivalent
CISPR 13 Limits and methods of measurement of radio interference characteristics of sound and television receivers	AS 1053 Limits and methods of measurement of radio interference characteristics of sound and television receivers
CISPR 14 Limits and methods of measurement of radio interference characteristics of household electrical appliances, portable tools and similar electrical apparatus	AS 1044 Limits and methods of measurement of radio interference characteristics of household electrical appliances, portable tools and similar electrical apparatus
CISPR 15 Limits and methods of measurement of radio interference characteristics of fluorescent lamps and luminaires	No Australian equivalent
CISPR 18-1 Radio interference characteristics of overhead power lines and high-voltage equipment	AS 2344 Limits of electromagnetic interference from overhead a.c. power lines

- | | | |
|----------|---|--------------------------|
| CISPR 20 | Measurement of the immunity of sound and television broadcast receivers and associated equipment in the frequency range 1.5 MHz to 30 MHz by the current-injection method. Guidance on immunity requirements for the reduction of interference caused by radio transmitters in the frequency range 26 MHz to 30 MHz | No Australian equivalent |
| CISPR 22 | Limits and methods of measurement of radio interference characteristics of information technology equipment | No Australian equivalent |
| IEC 96-1 | Radio frequency cables—
General requirements | No Australian equivalent |
- (c) *Page number references:* The text references apply to the IEC page numbers at the bottom left-hand corner of each page.

© Copyright — STANDARDS ASSOCIATION OF AUSTRALIA 1988

Users of Standards are reminded that copyright subsists in all SAA publications. Except where the Copyright Act otherwise allows, no part of this publication may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing of the Standards Association of Australia. Requests for permission should be directed to the Head Office of the Association. Where such requests relate to the reproduction of the whole or a substantial part of any Standard, permission may be conditional on an appropriate royalty payment.

CONTENTS

	Page
INTRODUCTION	7
SCOPE	9
SECTION ONE – MEASURING APPARATUS	
Clause	
1. Fundamental characteristics	10
2. Normal response of measuring apparatus to pulses	11
2.1 Amplitude relationship	11
2.2 Variation with repetition frequency	11
3. Selectivity	12
3.1 Overall selectivity (passband)	12
3.2 Intermediate-frequency rejection ratio	12
3.3 Image frequency rejection ratio	12
3.4 Other spurious responses	12
4. Limitation of intermodulation effects	12
5. Limitation of background noise and internally generated spurious signals	13
5.1 Random noise	13
5.2 Continuous wave (Band D only)	13
6. Screening	13
6.1 Band A	13
6.2 Bands B, C and D	13
7. Accuracy of measuring apparatus	14
7.1 Voltage measurement	14
7.2 Field-strength measurement	14
SECTION TWO – CONDUCTION MEASUREMENTS OF RADIO INTERFERENCE	
8. Artificial mains network	14
8.1 General	14
8.2 Arrangement for d.c. or single-phase a.c. supplies	14
8.3 Circuits with more than two conductors (three-phase a.c. supply with or without neutral)	15
8.4 Isolation	15
8.5 Connection between the artificial mains network and the measuring apparatus	15
9. Measurements of radio-interference voltages	16
9.1 Reduction of interference not produced by the device under test	16
9.2 Disposition of devices and their connection to the artificial mains network	17
10. Measurement of radio-interference current (10 kHz to 150 kHz)	17
10.1 Introduction	17
10.2 Characteristics	17
11. Methods of measurement of interference power from mains-operated devices (30 MHz to 300 MHz)	18
11.1 General	18
11.2 Measurement procedure	18
11.3 Calibration	18
11.4 Impedance measurement	18
11.5 Special procedure for the measurement of interference sources producing discontinuous interference	19
SECTION THREE – RADIATION MEASUREMENTS OF RADIO INTERFERENCE	
12. General	19
13. Types of aerials	20
13.1 Frequency range 10 kHz to 150 kHz	20
13.2 Frequency range 150 kHz to 30 MHz	20
13.3 Frequency range 30 MHz to 300 MHz	20
13.4 Frequency range 300 MHz to 1 000 MHz	21
14. Distances of measurement	22
15. Test site	22
15.1 Disposition of devices and their connection to the mains	23
16. Test procedure	23
16.1 Determination of field strength in direction of maximum radiation	23
16.2 Open-space tests (remote from test device)	23
16.3 Tests on installation	23
16.4 Methods of measurement of radiated power from devices with built-in batteries (30 MHz to 300 MHz)	24

SECTION FOUR – METHODS OF MEASUREMENT OF VARIOUS TYPES OF INTERFERENCE-PRODUCING DEVICES AND SYSTEMS

Clause	Page
17. Domestic appliances (excluding radio and television receivers)	25
17.1 Measurement of interference-producing voltages (0.15 MHz to 30 MHz)	25
17.2 Measurement of interference-producing current (10 kHz to 150 kHz)	25
17.3 Measurement of interference-producing power (30 MHz to 300 MHz)	25
17.4 Measurement of interference-producing fields	25
18. Radio and television receivers	25
19. Industrial, scientific and medical (ISM) radio-frequency equipment	25
19.1 Measuring apparatus (0.15 MHz to 1 000 MHz)	25
19.2 Methods of measurement	26
20. Ignition systems of motor vehicles and other devices	26
21. Measurement of high-voltage transmission systems	26
21.1 Measurement frequency	26
21.2 Noise voltage (current) measurement for line equipment	26
21.3 Radiation measurements (on overhead lines)	27

SECTION FIVE – RADIO INTERFERENCE MEASURING APPARATUS HAVING DETECTORS
OTHER THAN QUASI-PEAK

22. R.M.S. detector	31
22.1 Introduction	31
22.2 Fundamental characteristics	31
22.3 Normal response of measuring apparatus to pulses	31
23. Average detector	32
23.1 Fundamental characteristics	32
23.2 Normal response of measuring apparatus to pulses	32
24. Peak detectors	33
24.1 Introduction	33
24.2 Fundamental characteristics	33
24.3 Normal response to pulses	33

SECTION SIX – AUDIO-FREQUENCY INTERFERENCE MEASUREMENTS

25. Introduction	34
26. Fundamental characteristics	35
26.1 Input impedance	35
26.2 Measuring range	35
26.3 Filter network (frequency weighting) (c.w.)	35
26.4 Quasi-peak voltmeter	35
26.5 R.M.S. voltmeter	35
26.6 Balance (600 Ω input terminals)	35
26.7 Calibration error	35
26.8 Immunity from disturbances by alternating magnetic fields at the supply frequency	36
27. Specific requirements	36
27.1 For use as a quasi-peak voltmeter	36
27.2 For use as an r.m.s. voltmeter	36

SECTION SEVEN – MEASUREMENT OF DISTURBANCES DUE TO SWITCHING OPERATIONS

28. Introduction	37
29. Measurement of the duration of disturbances less than 10 ms	37
30. Disturbance analyzer for the automatic assessment of interference produced by switching operations	37

SECTION EIGHT – MEASUREMENT OF THE SHIELDING EFFICIENCY OF COAXIAL CABLES HAVING ONE OR TWO BRAIDS
IN THE METRIC WAVELENGTH RANGE

31. Introduction	38
32. Method of the absorbing clamp	38
32.1 Test arrangement	38
32.2 Definition of shielding efficiency	39
32.3 Precautions to be observed for measurement of doubled-shielded cables	39
33. Radiation method	40
33.1 Test arrangement	40
33.2 Definition of shielding efficiency	40
33.3 Precautions to be observed for measurement of multi-shielded cables	40
33.4 Correlation between the shielding efficiency and the surface transfer impedance	40

**SECTION NINE – STATISTICAL CONSIDERATIONS IN THE DETERMINATION OF LIMITS
OF RADIO INTERFERENCE**

Clause	Page
34. Introduction	41
35. Tests based on the non-central t distribution (sampling by variables)	41
35.1 Determination of the constant k	42
35.2 Determination of the sample size n	43
35.3 Example	44
36. Tests based on the binomial distribution (sampling by attributes)	44
36.1 Determination of constant c	44
36.2 Determination of sample size n	45
36.3 Control charts	46

**SECTION TEN – MEASUREMENT OF TOTAL RADIATED POWER FROM APPARATUS OPERATING IN
THE MICROWAVE RANGE (REVERBERATING CHAMBER METHOD)**

37. Introduction	46
38. Chamber	47
39. Stirrers	47
40. Coupling attenuation	48
41. Radiated power measurement	48
42. Effect of the chamber on total radiated power	49
APPENDIX A – Definitions and methods of measuring the fundamental characteristics of the receiver	50
APPENDIX B – Characteristics of a spectrum analyzer for use in the frequency range 0.3 GHz to 18 GHz	53
APPENDIX C – Determination of response to repeated pulses	55
APPENDIX D – Determination of pulse generator spectrum	57
APPENDIX E – Artificial mains networks	59
APPENDIX F – Connection of electrical equipment to the artificial mains network	62
APPENDIX G – Example of a device and its application for the measurement of interference from mains powered appliances as specified in Sub-clause 11.2	67
APPENDIX H – Field measurement at high frequencies	69
APPENDIX I – Propagation of interference from industrial radio-frequency equipment at frequencies between 30 MHz and 300 MHz	71
APPENDIX J – Interference from power lines	72
APPENDIX K – Calculation of the voltage gradient at the surface of conductors of high-voltage lines	75
APPENDIX L – Propagation of radio frequencies on high-voltage transmission lines	77
APPENDIX M – Correlation between measurements made with apparatus having characteristics differing from the C.I.S.P.R. characteristics and measurements made with C.I.S.P.R. apparatus	78
APPENDIX N – Definitions of the fundamental characteristics of a measuring apparatus employing an r.m.s. detector	83
APPENDIX O – Response of average and peak detectors	84
APPENDIX P – Determination of response of r.m.s. detector to pulses	87
APPENDIX Q – Accurate measurements of the output of nanosecond pulse generators	90
APPENDIX R – Disturbance analyzer performance checks	93
APPENDIX S – Historical background to the method of measurement of the interference power produced by electrical household and similar appliances in the v.h.f. range	96
FIGURES 1a to 47	99

STANDARDS ASSOCIATION OF AUSTRALIA**Australian Standard****ELECTROMAGNETIC INTERFERENCE—
MEASURING APPARATUS AND MEASUREMENT METHODS****SECTION ZERO—INTRODUCTION AND SCOPE****INTRODUCTION**

The initial objective of the C.I.S.P.R. method of voltage measurement was to provide, in the frequency range from 150 kHz to 1 605 kHz, an assessment of interference related to its effect on the reception of sound broadcasting. Much of the interference is impulsive in nature and its effect increases with increasing repetition rate in a way that has been shown to be approximated by a quasi-peak detector circuit with an appropriate set of time constants. Over the years, the quasi-peak technique has been extended in frequency to cover the range from 10 kHz to 1 GHz, and is applied for the protection of services other than radio broadcasting. Instruments using the quasi-peak detector still remain as the basic reference for determining compliance with C.I.S.P.R. limits.

Other measures can be useful in specific instances: included are the average, r.m.s. and peak measures. All of these are described in this publication. The quasi-peak detector is the most extensively used in the protection of broadcasting services. Its characteristics are best described in terms of its response to short, constant amplitude, pulses of adjustable level and whose repetition frequency may be varied from that of an isolated pulse to a high value. Following the usual practice, this response is expressed in terms of the r.m.s. value of the unmodulated sine-wave voltage (or field) injected at the input of the measuring apparatus under the same condition as for the pulses, and which produces the same indication on the measuring apparatus.

The measuring apparatus (usually of the superheterodyne type) has the following general characteristics:

- solely manual control of sensitivity,
- a defined overall bandwidth.

The apparatus is intended to measure the interference signal which may be conducted into the supply mains or radiated from the interference-producing device. In the general case when the device is connected to the supply mains, use is made of a special circuit known as the artificial mains network. The function of this network is, on the one hand, to isolate, at radio frequencies, the interfering device from the supply mains, and on the other, to provide a defined impedance across the terminals of the device. For measuring the interference power emitted by an appliance through its supply mains cord an absorbing clamp is used at frequencies above 30 MHz.

For the measurement of radiated interference, the measuring apparatus is connected to a suitable aerial. The response of the apparatus is expressed in terms of the r.m.s. value of the electric or magnetic component of the field which will give the same indication on the measuring apparatus.

The present instrument specification prescribes only those characteristics imposed by the principles of the method of measurement and refers primarily to the superheterodyne type of apparatus. Either a superheterodyne or tuned r.f. type of apparatus may, however, be used. Other characteristics which are subject to the conditions of use, such as the frequency coverage and the range of voltage or field levels, are left to individual choice.

Notes 1. — It is recommended that a conventional detector and audio-frequency amplifier (preferably gain control) be included in the measuring apparatus for aural monitoring of the interference to be measured.

2. — It is also recommended that a signal source be included so that the gain of the measuring apparatus may be set to the value used during the initial calibration.

The broadcasting services in the higher frequency range covered by this publication are very varied in nature and both aural and visual presentation are employed. Thus, it appears that while an apparatus similar to that used in the lower frequency ranges might be developed for each type of transmission to be considered, a universal measuring apparatus would hardly be possible.

For this reason, the tendency, which over the years is more and more marked, has been to subordinate agreement between subjective effect and objective measurement to the exigencies of the facility of making good measurements, is strongly emphasized in this publication.

The fundamental characteristics of the measuring apparatus over the higher frequency ranges have therefore been chosen in such a way as to obtain a compromise between the conditions appropriate to the frequencies under consideration and the measuring requirements, as well as maintaining a similarity to the lower frequency specification as regards the behaviour of the response to repeated pulses.

Cognizance has also been taken of the number of measuring apparatus in use which have fundamental characteristics approximating those chosen.

Further study will be necessary to establish the correlation between measurements made with apparatus complying with this specification and the different classes of subjective effect. They will assist in determining tolerable limits of interference voltages, currents, power and fields.

Scope

This publication stipulates performance requirements for radio interference measuring apparatus, including the associated artificial mains network.

The requirements of this publication shall be complied with at all frequencies and for all levels of radio-interference voltages, currents, power or field strengths within the range of the measuring apparatus.

The publication is divided into sections, as follows:

- Section One: Measuring apparatus.
- Section Two: Conduction measurements of radio interference.
- Section Three: Radiation measurements of radio interference.
- Section Four: Method of measurement of various types of interference-producing devices and systems.
- Section Five: Radio interference measuring apparatus having detectors other than quasi-peak.
- Section Six: Audio-frequency interference measurements.
- Section Seven: Measurement of disturbances due to switching operations.
- Section Eight: Measurement of the shielding efficiency of coaxial cables having one or two braids in the metric wavelength range.
- Section Nine: Statistical considerations in the determination of limits of radio interference.
- Section Ten: Measurement of total radiated power from apparatus in the microwave range (reverberating chamber method).

The appendices to this publication give additional information on the fundamental characteristics on which the requirements are based, information of general interest on propagation of interference from industrial equipment and the reproduction of the parts of C.I.S.P.R. Publication 1 about interference from power lines.

Sections Two and Three lay down only general requirements for the measurement of conduction and radiation characteristics of radio interference. Section Four lays down only general requirements applicable to the measurement of interference produced by various types of interference-producing devices and systems. Detailed requirements for the measurement of radio interference produced by specific types of devices are stipulated in other C.I.S.P.R. publications, as follows:

- C.I.S.P.R. 11 (1975): Limits and methods of measurement of radio interference characteristics of industrial, scientific and medical (ISM) radio-frequency equipment (excluding surgical diathermy apparatus).
- C.I.S.P.R. 12 (1985): Limits and methods of measurement of radio interference characteristics of vehicles, motor boats and spark-ignited engine-driven devices.
- C.I.S.P.R. 13 (1975): Limits and methods of measurement of radio interference characteristics of sound and television receivers.