



## **Medical electrical equipment**

### **Part 1.11: General requirements for basic safety and essential performance—Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment**



This Australian Standard® was prepared by Committee HE-003, Medical Electrical Equipment. It was approved on behalf of the Council of Standards Australia on 10 May 2017. This Standard was published on 23 June 2017.

---

The following are represented on Committee HE-003:

- Australasian College of Physical Scientists and Engineers in Medicine
  - Australian and New Zealand College of Anaesthetists
  - Australian Chamber of Commerce and Industry
  - Australian Radiation Protection and Nuclear Safety Agency
  - Australian Society of Anaesthetists
  - College of Biomedical Engineering Engineers Australia
  - Department of Defence (Australian Government)
  - Engineers Australia
  - Testing Interests (Australia)
  - Therapeutic Goods Administration
- 

This Standard was issued in draft form for comment as DR AS IEC 60601.1.11:2017.

Standards Australia wishes to acknowledge the participation of the expert individuals that contributed to the development of this Standard through their representation on the Committee and through the public comment period.

---

### **Keeping Standards up-to-date**

Australian Standards® are living documents that reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued.

Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments that may have been published since the Standard was published.

Detailed information about Australian Standards, drafts, amendments and new projects can be found by visiting [www.standards.org.au](http://www.standards.org.au)

Standards Australia welcomes suggestions for improvements, and encourages readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at [mail@standards.org.au](mailto:mail@standards.org.au), or write to Standards Australia, GPO Box 476, Sydney, NSW 2001.

---

Australian Standard®

## **Medical electrical equipment**

### **Part 1.11: General requirements for basic safety and essential performance—Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment**

First published as AS IEC 60601.1.11:2017.

#### **COPYRIGHT**

© Standards Australia Limited

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher, unless otherwise permitted under the Copyright Act 1968.

Published by SAI Global Limited under licence from Standards Australia Limited, GPO Box 476, Sydney, NSW 2001, Australia

ISBN 978 1 76035 822 8

## PREFACE

This Standard was prepared by the Standards Australia Committee HE-003, Medical Electrical Equipment.

The objective of this Standard is to specify requirements for the basic safety and essential performance of medical electrical (ME) equipment and ME systems for use in the home healthcare environment. It applies regardless of whether the ME equipment or ME system is intended for use by a lay operator or by trained healthcare personnel.

This Standard is identical with, and has been reproduced from IEC 60601-1-11:2015 (ED. 2), *Medical electrical equipment, Part 1-11: General requirements for basic safety and essential performance—Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment*.

As this Standard is reproduced from an International Standard, the following applies:

- (a) In the source text ‘this International Standard’ should read ‘this Australian Standard’.
- (b) A full point substitutes for a comma when referring to a decimal marker.

None of the normative references in the source document have been adopted as Australian or Australian/New Zealand Standards.

## CONTENTS

FOREWORD .....	5
INTRODUCTION .....	8
1 Scope, object and related standards .....	9
1.1 * Scope .....	9
1.2 Object .....	9
1.3 Related standards .....	9
1.3.1 IEC 60601-1 .....	9
1.3.2 Particular standards .....	10
2 Normative references .....	10
3 Terms and definitions .....	11
4 General requirements .....	12
4.1 * Additional requirements for SUPPLY MAINS for ME EQUIPMENT and ME SYSTEMS .....	12
4.2 * Environmental conditions for ME EQUIPMENT .....	12
4.2.1 General .....	12
4.2.2 * Environmental conditions of transport and storage between uses .....	13
4.2.3 * Environmental operating conditions .....	14
5 * General requirements for testing ME EQUIPMENT .....	16
6 * Classification of ME EQUIPMENT and ME SYSTEMS .....	17
7 ME EQUIPMENT identification, marking and documents .....	17
7.1 * USABILITY of the ACCOMPANYING DOCUMENTS .....	17
7.2 * Additional requirements for marking of IP classification .....	18
7.3 ACCOMPANYING DOCUMENTS .....	18
7.3.1 Contact information .....	18
7.3.2 LAY OPERATOR briefing information .....	18
7.4 Instructions for use .....	19
7.4.1 Additional requirements for warning and safety notices .....	19
7.4.2 * Additional requirements for an electrical power source .....	19
7.4.3 Additional requirements for ME EQUIPMENT description .....	20
7.4.4 Additional requirements for ME EQUIPMENT start-up PROCEDURE .....	20
7.4.5 Additional requirements for operating instructions .....	20
7.4.6 Additional requirements for ME EQUIPMENT messages .....	21
7.4.7 * Additional requirements for cleaning, disinfection and sterilization .....	21
7.4.8 Additional requirements for maintenance .....	21
7.4.9 Additional requirements for environmental protection .....	21
7.4.10 Additional requirements for ME EQUIPMENT and ME SYSTEMS .....	22
7.5 Technical description .....	22
7.5.1 PERMANENTLY INSTALLED CLASS I ME EQUIPMENT .....	22
7.5.2 Additional requirements for professional hygienic maintenance .....	22
8 Protection against excessive temperatures and other HAZARDS .....	22
8.1 * Additional requirements for cleaning, disinfection of ME EQUIPMENT and ME SYSTEMS .....	22
8.2 * Additional requirements for sterilization of ME EQUIPMENT and ME SYSTEMS .....	23
8.3 Additional requirements for ingress of water or particulate matter into ME EQUIPMENT and ME SYSTEMS .....	23
8.3.1 * Ingress of water or particulate matter into ME EQUIPMENT .....	23

8.3.2	* Ingress of water or particulate matter into ME SYSTEMS .....	23
8.4	Additional requirements for interruption of the power supply/SUPPLY MAINS to ME EQUIPMENT and ME SYSTEM .....	23
8.5	Additional requirements for an INTERNAL ELECTRICAL POWER SOURCE .....	24
8.5.1	* Indication of state .....	24
8.5.2	Accessibility of small INTERNAL ELECTRICAL POWER SOURCES .....	25
9	Accuracy of controls and instruments and protection against hazardous outputs .....	25
10	Construction of ME EQUIPMENT .....	25
10.1	* Additional requirements for mechanical strength .....	25
10.1.1	General requirements for mechanical strength .....	25
10.1.2	* Requirements for mechanical strength for non-TRANSIT-OPERABLE ME EQUIPMENT .....	27
10.1.3	* Requirements for mechanical strength for TRANSIT-OPERABLE ME EQUIPMENT .....	28
10.2	Additional requirements for actuating parts of controls of ME EQUIPMENT .....	29
11	* Protection against strangulation or asphyxiation .....	30
12	Additional requirements for ELECTROMAGNETIC EMISSIONS of ME EQUIPMENT and ME SYSTEMS .....	30
13	Additional requirements for ALARM SYSTEMS of ME EQUIPMENT and ME SYSTEMS .....	30
13.1	* Additional requirement for generation of ALARM SIGNALS .....	30
13.2	* Additional requirement for ALARM SIGNAL volume .....	30
Annex A (informative)	General guidance and rationale .....	31
A.1	General guidance .....	31
A.2	Rationale for particular clauses and subclauses .....	32
Annex B (informative)	Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS .....	51
B.1	Marking on the outside of ME EQUIPMENT, ME SYSTEMS or their parts .....	51
B.2	ACCOMPANYING DOCUMENTS, general .....	51
B.3	ACCOMPANYING DOCUMENTS, instructions for use .....	51
B.4	ACCOMPANYING DOCUMENTS, technical description .....	53
Annex C (informative)	Symbols on marking .....	54
Bibliography	.....	56
Index of defined terms used in this collateral standard	.....	58
Figure 1	– Small finger probe Ø 5,6 .....	17
Figure A.1	– Saturation water vapour pressure as function of temperature .....	36
Table 1	– Mechanical strength test applicability, non-TRANSIT-OPERABLE .....	26
Table 2	– Mechanical strength test applicability, TRANSIT-OPERABLE .....	27
Table A.1	– Saturation water vapour pressure as function of temperature .....	37
Table A.2	– Summary by use of HOME HEALTHCARE ENVIRONMENT ME EQUIPMENT ENCLOSURE ingress of water and particulate matter requirements .....	46
Table A.3	– Qualitative assessment of HOME HEALTHCARE ENVIRONMENT ME EQUIPMENT subjected to shock and vibration .....	47
Table B.1	– Marking on the outside of ME EQUIPMENT, ME SYSTEMS or their parts .....	51
Table B.2	– ACCOMPANYING DOCUMENTS, general .....	51
Table B.3	– ACCOMPANYING DOCUMENTS, instructions for use .....	52

Table B.4 – ACCOMPANYING DOCUMENTS, technical description.....	53
Table C.1 – General symbols .....	54

## FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-1-11 has been prepared by a joint working group of IEC subcommittee 62A: Common aspects of electrical equipment used in medical practice of IEC technical committee 62: Electrical equipment in medical practice and ISO subcommittee SC3: Lung ventilators and related devices, of ISO technical committee 121: Anaesthetic and respiratory equipment.

It is published as a double logo standard.

This second edition constitutes a collateral standard to IEC 60601-1 (third edition, including Amendment 1): *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*, hereafter referred to as the general standard.

This second edition cancels and replaces the first edition of IEC 60601-1-11, published in 2010, and constitutes a technical revision.

The most significant changes with respect to the previous edition include the following modifications:

- correction of test method for relative humidity control at temperatures above 35 °C;
- redrafting of subclauses that altered instead of adding to the general standard or other collateral standards; and
- harmonizing with the changes to the amendments to the general standard and other collateral standards.

The text of this collateral standard is based on the following documents:

FDIS	Report on voting
62A/959/FDIS	62A/978/RVD

Full information on the voting for the approval of this collateral standard can be found in the report on voting indicated in the above table. In ISO, the standard has been approved by 17 P-members out of 17 having cast a vote.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In the IEC 60601 series of publications, collateral standards specify general requirements for safety applicable to:

- a subgroup of MEDICAL ELECTRICAL EQUIPMENT (e.g. radiological equipment); or
- a specific characteristic of all MEDICAL ELECTRICAL EQUIPMENT, not fully addressed in the general standard (e.g. ALARM SYSTEMS).

In this collateral standard, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS COLLATERAL STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this collateral standard, the term

- “clause” means one of the numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.3.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this collateral standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;

- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

Clauses, subclauses and definitions for which a rationale is provided in informative Annex A are marked with an asterisk (\*).

A list of all parts of the IEC 60601 series, published under the general title: *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The attention of Member Bodies and National Committees is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised ISO or IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for mandatory implementation nationally not earlier than 3 years from the date of publication.

**IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.**

## INTRODUCTION

Medical practice is increasingly using MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS for monitoring, treatment or diagnosis of PATIENTS in the HOME HEALTHCARE ENVIRONMENT (see 3.1). The safety of MEDICAL ELECTRICAL EQUIPMENT in this uncontrolled environment with regard to the electrical installation and its related safety and protection means is a cause for concern.

The potential lack of training of the LAY OPERATOR and possibly of those supervising the use of the MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEM and their level of education need to be addressed in the development of the ACCOMPANYING DOCUMENTS and in the relevant marking on the equipment itself so that this material can be understood. This collateral standard gives special guidance on how this should be addressed in the instructions for use.

This collateral standard was developed with contributions from clinicians, engineers and regulators. The terminology, requirements, general recommendations and guidance of this collateral standard are intended to be useful for MANUFACTURERS of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS and for technical committees responsible for the development of particular standards.

## AUSTRALIAN STANDARD

**Medical electrical equipment**

## Part 1.11:

General requirements for basic safety and essential performance—  
Collateral Standard: Requirements for medical electrical equipment and  
medical electrical systems used in the home healthcare environment

**1 Scope, object and related standards****1.1 \* Scope**

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS for use in the HOME HEALTHCARE ENVIRONMENT, as defined in 3.1, and specified by the MANUFACTURER in the instructions for use. This International Standard applies regardless of whether the ME EQUIPMENT or ME SYSTEM is intended for use by a LAY OPERATOR or by trained healthcare personnel.

The HOME HEALTHCARE ENVIRONMENT includes:

- the dwelling place in which a PATIENT lives;
- other places where PATIENTS are present both indoors and outdoors, excluding professional healthcare facility environments where OPERATORS with medical training are continually available when PATIENTS are present.

This International Standard does not apply to ME EQUIPMENT and ME SYSTEMS intended solely for use in the EMERGENCY MEDICAL SERVICES ENVIRONMENT, covered by IEC 60601-1-12 or solely for use in professional healthcare facilities covered by IEC 60601-1 without the additions of IEC 60601-1-12 or this collateral standard. Nonetheless, ME EQUIPMENT or ME SYSTEMS can be intended for multiple use environments, and as such, if also intended for use in the HOME HEALTHCARE ENVIRONMENT, are within the scope of this standard.

EXAMPLE ME EQUIPMENT or ME SYSTEMS intended for both the HOME HEALTHCARE ENVIRONMENT and the professional healthcare facility environment.

NOTE HOME HEALTHCARE ENVIRONMENT ME EQUIPMENT and ME SYSTEMS can frequently be used in locations with unreliable electrical sources and poor electrical grounding.

**1.2 Object**

The object of this collateral standard is to specify general requirements that are in addition to those of the general standard and to serve as the basis for particular standards.

**1.3 Related standards****1.3.1 IEC 60601-1**

For ME EQUIPMENT and ME SYSTEMS, this collateral standard complements IEC 60601-1.

When referring to IEC 60601-1 or to this collateral standard, either individually or in combination, the following conventions are used:

- "the general standard" designates IEC 60601-1 alone;
- "this collateral standard" designates IEC 60601-1-11 alone;
- "this standard" designates the combination of the general standard and this collateral standard.