

Australian Standard[®]

**Recommended procedures and
principles of glow discharge mass
spectrometry (GD-MS)**

This Australian Standard was prepared by Committee CH/16, Spectroscopy. It was approved on behalf of the Council of Standards Australia on 15 October 1997 and published on 5 January 1998.

The following interests are represented on Committee CH/16:

Australian Chamber of Commerce and Industry
CSIRO, Division of Coal and Energy Technology
CSIRO, Division of Material Science and Technology
Department of Defence (Australia)
National Association of Testing Authorities Australia
Queensland Health Scientific Services

Review of Australian Standards. *To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.*

Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications; this information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

This Standard was issued in draft form for comment as DR 96549.

STANDARDS AUSTRALIA

RECONFIRMATION

OF

AS 3685—1998

**Recommended procedures and principles of glow discharge mass spectrometry
(GD-MS)**

RECONFIRMATION NOTICE

Technical Committee CH-016 has reviewed the content of this publication and in accordance with Standards Australia procedures for reconfirmation, it has been determined that the publication is still valid and does not require change.

Certain documents referenced in the publication may have been amended since the original date of publication. Users are advised to ensure that they are using the latest versions of such documents as appropriate, unless advised otherwise in this Reconfirmation Notice.

Approved for reconfirmation in accordance with Standards Australia procedures for reconfirmation on 5 October 2016.

The following are represented on Technical Committee CH-016:

Australian Institute of Physics
Environmental Laboratory Industry Group
National Association of Testing Authorities Australia
National Measurement Institute

NOTES

Australian Standard[®]

**Recommended procedures and
principles of glow discharge mass
spectrometry (GD-MS)**

First published as AS 3685—1998.

PREFACE

This Standard was prepared by the Standards Australia Committee CH/16, Spectroscopy.

The guidelines given in this Standard are intended primarily for the development and application of a standard method which may be applied less rigidly in ordinary routine operation of glow discharge mass spectrometry. The guidelines are intended to be of general applicability, and not restricted to a given class of materials.

The standard practice of specifying instrumental suitability in terms of fundamental performance criteria is described. The quality of results obtainable from a given instrument can depend markedly on its optimization and correct operation.

The term 'informative' has been used in this Standard to define the application of the appendix to which it applies. An 'informative' appendix is only for information and guidance.

© Copyright — STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

CONTENTS

	<i>Page</i>
SECTION 1 SCOPE AND GENERAL	
1.1 SCOPE	4
1.2 PRINCIPLE	4
1.3 REFERENCED DOCUMENTS	4
1.4 DEFINITIONS	4
1.5 REAGENTS	5
SECTION 2 SAMPLE PREPARATION	
2.1 GENERAL	6
2.2 SAMPLE TYPE	6
2.3 SAMPLE GEOMETRY	6
2.4 SAMPLE PREPARATION	6
SECTION 3 INSTRUMENTATION	
3.1 GENERAL	8
3.2 ION SOURCES	8
3.3 ION ANALYSER	9
3.4 DETECTOR SYSTEM	9
3.5 VACUUM SYSTEM	9
3.6 DATA ACQUISITION AND CONTROL	9
SECTION 4 SYSTEM OPERATION	
4.1 SYSTEM PRECAUTIONS	12
4.2 OBTAINING A DISCHARGE	12
4.3 PRE-SPUTTERING	13
4.4 OPTIMIZING THE ION BEAM	13
4.5 ANALYSIS SET-UP	13
4.6 DATA ANALYSIS	13
4.7 INSTRUMENT PRINTOUT	14
4.8 DEPTH PROFILE ANALYSIS	14
4.9 INSTRUMENT PERFORMANCE	14
SECTION 5 CALIBRATION AND ANALYSIS	
5.1 GENERAL	15
5.2 CALIBRATION AND ANALYSIS CRITERIA	15
APPENDIX A SAFETY	17

STANDARDS AUSTRALIA

Australian Standard

**Recommended procedures and principles
of glow discharge mass spectrometry (GD-MS)**

SECTION 1 SCOPE AND GENERAL

1.1 SCOPE This Standard is a guide to the operation and recommendations for the use of glow discharge mass spectrometry (GD-MS).

NOTES:

- 1 This Standard should be read in conjunction with the instrument manufacturer's recommendations.
- 2 The general safety precautions listed in Appendix A and in AS 2243.1 and AS 2243.2 should be followed to minimize safety hazards.

1.2 PRINCIPLE In a glow discharge source electrical power is supplied between the sample (cathode) and the anode by a power supply typically operated at 0.5 to 2 kV and 1 to 30 mA. Argon (or neon, helium, nitrogen or oxygen) is introduced into the cell, and the pressure near the cathode surface is regulated to a few hundred pascals (Pa). The potential difference between the cathode and the anode extracts electrons from the cathode, causing ionization of the gas and the creation of a plasma (negative glow). Sample atoms which are sputtered by fast ions and neutrals diffuse into the plasma.

Ions formed in the glow discharge are extracted from the cell and pass into a mass spectrometer. The mass spectrometer is scanned and ions are sequentially transmitted to the detector. The ion current reaching the detector is measured for each isotope in turn and this information is stored in the computer system. Elemental concentrations are calculated by the instrument software using these ion currents, by comparison with those arising from the corresponding isotopes of reference materials.

1.3 REFERENCED DOCUMENTS The following documents are referred to in this Standard:

AS

- | | |
|--------|---|
| 1894 | Code of practice for the safe handling of cryogenic fluids |
| 1940 | The storage and handling of flammable and combustible liquids |
| 2243 | Safety in laboratories |
| 2243.1 | Part 1: General |
| 2243.2 | Part 2: Chemical aspects |
| 2243.5 | Part 5: Non-ionizing radiations |
| 2865 | Safe working in a confined space |

ISO

- | | |
|------|--|
| 3696 | Water for analytical laboratory use—Specification and test methods |
|------|--|

1.4 DEFINITIONS For the purpose of this Standard, the definitions below apply.

1.4.1 Abundance sensitivity—the contribution to peak intensity from the tail of a peak at an adjacent mass.