

Australian Standard™

**Electroacoustics—Equipment for the  
measurement of real-ear acoustical  
characteristics of hearing aids**

This Australian Standard was prepared by Committee AV-003, Acoustics Human Effects. It was approved on behalf of the Council of Standards Australia on 14 March 2003 and published on 31 March 2003.

---

The following are represented on Committee AV-003:

Association of Australian Acoustical Consultants  
Association of Consulting Engineers Australia  
Australian Acoustical Society  
Australian Chamber of Commerce and Industry  
Australian Hearing  
Department of Consumer & Employment Protection, WorkSafe Division, W.A.  
Department of Labour, New Zealand  
N.S.W. Rural Fire Service  
New South Wales Nurses Association  
New Zealand Audiological Society  
Royal Institution of Naval Architects  
Safety Institute of Australia (Incorporated)  
Victorian WorkCover Authority  
WorkCover New South Wales

---

#### **Keeping Standards up-to-date**

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about Standards can be found by visiting the Standards Australia web site at [www.standards.com.au](http://www.standards.com.au) and looking up the relevant Standard in the on-line catalogue.

Alternatively, the printed Catalogue provides information current at 1 January each year, and the monthly magazine, *The Australian Standard*, has a full listing of revisions and amendments published each month.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at [mail@standards.com.au](mailto:mail@standards.com.au), or write to the Chief Executive, Standards Australia International Ltd, GPO Box 5420, Sydney, NSW 2001.

---

Australian Standard™

**Electroacoustics—Equipment for the  
measurement of real-ear acoustical  
characteristics of hearing aids**

First published as AS IEC 61669—2003.

**COPYRIGHT**

© Standards Australia International

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia International Ltd  
GPO Box 5420, Sydney, NSW 2001, Australia

ISBN 0 7337 5168 7

## PREFACE

This Standard was prepared by the Australian members of the Joint Standards Australia/Standards New Zealand Committee AV-003, Acoustics Human Effects. After consultation with stakeholders in both countries, Standards Australia and Standards New Zealand decided to develop this Standard as an Australian, rather than an Australian/New Zealand Standard.

This Standard is identical with and has been reproduced from IEC 61669:2001, *Electroacoustics—Equipment for the measurement of real-ear acoustical characteristics of hearing aids*.

The objective of this Standard is to specify the general requirements for test equipment designed for use in measuring the real-ear acoustical characteristics of hearing aids and describes the terminology used.

As this Standard is reproduced from an International Standard, the following applies:

- (a) Its number does not appear on each page of text and its identity is shown only on the cover and title page.
- (b) In the source text ‘this International Standard’ should read ‘this Australian Standard’.
- (c) A full point substitutes for a comma when referring to a decimal marker.

This Standard provides for the use of the following Australian/New Zealand Standards in place of particular International Standards referenced herein:

<i>Reference to International Standard</i>		<i>Australian/New Zealand Standard</i>	
IEC		AS/NZS	
60601	Medical electrical equipment	3200	Medical electrical equipment
60601-1	Part 1: General requirements for safety	3200.1.0	General requirements for safety—Parent Standard
60601-1-2	Part 1-2: General requirements for safety—Collateral Standard: Electromagnetic compatibility—Requirements and tests	3200.1.2	Part 1.2: general requirements for safety—Collateral Standard: Electromagnetic compatibility—Requirements and tests
60601-1-4	Part 1-4: General requirements for safety—Collateral Standard: Programmable electrical medical systems	3200.1.4	Part 1.4: General requirements for safety—Collateral Standard: Programmable electrical medical systems
60645	Audiometers	1591	Acoustics—Instrumentation for audiometry
60645-1	Part 1: Pure tone audiometers	1591.1	Part 1: Reference zero for the calibration of pure-tone bone conduction audiometers
		1591.2	Part 2: Reference zero for the calibration of pure-tone audiometers
ISO		AS	
266	Acoustics—Preferred frequencies	2533	Acoustics—Preferred frequencies and band centre frequencies

## CONTENTS

INTRODUCTION .....	v
Clause	
1 Scope .....	1
2 Normative references .....	1
3 Terms and definitions .....	1
4 Limitations .....	5
5 General requirements .....	6
5.1 General .....	6
5.2 Safety requirements .....	6
5.3 Warm-up time .....	6
5.4 Environmental conditions .....	6
5.5 Power supply variation .....	6
5.6 Immunity to power and radio frequency fields .....	7
5.7 Interface connections .....	7
6 Test signal characteristics .....	7
6.1 Test signal .....	7
6.2 Sound source .....	7
6.3 Test signal level .....	7
6.4 Test signal level indication .....	7
6.5 Equalization .....	7
6.6 Frequency .....	8
6.7 Harmonic distortion .....	8
6.8 Equipment for automatic sweep frequency recording .....	8
6.9 Integrity of test signal .....	8
7 Sound measurement .....	8
7.1 Probe microphone measurement .....	8
7.2 Noise floor of probe microphone measurement .....	8
7.3 Attenuation of probe microphone to external signals .....	8
8 Data acquisition, analysis and presentation .....	9
8.1 General .....	9
8.2 Output indication .....	9
8.3 Graphical printout .....	9
9 Calibration .....	9
10 Demonstration of conformity with specifications and test procedures .....	9
10.1 General .....	9
10.2 Environmental conditions and power supply variation .....	10
10.3 Immunity to power and radio frequency fields .....	10
10.4 Test signals .....	10
10.4.1 Signal accuracy .....	10
10.4.2 Accuracy of control .....	10

Clause	Page
10.5 Sound measurement.....	11
10.5.1 Measurement accuracy .....	11
10.5.2 Noise floor of probe microphone measurement .....	11
10.5.3 Attenuation of probe microphone to external signals .....	11
11 Maximum permitted expanded uncertainty of measurements.....	11
12 Marking and instruction manual.....	12
12.1 Marking.....	12
12.2 Software version.....	12
12.3 Instruction manual .....	12
Bibliography .....	13

## INTRODUCTION

The performance characteristics of hearing aids in actual use can differ significantly from those determined in accordance with standards such as IEC 60118-0, and IEC 60118-7, due to differing acoustic influence and coupling presented by individual ears. Measuring methods that take into account the acoustic coupling and the acoustic influence of the individual wearer on the performance of hearing aids are therefore important in the fitting of these devices. Such measuring methods have come to be known as “real-ear measurements” and are sometimes performed clinically in less than ideal acoustic environments. The accuracy and repeatability of measurements made under such conditions are complex functions of the sound field, the test environment, the nature of the test signal, the hearing aid under evaluation, the method of test signal control, the location of the sound source, the nature of the data acquisition, analysis and presentation as well as the degree of subject movement permitted.

This International Standard specifies performance requirements separate from the test requirements to show conformity. Conformance to the specifications in this International Standard is demonstrated only when the result of a measurement, extended by the actual expanded uncertainty of measurement of the testing laboratory, lies fully within the tolerances specified in this International Standard extended by the values for  $U_{\max}$  given in table 1.

NOTES

## AUSTRALIAN STANDARD

**ELECTROACOUSTICS –  
EQUIPMENT FOR THE MEASUREMENT  
OF REAL-EAR ACOUSTICAL CHARACTERISTICS OF HEARING AIDS****1 Scope**

This International Standard specifies the general requirements for test equipment designed for use in measuring the real-ear acoustical characteristics of hearing aids and describes the terminology used.

The purpose of this International Standard is to ensure that measurements of real-ear acoustical characteristics of a hearing aid on a given human ear, performed with different test equipment which comply with this International Standard using methods described in ISO 12124, shall give substantially the same results.

**2 Normative references**

The following normative documents contain provisions which, through reference in this text, constitute provisions of this International Standard. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this part of the ISO/IEC Directives are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

IEC 60601-1, *Medical electrical equipment – Part 1: General requirements for safety*

IEC 60601-1-2, *Medical electrical equipment – Part 1: General requirements for safety – 2. Collateral standard: Electromagnetic compatibility – Requirements and tests*

IEC 60601-1-4, *Medical electrical equipment – Part 1: General requirements for safety – 4. Collateral standard: Programmable electrical medical systems*

IEC 60645-1, *Audiometers – Part 1: Pure tone audiometers*

IEC 60942, *Electroacoustics – Sound calibrators*

ISO 266, *Acoustics – Preferred frequencies*

**3 Terms and definitions**

For the purpose of this International Standard, the following terms and definitions apply:

**3.1****test signal**

acoustic signal at the field reference point