

3536

(based on ISO 6142 - 1981)

AS 3536—1988

under revision see DR 97359.

WITHDRAWN:
19980705
SUPERSEDED BY:
AS 3536.1-1998



Standards
Association of
Australia



Australian Standard® 3536—1988

REFERENCE GASES— PREPARATION OF GRAVIMETRIC STANDARDS



This Australian Standard was prepared by Committee CH/27, Reference Gases. It was approved on behalf of the Council of the Standards Association of Australia on 6 January 1988 and published on 5 April 1988.

The following interests are represented on Committee CH/27:

Australian Environment Council
Australian Gas Association
Confederation of Australian Industry
CSIRO, Division of Fossil Fuels
Department of Environment and Planning, S.A.
Department of Industrial Relations and Employment, N.S.W.
Environment Protection Authority of Victoria
Federal Chamber of Automotive Industries
National Association of Testing Authorities, Australia
Royal Australian Chemical Institute
State Pollution Control Commission, N.S.W.

Review of Australian Standards. To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up-to-date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all SAA publications will be found in the Catalogue of SAA Publications; this information is supplemented each month by SAA's journal 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of the Association, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

This Standard was issued in draft form for comment as DR 86088.

AUSTRALIAN STANDARD

**REFERENCE GASES—
PREPARATION OF
GRAVIMETRIC STANDARDS**

AS 3536—1988

First published as AS 3536—1988.

**PUBLISHED BY THE STANDARDS ASSOCIATION OF AUSTRALIA
STANDARDS HOUSE, 80 ARTHUR ST, NORTH SYDNEY, N.S.W.**

ISBN 0 7262 4909 2

PREFACE

This Standard was prepared by the Association's Committee on Reference Gases, under the direction of the Chemical Standards Board, to standardize a gravimetric procedure for the preparation of reference gases.

This Standard method is based on ISO 6142, *Gas analysis—Preparation of calibration gas mixtures—Weighing methods*.

CONTENTS

	<i>Page</i>
1 SCOPE	3
2 APPLICATION	3
3 REFERENCED DOCUMENTS	3
4 PRINCIPLE	3
5 APPARATUS	3
6 GASES	3
7 PROCEDURE	3
8 CALCULATIONS	4
9 ESTIMATION OF ERRORS	4
10 CERTIFICATION OF COMPOSITION	4
APPENDICES	
A BUOYANCY CORRECTION	5
B TYPICAL CALCULATION FOR GRAVIMETRIC GAS MIXTURE PREPARATION	6
C ORDERING INFORMATION FOR GRAVIMETRIC STANDARDS	7
D REFERENCES	8

© Copyright — STANDARDS ASSOCIATION OF AUSTRALIA 1988

Users of Standards are reminded that copyright subsists in all SAA publications. Except where the Copyright Act otherwise allows, no part of this publication may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing of the Standards Association of Australia. Requests for permission should be directed to the Head Office of the Association. Where such requests relate to the reproduction of the whole or a substantial part of any Standard, permission may be conditional on an appropriate royalty payment.

STANDARD ASSOCIATION OF AUSTRALIA

Australian Standard

REFERENCE GASES—PREPARATION OF GRAVIMETRIC STANDARDS

1 SCOPE. This Standard sets out a method for the preparation of gravimetric standards for which the accuracy of each gas component is better than 0.5 percent relative or 0.02 percent absolute on a mole to mole basis, whichever is smaller and whatever the values of the selected concentrations.

NOTE: Appendix C sets out information which should be provided when ordering reference gases to be supplied to this Standard.

2 APPLICATION. The method is applicable only to gaseous components which do not react with each other or with the cylinder walls, and to condensable components which remain totally vaporized in the cylinder at all pressures and temperatures anticipated during the use of the contents.

NOTE: Retrograde condensation may occur with some hydrocarbon gas mixtures.

3 REFERENCED DOCUMENTS. The following Standards are referred to in this Standard:

- AS
 1777 Aluminium cylinders for compressed gases—Seamless—0.1 kg to 130 kg
 2030 SAA Gas Cylinders Code
 2473 Valves for compressed gas cylinders (threaded outlet)
 B114 Alloy steel cylinders for the storage and transport of permanent gases and high pressure liquefiable gases

4 PRINCIPLE. The compositions of gas mixtures are determined from the defined units of mass and molecular mass. Using an evacuated cylinder of known mass, standard mixtures are prepared by adding and weighing each gas. The mass of each gas is found by difference, and the number of moles of each component is calculated. The mole fraction is calculated from the total number of moles.

To avoid having to weigh very small masses of gas, a lower mass limit is fixed for each of the components in the final gas mixture. Where the mass required to produce the desired concentration is less than this limiting value, a known quantity of gas mixture previously prepared in accordance with this Standard is diluted with a known mass of a given gas.

5 APPARATUS. Materials used for surfaces with which the gases may come into contact shall not promote changes in chemical composition of the gases. All devices (regulators, valves, filling connections) shall be leak tight.

5.1 Balance. With a minimum ratio of capacity to resolution (R) of 10^5 and minimum weighing accuracy of $\pm R$. The balance shall be calibrated with balance weights at least equivalent to Class F_1 (Ref. 1.).

5.2 Balance weights. At least equivalent to Class F_1 (Ref. 1.).

5.3 Barometer. With a reading resolution of 10 Pa (1 mm Hg is satisfactory) and calibrated against current meteorological data.

5.4 Gas cylinders. Complying with either AS 1777, AS 2030 or AS B114, as appropriate.

5.5 Valves. Manufactured in accordance with AS 2473.

5.6 Wet and dry bulb thermometer. Accurate to at least 0.1°C .

5.7 Vacuum pump. Capable of maintaining a vacuum of less than 100 Pa.

5.8 Manifold. A suitable arrangement to allow interconnection of the gas cylinders, vacuum pump, vents and gauges.

6 GASES. All component gases used in primary preparation (before dilution) shall be subject to strict quality control and the purity shall be accurately known to at least ± 0.1 percent absolute.

7 PROCEDURE.

7.1 Preparation of cylinder. The cylinder shall be prepared as follows:

- The external surface of the cylinder shall be free from dust, moisture, grease and particles. Cylinders suspected of being internally contaminated shall not be used.
- Fill the cylinder to full pressure and check for leaks.
- Evacuate the cylinder to at least 100 Pa.
- Fill with major component gas to a pressure greater than 200 kPa.
- Vent the cylinder.
- Repeat Steps (c), (d) and (e).
- Evacuate the cylinder to less than 100 Pa.

7.2 Weighing the cylinder. The weighing procedure shall be as follows:

- Check the mass of the cylinder on a rough scale.
- Zero the balance.
- Select the balance weight (5.2) corresponding to the mass of the cylinder obtained in Step (a).
- Carefully place the cylinder on the centre of the pan and release the beam. Record the mass of the cylinder.
- Remove the cylinder and check the balance zero. This should agree with the zero in Step (b) within the resolution of the balance; if not, re-zero the balance.
- Repeat Step (d).