

Australian Standard[®]

Acoustics—Audiometric test methods

Part 3: Speech audiometry



This Australian Standard® was prepared by Committee AV-003, Acoustics Human Effects. It was approved on behalf of the Council of Standards Australia on 2 December 2008. This Standard was published on 9 March 2009.

The following are represented on Committee AV-003:

- Association of Accredited Certification Bodies
 - Association of Australian Acoustical Consultants
 - Association of Consulting Engineers Australia
 - Audiological Society of Australia
 - Australasian Faculty of Occupational & Environmental Medicine
 - Australian Acoustical Society
 - Australian Chamber of Commerce and Industry
 - Australian Council of Trade Unions
 - Australian Industry Group
 - Australian Society of Otolaryngological Head and Neck Surgery
 - Department of Defence, Australia
 - Department of Labour, New Zealand
 - Department of the Premier and Cabinet, SA
 - Engineers Australia
 - Environment Protection and Heritage Council
 - Institute of Marine Engineers Australia/New Zealand Division
 - National Acoustic Laboratories
 - New South Wales Nurses Association
 - New Zealand Audiological Society
 - Office of the Australian Safety and Compensation Council
 - Royal Institute of Naval Architects
 - Safety Institute of Australia
 - University of Queensland
 - Victorian WorkCover Authority
 - WorkCover New South Wales
 - WorkSafe Division Department of Commerce Government of WA
-

This Standard was issued in draft form for comment as DR 08072.

Standards Australia wishes to acknowledge the participation of the expert individuals that contributed to the development of this Standard through their representation on the Committee and through the public comment period.

Keeping Standards up-to-date

Australian Standards® are living documents that reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued.

Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments that may have been published since the Standard was published.

Detailed information about Australian Standards, drafts, amendments and new projects can be found by visiting www.standards.org.au

Standards Australia welcomes suggestions for improvements, and encourages readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at mail@standards.org.au, or write to Standards Australia, GPO Box 476, Sydney, NSW 2001.

Australian Standard[®]

Acoustics—Audiometric test methods

Part 3: Speech audiometry

First published as AS ISO 5253.3—2009.

COPYRIGHT

© Standards Australia

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia GPO Box 476, Sydney, NSW 2001, Australia

ISBN 0 7337 9045 3

PREFACE

This Standard was prepared by the Joint Standards Australia/Standards New Zealand Committee AV-003, Acoustics Human Effects. After consultation with stakeholders in both countries, Standards Australia and Standards New Zealand decided to develop this Standard as an Australian Standard rather than an Australian/New Zealand Standard.

This Standard is identical with, and has been reproduced from ISO 8253-3:1996, *Acoustics—Audiometric test methods, Part 3: Speech audiometry*.

The objective of this Standard is to specify procedures and requirements for speech audiometry where the recorded test material is presented by air conduction through an earphone, by bone conduction through a bone vibrator, or from a loudspeaker for sound field audiometry.

Committee AV-003 agreed that the following sentences should be added:

- (a) Clause 9.1, paragraph 1—Guidance on recent exposure can be obtained from AS/NZS 1269.4.
- (b) Clause 17, paragraph 1—It is recommended that for Stages B and C the calibration be recorded and any non-compliance noted.

As this Standard is reproduced from an International Standard, the following applies:

- (i) Its number does not appear on each page of text and its identity is shown only on the cover and title page.
- (ii) In the source text ‘this part of ISO 8253’ should read ‘this Australian Standard’.
- (iii) Substitute a full point for a comma as a decimal marker.

References to International Standards should be replaced by references to Australian or Australian/New Zealand Standards, as follows:

<i>Reference to International Standard</i>		<i>Australian Standard</i>	
ISO		AS	
266	Acoustics—Preferred frequencies	2533	Acoustics—Preferred frequencies and band centre frequencies
		AS ISO	
8253	Acoustics—Audiometric test methods	8253	Acoustics—Audiometric test methods
8253-1	Part 1: Basic pure tone air and bone conduction threshold audiometry	8253.1	Part 1: Basic pure tone air and bone conduction threshold audiometry
8253-2	Part 2: Sound field audiometry with pure tone and narrow band test signals	8253.2	Part 2: Sound field audiometry with pure tone and narrow-band test signals
IEC		AS IEC	
645	Audiometers	60645	Electroacoustics—Audiological equipment
645-1	Part 1: Pure tone audiometers	60645.1	Part 1: Pure-tone audiometers (IEC 60645-1:2001, MOD)
645-2	Part 2: Equipment for speech audiometry	60645.2	Part 2: Equipment for speech audiometry

Only international references that have been adopted as Australian or Australian/New Zealand Standards have been listed.

CONTENTS

	<i>Page</i>
1 Scope	1
2 Normative references.....	1
3 Definitions.....	2
4 Requirements for recorded speech material	5
5 Speech signal level	7
6 Masking noise level.....	7
7 Ambient sound pressure levels in test room	7
8 Sound field speech audiometry	7
9 Preparation and instruction of test subject	8
10 Subject's response mode	8
11 Determination of speech detection threshold level.....	9
12 Determination of speech recognition threshold level	9
13 Determination of speech recognition scores	11
14 Contralateral masking	11
15 Speech recognition with competing sound.....	12
16 Format of speech audiogram.....	14
17 Maintenance and calibration of equipment.....	15
Annexes	
A Live-voice testing	17
B Bibliography	18

INTRODUCTION

Speech audiometry is used in connection with diagnostic evaluation, audiological rehabilitation and the evaluation of hearing disability. The purpose of a particular test is to assist in the choice of speech test material and mode of presentation.

AUSTRALIAN STANDARD

Acoustics — Audiometric test methods —

Part 3: Speech audiometry

1 Scope

This part of ISO 8253 specifies procedures and requirements for speech audiometry with the recorded test material being presented by air conduction through an earphone, by bone conduction through a bone vibrator, or from a loudspeaker for sound field audiometry. Methods for using noise either for masking the non-test ear or as a competing sound are described. All test procedures are based on the use of open-set test material.

This part of ISO 8253 also contains requirements on recorded speech material and recommended procedures for the maintenance and calibration of speech audiometric equipment.

Some test subjects, for example children, may require amended test procedures not specified in this part of ISO 8253.

It is recognized that standards for speech audiometry cannot be met using live voice speech tests, however, guidance is given in an informative annex so as to make such tests as reliable as possible.

This part of ISO 8253 does not specify the contents of the speech material because of the variety of languages. The type of test material may significantly influence the result of speech audiometry.

Specialized tests such as those used for evaluating directional hearing and dichotic hearing are not included in this part of ISO 8253.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this part of ISO 8253. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this part of ISO 8253 are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 266:—¹⁾, *Acoustics — Preferred frequencies*.

ISO 8253-1:1989, *Acoustics — Audiometric test methods — Part 1: Basic pure tone air and bone conduction threshold audiometry*.

ISO 8253-2:1992, *Acoustics — Audiometric test methods — Part 2: Sound field audiometry with pure tone and narrow band test signals*.

IEC 645-1:1992, *Audiometers — Part 1: Pure-tone audiometers*.

IEC 645-2:1993, *Audiometers — Part 2: Equipment for speech audiometry*.

1) To be published. (Revision of ISO 266:1975)