

Australian Standard<sup>®</sup>

---

**One-mark pipettes**

---

This Australian Standard was prepared by Committee CH/1, Laboratory Glassware and Related Apparatus. It was approved on behalf of the Council of Standards Australia on 1 December 1994 and published on 5 April 1995.

---

The following interests are represented on Committee CH/1:

Australian Chamber of Commerce and Industry  
Australian Government Analytical Laboratories  
National Association of Testing Authorities Australia  
National Standards Commission  
N.S.W. Agriculture  
Royal Australian Chemical Institute  
Royal College of Pathologists Australasia  
Royal Melbourne Institute of Technology  
University of Sydney

---

**Review of Australian Standards.** *To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.*

*Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications; this information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.*

*Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.*

Australian Standard<sup>®</sup>

---

**One-mark pipettes**

---

First published as AS R16—1953.  
Second edition 1962 (being endorsement of  
BS 1583—1961 with amendments).  
Revised and redesignated AS 2166—1978.  
Second edition 1995.

PUBLISHED BY STANDARDS AUSTRALIA  
(STANDARDS ASSOCIATION OF AUSTRALIA)  
1 THE CRESCENT, HOMEBUSH, NSW 2140

ISBN 0 7262 9537 X

## PREFACE

This Standard was prepared by the Standards Australia Committee CH/1, Laboratory Glassware and Related Apparatus, under the direction of the Multitechnics Standards Policy Board, to supersede the 1978 edition.

The objective of this Standard is to provide a specification for one-mark pipettes required for general use in laboratories.

This Standard is based on ISO 648, *Laboratory glassware—One-mark pipettes*. It varies from the 1978 edition of AS 2166 in that a correction has been made to 100 mL pipette dimensions in Table 2, the range of pipettes covered has been extended to 200 mL, and Clause 8, Clause 9 and Table 4 have been modified to cover pipettes with a 15 s waiting time. The format of AS 2166 has been retained to assist Australian manufacturers.

---

## CONTENTS

	<i>Page</i>
1 SCOPE . . . . .	3
2 REFERENCED DOCUMENTS . . . . .	3
3 DEFINITIONS . . . . .	3
4 NOMINAL CAPACITIES AND TOLERANCES ON CAPACITIES . . . . .	3
5 CONSTRUCTION . . . . .	3
6 GRADUATION LINE . . . . .	6
7 DELIVERY TIME . . . . .	6
8 WAITING TIME . . . . .	6
9 MARKING . . . . .	8
10 COLOUR CODING . . . . .	8

© Copyright — STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

# STANDARDS AUSTRALIA

## Australian Standard

### One-mark pipettes

**1 SCOPE** This Standard sets out requirements for a series of one-mark pipettes, suitable for general laboratory purposes. Two classes of accuracy are specified, Class A pipettes being of higher accuracy than Class B pipettes.

NOTE: The method of verification and notes for the use of one-mark pipettes are given in AS 2162.

**2 REFERENCED DOCUMENTS** The following documents are referred to in this Standard:

AS

2162 Verification and use of volumetric glassware

ISO

1769 Laboratory glassware—Pipettes—Colour coding

BS

1797 Tables for use in the calibration of volumetric glassware

**3 DEFINITIONS** For the purpose of this Standard, the definitions below apply.

**3.1 Capacity** The volume of water at 20°C, expressed in millilitres, delivered by the pipette at 20°C, when it is filled to the graduation line and emptied in accordance with the method given in AS 2162.

**3.2 Delivery time** The time occupied by the descent of the water meniscus from the graduation line to the point at which it appears to come to rest in the jet.

**3.3 Waiting time** The time allowed, after the delivery time, to ensure that the total amount of liquid for which the pipette is calibrated has been delivered.

**4 NOMINAL CAPACITIES AND TOLERANCES ON CAPACITIES** The nominal capacities and tolerances on the capacities of one-mark pipettes shall be as specified in Table 1.

## 5 CONSTRUCTION

**5.1 Material** The pipettes shall be constructed of glass having suitable chemical and thermal properties, shall be as free as possible from visible defects and shall be reasonably free from internal stress.

NOTES:

- 1 BS 1797 is applicable to pipettes manufactured almost exclusively from soda glass or borosilicate glass. Corrections for calibrating pipettes made of other types of glass may be obtained from the tables by interpolation or extrapolation.
- 2 Pipettes made from borosilicate glass are unlikely to be as free from visible defects as those made from soda glass.
- 3 Stress is commonly referred to as 'strain' in the glass manufacturing industry.

### 5.2 Shape

**5.2.1 0.5 mL pipettes** Pipettes of 0.5 mL capacity shall consist of a straight tube with a jet at the lower end.