

Australian Standard™

**Laboratory glassware—Interchangeable
spherical ground glass joints**



This Australian Standard was prepared by Committee CH-001, Laboratory Glassware and Related Apparatus. It was approved on behalf of the Council of Standards Australia on 14 October 2005.

This Standard was published on 2 November 2005.

The following are represented on Committee CH-001:

National Association of Testing Authorities Australia
National Measurement Institute
Royal Australian Chemical Institute
Royal College of Pathologists of Australasia
Science Industry Australia
University of New South Wales
University of Sydney

Keeping Standards up-to-date

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about Standards can be found by visiting the Standards Web Shop at www.standards.com.au and looking up the relevant Standard in the on-line catalogue.

Alternatively, the printed Catalogue provides information current at 1 January each year, and the monthly magazine, *The Global Standard*, has a full listing of revisions and amendments published each month.

Australian Standards™ and other products and services developed by Standards Australia are published and distributed under contract by SAI Global, which operates the Standards Web Shop.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at mail@standards.org.au, or write to the Chief Executive, Standards Australia, GPO Box 476, Sydney, NSW 2001.

This Standard was issued in draft form for comment as DR 05079.

Australian Standard™

**Laboratory glassware—Interchangeable
spherical ground glass joints**

Originated as AS 2410—1981.
Second edition 2005.

COPYRIGHT

© Standards Australia

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia, GPO Box 476, Sydney, NSW 2001, Australia

ISBN 0 7337 6975 6

PREFACE

This Standard was prepared by the Standards Australia Committee CH-001, Laboratory Glassware and Related Apparatus to supersede AS 2410—1981, *Interchangeable spherical ground glass joints*. It is based on, but not equivalent to, ISO 641:1975 (confirmed in 2000), *Laboratory glassware—Interchangeable spherical ground joints*.

This Standard is intended to ensure interchangeability between spherical ground glass joints, irrespective of where they are manufactured.

Limiting dimensions for the ground glass zone and the external diameter of adjacent tubing, which will ensure interchangeability, are given in Table 1. The finish of the ground surface is also specified, and this is similar to the finish specified in AS 2409, *Laboratory glassware—Interchangeable conical ground glass joints*.

A conventional designation is adopted, which consists of the code letter ‘S’ in conjunction with the approximate spherical diameter of the joint in millimetres. The bore diameter which may be employed with each size of joint is controlled, for the purpose of this Standard, only the maximum diameter at the narrow end of the ground zone.

The objective of this revision is to reconfirm the Standard with minor changes including the following:

- (a) Referenced documents have been updated.
- (b) The Standard has been brought into line with current editorial practices.

The terms ‘normative’ and ‘informative’ have been used in this Standard to define the application of the appendix to which they apply. A ‘normative’ appendix is an integral part of a Standard, whereas an ‘informative’ appendix is only for information and guidance.

CONTENTS

	<i>Page</i>
1 SCOPE.....	4
2 REFERENCED DOCUMENTS.....	4
3 NOMINAL SPHERICAL GROUND JOINT DIAMETER.....	4
4 DIMENSIONS.....	4
5 DIAMETER OF TUBING	4
6 SURFACE FINISH OF GROUND GLASS JOINTS	5
7 TESTING OF INNER AND OUTER COMPONENTS (BALL AND CUP) OF SPHERICAL JOINTS	5
8 MARKING	5
 APPENDICES	
A LEAKAGE TEST FOR SPHERICAL JOINTS	7
B SIZE DESIGNATIONS OF SPHERICAL GROUND GLASS JOINTS	10

STANDARDS AUSTRALIA**Australian Standard****Laboratory glassware—Interchangeable spherical ground glass joints****1 SCOPE**

This Standard specifies the essential geometric requirements for interchangeability in relation to a series of 10 spherical ground glass joints for incorporation in laboratory glassware. The joints range in size from a nominal diameter of 7 mm to a nominal diameter of 102 mm.

NOTE: Spherical ground glass joints are normally assembled with the use of an appropriate clamp. Details of clamps are not included in this Standard.

2 REFERENCED DOCUMENTS

The following documents are referred to in this Standard.

AS

1349 Bourdon tube pressure and vacuum gauges

HB 47 Dimensioning and tolerancing to AS 1100.101 and AS 1100.201

BS

2761 Spherical ground joints

ASTM

E677-94 Specification for interchangeable spherical-ground joints

3 NOMINAL SPHERICAL GROUND JOINT DIAMETER

The diameter of the ground spherical zone shall conform to the dimensions specified in Table 1, columns 2, 3 and 4.

NOTE: The tolerances ensure that the diameter of the inner component (or ball member) is not greater than the nominal diameter, and that the diameter of the outer component (or cup member) is not less than the nominal diameter.

4 DIMENSIONS

The diameter at the wide end of the ground zone shall be not less than the corresponding dimension given in Table 1, column 5, and the diameter at the narrow end of the ground zone shall be not greater than the corresponding dimension given in Table 1, column 6.

NOTE: The relationship of the dimensions is illustrated in Figure 1.

5 DIAMETER OF TUBING

The external diameter of tubing adjacent to the joint shall not exceed the dimensions given in Table 1, column 7.

NOTE: It is important to limit the external diameter of the tubing in order to facilitate interchangeability of appropriate clamps.