

Australian/New Zealand Standard™

**Ultrasonics—Pulsed Doppler diagnostic
systems—Test procedures to determine
performance**



Standards Australia



STANDARDS
NEW ZEALAND
Te Kaitiaki Take Kōwhiri

AS/NZS 4966:2002

This Joint Australian/New Zealand Standard was prepared by Joint Technical Committee HE-003, Medical Electrical Equipment. It was approved on behalf of the Council of Standards Australia on 28 June 2001 and on behalf of the Council of Standards New Zealand on 5 October 2001. It was published on 4 January 2002.

The following interests are represented on Committee HE-003:

Australasian College of Physical Scientists and Engineers in Medicine
Australasian Society for Ultrasound in Medicine
Australian Dental Association
Australian Institute of Radiography
Australian Radiation Protection and Nuclear Safety Agency
Australian Society of Anaesthetists
Australian and New Zealand College of Anaesthetists
College of Biomedical Engineering Institution of Engineers Australia
Commonwealth Department of Health and Aged Care
Medical Industry Association of Australia
Ministry of Economic Development, New Zealand
Royal Australasian College of Radiologists

Additional interests participating in the preparation of this Standard:

CSIRO National Measurement Laboratory
CSIRO Telecommunications and Industrial Physics
National Radiation Laboratory, New Zealand
N.S.W. Department of Public Works and Services
University of Queensland
University of Sydney

Keeping Standards up-to-date

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about joint Australian/New Zealand Standards can be found by visiting the Standards Australia web site at www.standards.com.au or Standards New Zealand web site at www.standards.co.nz and looking up the relevant Standard in the on-line catalogue.

Alternatively, both organizations publish an annual printed Catalogue with full details of all current Standards. For more frequent listings or notification of revisions, amendments and withdrawals, Standards Australia and Standards New Zealand offer a number of update options. For information about these services, users should contact their respective national Standards organization.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Please address your comments to the Chief Executive of either Standards Australia International or Standards New Zealand at the address shown on the back cover.

Australian/New Zealand Standard™

Ultrasonics— Pulsed Doppler diagnostic systems—Test procedures to determine performance

First published as AS/NZS 4966:2002.

COPYRIGHT

© Standards Australia/Standards New Zealand

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Jointly published by Standards Australia International Ltd, GPO Box 5420, Sydney, NSW 2001 and Standards New Zealand, Private Bag 2439, Wellington 6020

ISBN 0 7337 3485 5

PREFACE

This Standard was prepared by the Joint Standards Australia/Standards New Zealand Committee HE-003, Medical Electrical Equipment.

This Standard is identical with and has been reproduced from IEC TS 61895:1999, *Ultrasonics—Pulsed Doppler diagnostic systems—Test procedures to determine performance*.

As this publication has been reproduced from an international Standard, the following modifications apply:

- (a) Its number does not appear on each page and its identity is shown on the cover and title page.
- (b) The words ‘this Australian/New Zealand Standard’ should replace the words ‘this technical specification’ wherever they appear.
- (c) Substitute a full point (.) for a comma (,) where it appears as a decimal marker.

Some pages of the original, which relate to IEC administrative matters, are omitted from this edition.

The term ‘normative’ has been used in this Standard to define the application of the annex or appendix to which it applies. A ‘normative’ annex or appendix is an integral part of a Standard.

The references to international Standards should be replaced by references to the following Australian or Joint Australian/New Zealand Standards:

<i>Reference to International Standard</i>	<i>Australian/New Zealand Standard</i>
IEC	AS/NZS
61206 Ultrasonics—Continuous-wave Doppler systems—Test procedures	4188 Ultrasonics—Continuous-wave Doppler systems—Test procedures

Any International Standards not listed do not have an Australian/New Zealand equivalent.

CONTENTS

	<i>Page</i>
INTRODUCTION	v
Clause	
1 Scope	1
2 Normative references	1
3 Definitions	1
4 Symbols	6
5 Overall tests of complete systems	6
5.1 General considerations	6
5.1.1 Types of pulsed Doppler ultrasound systems	6
5.1.2 Worst-case conditions	7
5.1.3 Doppler beam axes	8
5.1.4 Probe/target distance variation and measurement	8
5.2 Initial conditions	9
5.2.1 Power supply	9
5.2.2 Target movement direction	9
5.2.3 Propagation medium	9
5.2.4 Penetration depth	10
5.2.5 Working depth	10
5.2.6 Focusing	10
5.2.7 Working Doppler angle	10
5.2.8 Wall-thump filter cut-off frequency	11
5.2.9 Transmitter output power	11
5.2.10 Working pulse repetition frequency (PRF)	11
5.2.11 Doppler (receiver) gain	11
5.2.12 Test frequency	11
5.2.13 Working sample volume length	11
5.2.14 Doppler signal power measurement	11
5.3 Zero signal noise level	11
5.4 Doppler frequency response	12
5.4.1 Frequency response range	12
5.4.2 Deviation from flat response	12
5.4.3 Large signal performance	12
5.5 Spatial response	13
5.5.1 Sample volume response	13
5.6 Sample volume position registration error	14
5.7 Beam position and orientation	15
5.8 Intrinsic broadening	16
5.9 Dead zone	16
5.10 Acoustic working frequency	16
5.11 Flow direction separation	16
5.12 Velocity estimation accuracy	16

	<i>Page</i>
5.13 Volume flow estimation accuracy.....	16
5.14 Maximum, mean, mode and median frequency estimation accuracy.....	17
5.15 Velocity waveform indices estimation accuracy.....	19
6 Doppler test objects	20
6.1 Test objects	20
6.2 Electronic test object	20
 Annex A (normative) Description of pulsed Doppler ultrasound systems	 22
A.1 Single-channel system	22
A.2 Multi-channel system.....	23
A.3 Aliasing.....	23
A.4 Duplex and triplex scanners	23
 Bibliography	 26

INTRODUCTION

Pulsed ultrasonic Doppler flowmeters and velocimeters are widely used in clinical practice, usually in combination with real-time **B-mode** imaging and colour-flow imaging instruments. The device periodically transmits pulses of ultrasound from an ultrasound transducer and measures the Doppler shift in the frequency of ultrasound reflected and scattered from moving tissues. This Doppler shift is proportional to the component of reflector or scatterer velocity along the ultrasound beam. By looking for Doppler shifts in the received signal at specific times after transmission (range-gating), the device can be used to determine the variation of tissue velocity with distance along the ultrasound beam. The device is sensitive to movement only within a region of the beam called the sample volume. The position of the sample volume along the beam may be adjusted by altering the delay between transmission and range-gating. Multi-channel devices have a number of sample volumes operating simultaneously.

The pulsed ultrasonic device is most commonly used to investigate blood flow when the ultrasound is scattered from red blood cells.

This technical specification describes a range of tests which may be used to measure performance and the test objects required. In many cases, the test method and test object have been described in IEC 61206 and in these cases reference is simply made to this document. Other tests and test objects are described in [1] and [2]. The test methods may be considered as falling into one of the following three categories. The first is routine quality control tests that can be carried out by a clinician or technologist to ensure that the system is working adequately or has adequate sensitivity. The second is more elaborate test methods, conducted less frequently, when, for example, the system is suspected of malfunctioning. The third represents tests that would be carried out by a manufacturer on complete systems in order to guarantee compliance with specification.

NOTES

AUSTRALIAN/NEW ZEALAND STANDARD

Ultrasonics—Pulsed Doppler diagnostic systems—Test procedures to determine performance**1 Scope**

This technical specification describes

- test methods for measuring the performance of pulsed **Doppler ultrasound systems**;
- **Doppler test objects** for carrying out these tests;

and applies to

- tests made on an overall pulsed **Doppler ultrasound system**, a system which is not disassembled or disconnected;
- tests made on pulsed **Doppler ultrasound systems** whether they are stand-alone or as part of another ultrasound instrument.

Electrical safety, acoustic output and electromagnetic compatibility (EMC) are not covered in this technical specification.

The workload to perform all described tests is, in general, prohibitive. It is intended that a subset of the described tests is adopted for regular use. However, experience to give guidance for selection has still to be gathered and will be the subject of ongoing work.

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this technical specification. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this technical specification are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of IEC and ISO maintain registers of currently valid International Standards.

IEC 61102:1991, *Measurement and characterisation of ultrasonic fields using hydrophones in the frequency range 0,5 MHz to 15 MHz*

IEC 61206:1993, *Ultrasonics – Continuous-wave Doppler systems – Test procedures*

IEC 61390:1996, *Ultrasonics – Real-time pulse-echo systems – Test procedures to determine performance specifications*

3 Definitions

For the purposes of this technical report, the following definitions apply.

3.1**6 dB spectral width**

width of a frequency spectrum between the frequencies at which the spectral power is 6 dB less than the maximum power