

Australian/New Zealand Standard™

**Management programs for medical
equipment**



AS/NZS 3551:2012

This Joint Australian/New Zealand Standard was prepared by Joint Technical Committee HE-003, Medical Electrical Equipment. It was approved on behalf of the Council of Standards Australia on 19 July 2012 and on behalf of the Council of Standards New Zealand on 10 July 2012.

This Standard was published on 31 October 2012.

The following are represented on Committee HE-003:

Australasian College of Physical Scientists and Engineers In Medicine
Australasian Society for Ultrasound in Medicine
Australian and New Zealand College of Anaesthetists
Australian Dental Association
Australian Radiation Protection and Nuclear Safety Agency
Australian Society of Anaesthetists
Canterbury District Health Board
College of Biomedical Engineering Engineers Australia
Department of Defence, Australia
Engineers Australia
Medical Technology Association of Australia
Ministry of Economic Development, New Zealand
New Zealand Institute of Healthcare Engineering
Queensland Health
Testing Interests, Australia
The Royal Australian and New Zealand College of Radiologists
Therapeutic Goods Administration
Wairarapa District Health Board

Keeping Standards up-to-date

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about joint Australian/New Zealand Standards can be found by visiting the Standards Web Shop at www.saiglobal.com or Standards New Zealand web site at www.standards.govt.nz and looking up the relevant Standard in the on-line catalogue.

For more frequent listings or notification of revisions, amendments and withdrawals, Standards Australia and Standards New Zealand offer a number of update options. For information about these services, users should contact their respective national Standards organization.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Please address your comments to the Chief Executive of Standards Australia or the New Zealand Standards Executive at the address shown on the back cover.

This Standard was issued in draft form for comment as DR 10023.

Australian/New Zealand Standard™

Management programs for medical equipment

Originated in Australia as AS 3551—1988.
Jointly revised and designated as AS/NZS 3551:1998.
Previous edition 2004.
Fourth edition 2012.
Reissued incorporating Amendment No. 1 (October 2013).
Reissued incorporating Amendment No. 2 (September 2016).

COPYRIGHT

© Standards Australia Limited/Standards New Zealand

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher, unless otherwise permitted under the Copyright Act 1968 (Australia) or the Copyright Act 1994 (New Zealand).

Jointly published by SAI Global Limited under licence from Standards Australia Limited, GPO Box 476, Sydney, NSW 2001 and by Standards New Zealand, PO Box 1473, Wellington 6011.

PREFACE

This Standard was prepared by the Joint Standards Australia/Standards New Zealand Committee HE-003, Medical Electrical Equipment, to supersede AS/NZS 3551:2004.

This Standard incorporates Amendment No. 1 (October 2013) and Amendment No. 2 (September 2016). The changes required by the Amendment are indicated in the text by a marginal bar and amendment number against the clause, note, table, figure or part thereof affected.

The principal differences between this edition and the previous edition are as follows:

- (a) Principles from quality management and risk management standards have been incorporated throughout the document.
- (b) Introduction of the requirements for hire and loan equipment.
- (c) Introduction of the term ‘performance verification’ which encompasses physical, functional, and electrical testing.
- (d) Change in earth leakage current tests to align with IEC 60601-1, Ed. 3.0.
- (e) Addition of touch current testing to align with IEC 60601-1, Ed. 3.0.
- (f) Addition of the medical electrical systems section (Section 7) to align with IEC 60601-1, Ed. 3.0.
- (g) Management of medical equipment configurations during medical equipment maintenance.
- (h) Introduction of test requirements for non-medical equipment used in a patient environment.
- (i) Introduction of terms ‘service entity’ and ‘responsible organization’.
- (j) Introduction of requirements for management of medical equipment used in the home.

The Standard was prepared for users of medical equipment who need to—

- (i) ensure that medical equipment used in the responsible organization perform as intended by the manufacturer and is safe for clinical use;
- (ii) ensure that the medical equipment complies with the essential principles of safety and performance;
- (iii) inspect and test new medical equipment, before commissioning, in order to ensure against manufacturing defects; and
- (iv) perform maintenance, routine inspections or tests on medical equipment during its service life, in order to ensure its continued safety and performance.

This Standard is intended to provide a practical approach and application of the testing philosophies of AS/NZS 3200.1.0 and associated Part 2 Standards, along with other essential performance verification principles, to ensure that medical equipment, in clinical use in the responsible organization, performs as intended by the manufacturer and does not compromise the safety of the patient, the operator or the environment.

Recommendations linking the type of medical procedures with the appropriate type and class of medical electrical equipment and the type of protective facilities in the reticulated mains wiring are not covered in this Standard but are published as AS/NZS 2500, *Guide to safe use of electricity in patient care*.

Design requirements for type examination and approval of medical electrical equipment are not considered in this Standard but are set out in AS/NZS 3200.1.0 and Collateral Standards, and the relevant Part 2 Standards in the AS/NZS 3200 series. In this regard, it is important to note the following:

- (A) The test requirements of the AS/NZS 3200 Part 2 Standards can be applied to a sample of the medical equipment to verify that the basic design and manufacture complies with these Standards before the medical equipment is introduced to the market.
- (B) Many of the tests in the AS/NZS 3200 Part 2 Standards are too extensive to be applied on a routine basis to each item of equipment delivered to a hospital or healthcare facility. Some inspection and testing requirements of the AS/NZS 3200 Part 2 Standards are potentially destructive of the medical equipment or its components. Some of those type tests, if applied to each item of equipment purchased, may render some or all of them inoperative or unsafe for use.
- (C) Few organizations and healthcare facilities have the resources to carry out type testing to the full requirements of the AS/NZS 3200 Part 2 Standards.
- (D) For these reasons, it is not appropriate, nor would it be cost effective, to carry out all the inspection and testing required in the AS/NZS 3200 Part 2 Standards on every item of equipment to be commissioned.
- (E) The AS/NZS 3200 Part 2 Standards modify the Parent Standard and are totally dependent on AS/NZS 3200.1.0 for the bulk of requirements.

The terms 'normative' and 'informative' have been used in this Standard to define the application of the appendix to which they apply. A 'normative' appendix is an integral part of a Standard, whereas an 'informative' appendix is only for information and guidance.

Statements expressed in mandatory terms in notes to tables and figures are deemed to be requirements of this Standard.

CONTENTS

	<i>Page</i>
FOREWORD.....	6
SECTION 1 SCOPE AND GENERAL	
1.1 SCOPE.....	7
1.2 APPLICATION	8
1.3 REFERENCED DOCUMENTS.....	8
1.4 DEFINITIONS.....	9
SECTION 2 MEDICAL EQUIPMENT MANAGEMENT PROGRAM	
2.1 GENERAL.....	17
2.2 PROGRAM SUPPORT FUNCTIONS.....	17
2.3 TEST EQUIPMENT	19
2.4 DOCUMENTATION REQUIREMENTS	20
2.5 EQUIPMENT DATABASE.....	20
2.6 MAJOR PROGRAM FUNCTIONS.....	22
SECTION 3 REGULATORY COMPLIANCE	
3.1 GENERAL EVIDENCE OF REGULATORY COMPLIANCE	24
3.2 MARKETING AUTHORIZATION.....	24
3.3 ELECTROMAGNETIC COMPATIBILITY.....	24
3.4 MEDICAL EQUIPMENT FOR CLINICAL TRIAL.....	25
SECTION 4 PROCUREMENT	
4.1 SCOPE.....	26
4.2 PROCUREMENT PROCEDURES.....	26
SECTION 5 ACCEPTANCE	
5.1 SCOPE	29
5.2 GENERAL.....	29
5.3 ACCEPTANCE CHECK AND INSPECTION	29
5.4 TEST SELECTION AND ACCEPTANCE TESTING	31
5.5 DOCUMENTATION OF RESULTS	32
5.6 FAILED MEDICAL EQUIPMENT	33
SECTION 6 PERFORMANCE VERIFICATION AND MAINTENANCE	
6.1 GENERAL.....	34
6.2 DOCUMENTING OF PERFORMANCE VERIFICATION	34
6.3 PERFORMANCE VERIFICATION.....	34
6.4 PREVENTIVE MAINTENANCE	39
6.5 MANAGEMENT OF MEDICAL EQUIPMENT FAILURE OR FAULTS	40
6.6 CONFIGURABLE MEDICAL EQUIPMENT.....	40
6.7 REVIEW AND RECORDING OF RESULTS	41
6.8 DISPOSAL	43
6.9 SAFETY-RELATED NOTIFICATIONS, ALERTS AND RECALLS	43
6.10 MEDICAL EQUIPMENT ADVERSE EVENT OR FAILURE	44

SECTION 7	MEDICAL ELECTRICAL SYSTEMS	
7.1	GENERAL.....	45
7.2	TYPES OF MEDICAL ELECTRICAL SYSTEM	45
7.3	TESTING CONFIGURATION.....	47
SECTION 8	MEDICAL DEVICES ON LOAN, LEASE OR HIRE	
8.1	HIRE OR LOAN MEDICAL EQUIPMENT	48
8.2	HIRE OR LOAN TO THE RESPONSIBLE ORGANIZATION.....	48
8.3	HIRE OR LOAN FROM THE RESPONSIBLE ORGANIZATION	48
8.4	MEDICAL EQUIPMENT ON TRIAL IN THE ORGANIZATION.....	49
SECTION 9	MODIFICATION OF MEDICAL EQUIPMENT	50
SECTION 10	ASSESSMENT INTERVALS.....	51
APPENDICES		
A	APPLICATION OF RISK MANAGEMENT PRINCIPLES IN ESTABLISHING A MEDICAL EQUIPMENT MANAGEMENT PROGRAM.....	52
B	ELECTRICAL SAFETY TESTING	58
C	MEASURING DEVICE	72
D	EXAMPLE—TYPICAL FAULT REPORT	73
E	ESSENTIAL SAFETY AND PERFORMANCE PARAMETERS.....	74
F	EXAMPLE—TYPICAL PURCHASE/ACCEPTANCE RECORD.....	79
G	EXAMPLE—TYPICAL PERFORMANCE VERIFICATION REPORT— INFUSION PUMP	81
H	EXAMPLE—TYPICAL PERFORMANCE VERIFICATION REPORT— SYRINGE PUMP	82
I	EXAMPLE—TYPICAL PERFORMANCE VERIFICATION REPORT— DEFIBRILLATOR	83
J	EXAMPLE—TYPICAL PERFORMANCE VERIFICATION REPORT— BABY INCUBATORS	84
K	EXAMPLE—TYPICAL ELECTRICAL TEST REPORT	85
L	APPLIED PARTS OF MEDICAL ELECTRICAL EQUIPMENT	87
M	CONTACT INFORMATION	89

FOREWORD

The safe application of medical equipment depends on a variety of factors. The minimum requirements are as follows:

- (a) Medical electrical equipment is used only in a patient area provided with appropriate protection and the reticulated mains wiring in accordance with AS/NZS 3003, *Electrical installations—Patient areas*.
- (b) Appropriate equipment is used for each particular application and in accordance with an appropriate set of rules linking the type of procedure with the class of medical equipment and the electrical safety facilities provided in the patient area.
- (c) Each new item of equipment is—
 - (i) acceptance tested prior to clinical use;
 - (ii) subjected to routine performance verification during its useable life to detect damage, wear, component failure or changed component value which might render it unsafe; and
 - (iii) maintained with reference to the manufacturer's instructions using formal risk analysis within a risk management framework.
- (d) The users of the medical equipment must know, not only the medical procedure, but also the safety characteristics and operational details of the medical equipment. This can be achieved by learning and training under the supervision of the manufacturer, the local representative, experienced users, or staff from biomedical engineering departments or the service entity.
- (e) The responsible organization needs to ensure that safety and performance of the medical equipment is maintained by an effective maintenance scheme with regular assessment, testing and maintenance in accordance with this Standard.

STANDARDS AUSTRALIA/STANDARDS NEW ZEALAND

Australian/New Zealand Standard
Management programs for medical equipment

SECTION 1 SCOPE AND GENERAL

1.1 SCOPE

This Standard outlines procedures required to develop management programs for medical equipment. These include protocols and procedures for procurement, acceptance, maintenance activities throughout the service life of the medical equipment, and finally disposal of medical equipment.

This Standard applies to non-electrical medical equipment, such as ventilators, which may be solely pneumatic or fluidic in operation, as well as medical electrical equipment. Also included in the scope are products that are not generally considered to be medical equipment, but are used in the patient environment, such as computers, imaging cameras and recorders.

This Standard applies whether the medical equipment is owned by the responsible organization, privately owned, on loan, on hire, on trial or donated, and includes medical equipment provided by a responsible organization as part of a pool of medical equipment for hire. It does not apply to the electrical commissioning requirements of permanently installed medical electrical devices as specified in AS/NZS 3003.

Although this Standard is focused on the management of medical equipment within an organization, it is acknowledged that some responsible organizations may choose to use many of the protocols and procedures of this document to manage medical equipment in other areas, for example laboratories. In such circumstances, the organization, or its service entity, will have to consider the relevance of some of the requirements outlined in this document.

In developing this Standard, reference has been made to manufacturer's specifications for a wide range of commercially available medical equipment. It is not possible in this Standard to address every type of medical equipment, particularly since new treatment modalities and medical equipment will be marketed during the life of the Standard. It is therefore important to make reference to manufacturer's specifications for the operation, performance testing and calibration of any new medical equipment, and to balance this with a knowledge of professional biomedical engineering practice for similar equipment.

Where the acceptance values of an AS/NZS 3200 Part 2 Standard for a particular type of medical electrical equipment are not in agreement with the requirements of AS/NZS 3200.1.0:1998 or IEC 60601-1, Ed. 3.0 (2005), the requirements of the Part 2 Standard override these documents. For example, the normal limit (specified in AS/NZS 3200.1.0) for mains contact current is 50 μA per Applied Part, while AS/NZS 3200.2.4 specifies an allowable current of 100 μA for Type CF defibrillator Applied Parts. It should also be noted that, in some instances, allowable limits applicable for acceptance testing may be more stringent than for ongoing performance verification.