

Australian Standard<sup>®</sup>

---

**Electrical equipment for coal  
mines—Maintenance and overhaul**

**Part 3: Maintenance of gas  
detecting and monitoring  
equipment**

---

This Australian Standard was prepared by Committee EL/23, Electrical Equipment in Coal Mines. It was approved on behalf of the Council of Standards Australia on 13 November 1989 and published on 2 April 1990.

---

The following interests are represented on Committee EL/23:

Australian Chamber of Commerce  
Australian Coal Association  
Australian Electrical and Electronic Manufacturers Association  
Confederation of Australian Industry  
Department of Industrial Relations and Employment, N.S.W.  
Department of Minerals and Energy, N.S.W.  
Department of Mines, Qld  
Institution of Mining Electrical and Mining Mechanical Engineers  
Joint Coal Board  
Regulatory authorities (electrical)

Additional interests participating in preparation of Standard:

Australian Coal Industrial Research Laboratories  
Southern Mines Rescue Station

---

**Review of Australian Standards.** To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications; this information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

---

*This Standard was issued in draft form for comment as DR 89064.*

Australian Standard<sup>®</sup>

---

**Electrical equipment for coal  
mines—Maintenance and overhaul**

**Part 3: Maintenance of gas  
detecting and monitoring  
equipment**

---

First published as AS 2290.3—1990.

## PREFACE

This Standard was prepared by the Standards Australia Committee on Electrical Equipment in Coal Mines. It is intended for the guidance of users and relevant regulatory authorities concerned with the maintenance of electrical equipment in coal mines and is part of a series of Standards on the maintenance and overhaul of electrical equipment used in association with underground mining machines.

## CONTENTS

	<i>Page</i>
FOREWORD .....	3
SECTION 1. SCOPE AND GENERAL	
1.1 SCOPE .....	4
1.2 REFERENCED DOCUMENTS .....	4
1.3 DEFINITIONS .....	4
1.4 EQUIPMENT CLASSIFICATION .....	5
1.5 EXAMINATION FREQUENCY .....	5
1.6 EXAMINATION RESPONSIBILITY .....	5
1.7 EXAMINATION LOCATIONS .....	5
1.8 CERTIFICATE OF COMPLIANCE .....	5
1.9 RECORDS .....	6
1.10 ACCURACY REQUIREMENTS .....	6
1.11 OVERHAUL .....	6
1.12 USER/FIELD SERVICEABLE ITEMS .....	7
SECTION 2. EXAMINATION SCHEDULES	
2.1 GENERAL .....	8
2.2 CLASS I EQUIPMENT .....	8
2.3 CLASS II EQUIPMENT .....	9
2.4 CLASS III EQUIPMENT .....	10
2.5 CLASS IV EQUIPMENT .....	11
APPENDICES	
A TYPICAL CERTIFICATE OF COMPLIANCE .....	13
B SAMPLE LAYOUTS FOR RECORDING EQUIPMENT DEFECTS ..	14
C ACCURACY REQUIREMENTS .....	17
D MINIMUM REQUIREMENTS FOR AUTHORIZED PERSONS AND ACCREDITED TEST AUTHORITIES .....	18
E TECHNIQUE AND EQUIPMENT GUIDELINES FOR VERIFICA- TION CHECKS .....	19

## © Copyright — STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

## FOREWORD

Various Australian Standards and statutory requirements lay down criteria for the manufacture and regulations for the installation and use of electrical equipment.

This Standard recognizes the need to maintain equipment properly while it is in service, and lays down guidelines in the form of checklists and examination procedures.

Where maintenance and testing of gas detecting and monitoring equipment is required, the Standard specifies that the work be carried out by organizations approved for the purpose. Only persons authorized by the relevant Regulatory Authority are to supervise or carry out such work.

Guidelines setting out the manner in which the work is to be performed and recorded are included.

## STANDARDS AUSTRALIA

## Australian Standard

## Electrical equipment for coal mines — Maintenance and overhaul

## Part 3: Maintenance of gas detecting and monitoring equipment

## SECTION 1. SCOPE AND GENERAL

**1.1 SCOPE.** This Standard sets out requirements for the examination and calibration of gas detecting and monitoring equipment for use in and around underground coal mines. It details methods of examination, maintenance and testing required to ensure safety, compliance with relevant Standards and trouble-free operation while the equipment is in service.

Organizations and personnel engaged in the examination, repair and testing of gas detecting equipment may be required to be approved for the purpose by the relevant Regulatory Authority. Minimum requirements for authorized persons and accredited test authorities are given in Appendix D.

## NOTES:

1. Gas detector tubes, of either sample draw or diffusion types, which rely on a visually detectable chemical change in the presence of a gas, are not subject to the requirements of this Standard.
2. Attention is drawn to the various Mines Acts, the Regulations under these Acts, and also any other appropriate statutory requirements or by-laws. These place the prime responsibility for complying with specific safety requirements on the manufacturer or user of the equipment.

**1.2 REFERENCED DOCUMENTS.** The following documents are referred to in this Standard:

AS 2706	Numerical values — Rounding and interpretation of limiting values
3573	Electrical equipment for coal mines — Glossary of terms

**1.3 DEFINITIONS.** For the purpose of this Standard, the definitions given in AS 3573 and those below apply.

**1.3.1 Accredited test authority**—authority approved, or otherwise recognized, by the relevant Regulatory Authority, for the purpose of examination and calibration of gas detecting equipment as required by this Standard, and approved to issue a Certificate of Compliance.

**1.3.2 Accredited workshop**—workshop approved, or otherwise recognized, by the relevant Regulatory Authority, for the purpose of repair of gas detecting equipment as required by this Standard.

**1.3.3 Authorized person**—person authorized by the relevant Regulatory Authority as being competent to carry out or supervise the examinations as required by this Standard.

**1.3.4 Certified test gas**—mixture of a test gas and diluent in which the true content of the test gas, with a confidence level, is accepted by the relevant Regulatory Authority.

**1.3.5 Certificate of Compliance**—certificate issued by an accredited test authority verifying that the equipment complies with the relevant requirements of this Standard.

**1.3.6 Verification check**—check on the response of gas detecting equipment when exposed to zero gas and certified test gas(es) or, where alarm levels are the subject of the test, the accuracy of alarm settings.

NOTE: Verification checks should not be confused with full calibration of equipment which is, in general, more rigorous.

**1.3.7 Full scale concentration**—that gas concentration equivalent to the maximum variation in gas concentration which may be read from the equipment scale and range being used.

**1.3.8 Gas sensing element**—that part of gas detecting equipment which is in direct contact with sampled gas and at which a physical or chemical change, dependent on gas concentration, occurs.

**1.3.9 Gas detector head**—discrete assembly within gas detecting or monitoring equipment which contains the gas sensing element.

**1.3.10 Examination**—inspection to determine the condition of gas detecting equipment. This may involve visual inspection, the application of zero or span tests or the verification of equipment functions.

**1.3.11 Electrical zero**—the reading produced by gas detecting equipment under zero gas conditions, with the equipment electrically activated.

**1.3.12 Mechanical zero**—the reading produced by gas detecting equipment in the absence of electrical activation of the equipment. This reading corresponds to the reading produced by an error in the mechanical adjustment of moving coil meters.

**1.3.13 Zero gas**—for equipment other than oxygen sensing, a gas mixture free from the component which the item of gas detecting equipment under test is designed to detect. For oxygen sensing equipment, clean air is used as a zero gas.

NOTE: The convention has been adopted that the normal null point for oxygen detecting equipment is that corresponding to fresh air. Normal instrument response from this point is in a negative direction with decreasing oxygen concentration.

**1.3.14 Zero gas conditions**—application of zero gas to the gas sensing element of equipment under test.

**1.3.15 Zero test**—test of response to zero gas conditions. The 'zero' reading for oxygen detecting equipment is that corresponding to clean air.

**1.3.16 Span test**—test of response to certified test gas(es). The 'span' reading for oxygen detecting equipment is that corresponding to certified test gas(es) containing less oxygen than clean air.