

Australian Standard[®]

**Measurement of water flow in open
channels**

**Part 5.3: Liquid flow in open channels
and partly filled pipes—Guidelines for
the application of Doppler-based flow
measurements**



This Australian Standard® was prepared by Committee CE-024, Measurement of Water Flow in Open Channels and Closed Conduits. It was approved on behalf of the Council of Standards Australia on 5 April 2007.
This Standard was published on 6 August 2007.

The following are represented on Committee CE-024:

- Australian Industry Group
 - Australian National Committee on Irrigation and Drainage
 - Department of Environment and Water Resources
 - Institute of Instrumentation, Control and Automation Australia
 - Irrigation Association of Australia
 - National Measurement Institute
 - Plumbing Products Industry Group
 - University of New South Wales
 - University of South Australia
 - Water Services Association of Australia
-

This Standard was issued in draft form for comment as DR 07113.

Standards Australia wishes to acknowledge the participation of the expert individuals that contributed to the development of this Standard through their representation on the Committee and through the public comment period.

Keeping Standards up-to-date

Australian Standards® are living documents that reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued.

Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments that may have been published since the Standard was published.

Detailed information about Australian Standards, drafts, amendments and new projects can be found by visiting **www.standards.org.au**

Standards Australia welcomes suggestions for improvements, and encourages readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at **mail@standards.org.au**, or write to Standards Australia, GPO Box 476, Sydney, NSW 2001.

Australian Standard[®]

Measurement of water flow in open channels

Part 5.3: Liquid flow in open channels and partly filled pipes—Guidelines for the application of Doppler-based flow measurements

First published as AS 3778.5.3—2007

COPYRIGHT

© Standards Australia

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia GPO Box 476, Sydney, NSW 2001, Australia

ISBN 0 7337 8309 0

PREFACE

This Standard was prepared by the Standards Australia Committee CE-024, Measurement of Water Flow in Open Channels and Closed Conduits.

The objective of this Standard is to specify methods for measuring discharge in large rivers and estuaries by the moving-boat technique.

This Standard is identical to and reproduced from ISO/TS 15769:2000, *Hydrometric determinations—Liquid flow in open channels and partly filled pipes—Guidelines for the application of Doppler-based flow measurements*.

As this Standard is reproduced from an international standard, the following applies:

- (a) its number appears on the cover and title page while the international standard number appears only on the cover
- (b) In the source text 'ISO/TS 15769' should read 'AS 3776.5.3'
- (c) A full point substitutes for a comma when referring to a decimal marker.

The term 'informative' has been used in this Standard to define the application of the annex to which it applies. An 'informative' annex is only for information and guidance.

CONTENTS

	<i>Page</i>
1	Scope 1
2	Normative references 1
3	Terms and definitions 1
4	Principle 1
5	Factors affecting operation and accuracy 3
5.1	General 3
5.2	Characteristics of the instrument 3
5.3	Channel and water characteristics 6
6	General expectations of performance 10
7	Considerations when selecting equipment 10
8	Evaluation and verification 10
Annex A (informative)	Other practical considerations 11
Annex B (informative)	Sample questionnaire — Doppler-based flowmeter 16

INTRODUCTION

Flow measurement using Doppler-based flowmeters fall into the category of velocity-area methods and, as with all variants of this approach, flow estimation is a two-stage process. Measurements are made to derive estimates of:

- a) mean channel velocity, using the principle of Doppler shift,
- b) depth, from which cross-sectional area is computed with a knowledge of the relationship between depth and area (i.e. the profile of the cross-section).

Most difficulties governing this method relate to the estimation of mean channel velocity and the degree to which computed velocities can be assumed to be representative of the true mean velocity through the measurement section. The accuracy with which the calculated velocities represent the mean velocity is influenced by various factors which are considered in clause 4. This guide focuses on the process of velocity estimation and the conditions and practices which may help to deliver optimum results. However, it should be recognized that the accuracy of overall flow determination also depends on the accuracy of depth measurement. The influence of meter location and sensitivity of cross-sectional area to depth variation will have a bearing on performance.

STANDARDS AUSTRALIA

Australian Standard

**Measurement of water flow in open channels
Part 5.3: Liquid flow in open channels and partly filled pipes—Guidelines
for the application of Doppler-based flow measurements****1 Scope**

This Technical Specification gives guidelines for the selection and use of Doppler-based flowmeters for the measurement of liquid flow in small open channels and partly filled pipes. It is applicable to whole-channel flow measurements using a single meter at a fixed point in the cross-section.

It is not applicable to flow measurements made using Doppler-based current meters for point velocity measurement nor using mobile profiling systems.

A limitation of this technique is that measurement is made of the velocity of particles or other reflectors being transported by the liquid rather than that of the liquid itself.

NOTE Though the title refers to liquid flow in general, the main use of flowmeters which use the Doppler principle is for water or water-based liquid and particular reference is often made in the text to water.

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this Technical Specification. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this Technical Specification are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 772, *Hydrometric determinations — Vocabulary and symbols*.

ISO/TR 8363, *Measurement of liquid flow in open channels — General guidelines for selection of method*.

3 Terms and definitions

For the purposes of this Technical Specification, the terms and definitions given in ISO 772 apply.

4 Principle

This method of velocity measurement is based on a phenomenon first identified by Christian Doppler in 1843. The principle of “Doppler shift” describes the difference or shift which occurs in the frequency of emitted sound waves as they are reflected back from a moving body.

The sensors of Doppler systems normally contain a transmitting and a receiving device (see Figure 1). A sound wave of high frequency (f_s) is transmitted into the flow of water and intercepted and reflected back at a different frequency by tiny particles or air bubbles (reflectors). A typical reflector “ n ” produces a frequency shift $f_{D,n}$. The “shift” between transmitted and reflected frequencies is proportional to the movement of particles relative to the position of the sound source (i.e. the sensor).