

Superseded by AS 2243.7-1991

AS 2243, Part 7—1980
UDC 614.8.006.25

Under revision see PR 90046

Australian Standard 2243, Part 7—1980

SAFETY IN LABORATORIES Part 7—ELECTRICAL ASPECTS



STANDARDS ASSOCIATION OF AUSTRALIA
Incorporated by Royal Charter



THE FOLLOWING SCIENTIFIC, INDUSTRIAL AND GOVERNMENTAL ORGANIZATIONS and departments were officially represented on the committee entrusted with the preparation of this standard:

Australian Atomic Energy Commission
Australian Institute of Petroleum Ltd
The Broken Hill Pty Co Ltd
Commonwealth Scientific and Industrial Research Organization
Department of Productivity
Department of Science and Technology
State Laboratories, Victoria

This standard, prepared by Committee CH/26, Safety in Laboratories, was approved by the Chemical Standards Board on behalf of the Council of the Standards Association of Australia on 10 October 1980, and was published on 31 December 1980.

To keep abreast of progress in industry, Australian standards are subject to continuous review and are kept up-to-date by the issue of amendments or new editions as necessary. It is important therefore that standards users ensure that their standards are up-to-date. Full details of all SAA publications will be found in the Annual List of Australian Standards; these details are supplemented by listings in the SAA monthly journal 'The Australian Standard'. Information on the Annual List and 'The Australian Standard' may be obtained from any sales office of the Association, where details are also available of the current status of individual standards. Suggestions for improvements to published standards, addressed to the head office of the Association, are welcomed.

This standard was issued in draft form for public review as DR 79141.

AUSTRALIAN STANDARD

**CODE OF PRACTICE FOR
SAFETY IN LABORATORIES
Part 7
ELECTRICAL ASPECTS**

AS 2243, Part 7—1980

First published	1980
Reprinted (incorporating Corrig.)	1982

**PUBLISHED BY THE STANDARDS ASSOCIATION OF AUSTRALIA
STANDARDS HOUSE, 80 ARTHUR ST, NORTH SYDNEY, N.S.W.**

ISBN 0 7262 2127 9



16 FEB 1982

PREFACE

This standard was prepared by the Association's Committee on Safety in Laboratories under the direction of the Chemical Standards Board. It represents the seventh part of the series on the subject of safety in laboratories, and it supplements the information given in the other parts. This standard is aimed at the provision of a sound, basic working knowledge of electrical safety in the laboratory situation and the encouragement of systematic procedures for preventing electrical accidents.

Other parts are as follows:

- Part 1—General
- Part 2—Chemical
- Part 3—Microbiology
- Part 4—Ionizing Radiations
- Part 5—Non-ionizing Radiations
- Part 6—Mechanical Aspects

In addition to the reference standards and publications listed in Appendix B, this standard makes reference to the following Australian standards:

- | | |
|----------|---|
| AS 1076 | Code of Practice for Selection, Installation and Maintenance of Electrical Apparatus and Associated Equipment for Use in Explosive Atmospheres (Other than Mining Applications) |
| AS 1188 | Code for Safety of Electronic Equipment |
| AS 1848 | Halogenated Hydrocarbon Type Portable Fire Extinguishers |
| AS 1895 | Code of Practice for Guarding and Safe Use of Portable Electric Tools for Domestic Use |
| AS 2243 | Code of Practice for Safety in Laboratories
Part 4—Ionizing Radiations |
| AS 3000 | SAA Wiring Rules |
| AS 3160 | Approval and Test Specification for Hand-held Portable Electric Tools |
| AS 3190 | Approval and Test Specification for Current-operated (Core Balance) Earth-leakage Devices |
| AS 3191 | Approval and Test Specification for Electric Flexible Cords |
| AS C100 | Approval and Test Specification for Definitions and General Requirements for Electrical Materials and Equipment |
| AS C109 | Approval and Test Specification for Appliance Plugs and Appliance Inlet-sockets |
| AS C112 | Approval and Test Specification for Plugs and Plug Sockets |
| AS C126 | Approval and Test Specification for Extra-low Voltage Transformers |
| AS C167 | Approval and Test Specification for Protective Isolating Transformers |
| AS | Laboratory Design and Construction for Safe Working Conditions* |
| SAA MP30 | Report on Effects of Current Passing Through the Human Body |

*In course of preparation.

©Copyright — STANDARDS ASSOCIATION OF AUSTRALIA 1980

Users of standards are reminded that copyright subsists in all SAA publications. No part of this publication may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing of the Standards Association of Australia.

CONTENTS

	<i>Page</i>		<i>Page</i>
FOREWORD	4	5.2 Officer-in-charge	11
SECTION 1. SCOPE AND DEFINITIONS		5.3 Electrical Appliance Register	11
1.1 Scope	5	5.4 Inspection	11
1.2 Definitions	5	5.5 Safety Tests	11
SECTION 2. GENERAL SAFETY PROCEDURES		5.6 General Purpose Outlets	11
2.1 Scope of Section	6	SECTION 6. EQUIPMENT	
2.2 Checking of Equipment Before Use	6	6.1 Autotransformers	12
2.3 Conducting Materials and Apparel	6	6.2 High Grade Capacitors	12
2.4 Safety Observer	6	6.3 Experimental Equipment	12
2.5 Electrical Installations	6	6.4 Working on Live Equipment	12
2.6 Emergency Isolating Switch	6	6.5 Unattended Equipment	12
2.7 Installations in Hazardous Locations.....	6	6.6 Batteries	13
2.8 Prevention of Electrical Ignition	6	6.7 Inspection and Testing of Permanently Wired Equipment	13
2.9 Colour Coding of Conductors	7	SECTION 7. ELECTROMAGNETIC RADIATION	13
2.10 Imported Equipment	7	SECTION 8. ELECTRICAL SAFETY SYSTEMS	
2.11 Plug Socket Polarisation	7	8.1 General	14
SECTION 3. SAFE PROCEDURES FOR OFFICE MACHINES		8.2 Tag System	14
3.1 General	8	8.3 Lock-out System	14
3.2 Servicing	8	8.4 Permit System	14
3.3 Battery Recharging	8	SECTION 9. ACTION IN CASE OF FIRE	
3.4 Periodic Inspection	8	9.1 General	16
SECTION 4. SAFE PROCEDURES FOR ELECTRIC POWERED TOOLS OR PORTABLE APPLIANCES		9.2 Fire Extinguishers	16
4.1 General	9	9.3 Action if a Fire Starts	16
4.2 Double Insulated Appliances	9	SECTION 10. ELECTRIC SHOCK	
4.3 Supplementary Protection	9	10.1 General	17
4.4 Flexible Cord	9	10.2 Electrical Characteristics of the Human Body	17
4.5 Extension Cords	9	10.3 Severity of Electric Shock	17
4.6 Plugs and Sockets	10	10.4 Treatment of Electric Shock	17
4.7 Operation of Portable Electric Tools	10	APPENDICES	
SECTION 5. ISSUE, INSPECTION AND MAINTENANCE OF PORTABLE APPLIANCES		A Treatment of Victims of Electric Shock	18
5.1 General	11	B Bibliography	24

STANDARDS ASSOCIATION OF AUSTRALIA

Australian Standard

CODE OF PRACTICE FOR SAFETY IN LABORATORIES

PART 7—ELECTRICAL ASPECTS

FOREWORD

Electrical safety entails safe practices, safe equipment and common sense to effectively contain occupational hazards due to electricity. Electrical accidents in the laboratory account for only a small percentage of all accidents, but they may involve serious injury.

In the event of an electric shock, it is of the utmost importance, if the victim fails to breathe and/or has lost pulse, that resuscitation be commenced immediately after rescue. Any delay may seriously reduce the chance of recovery.

To ensure that rescue and resuscitation will be quickly and competently applied, all staff who are required to work or assist on any work in the vicinity of electrical apparatus should be trained in the methods of rescue and resuscitation, and be given refresher courses at least annually.

Electrical hazards can also cause burns to the body and can start fires and explosions. These hazards can be minimized if all equipment and installations are inspected regularly by competent staff.

SECTION 1. SCOPE AND DEFINITIONS

1.1 SCOPE. This Part of the code sets out the general principles for the use of electrical equipment in general purpose laboratories to ensure the electrical safety of personnel.

The code is not intended for use in laboratories where high electrical energies are generated or used, nor is it intended to replace any legislation, statutory requirements or electricity supply authority regulations.

1.2 DEFINITIONS. For the purpose of this Part of the code, the following definitions apply:

1.2.1 Head of Laboratory—the person on whom ultimate responsibility devolves for the conduct of a laboratory (see Note.)

NOTE: Where the term 'Head of Laboratory' is used in the text it is intended to mean that the ultimate responsibility rests with the Head of Laboratory. However, in the delegation of any normal managerial processes, the person physically concerned may be a delegate of appropriate level, e.g. section head of appropriate technical expertise, or professional engineer, scientist or safety officer. As administrative appointments and terms vary from establishment to establishment the title of the actual person delegated will vary from place to place.

1.2.2 Active—any terminal or conductor of a supply system which is maintained at a difference of potential from the neutral or earthed conductor. In a system which does not include a neutral or earthed conductor, all conductors shall be considered to be active.

Alive—see live.

1.2.3 Appliance—a consuming device, other than a lamp, in which electricity is converted into heat, motion or any other form of energy, or is substantially changed in its electrical character.

1.2.4 Appliance, portable—either an appliance which is moved in operation or an appliance which can easily be moved from one place to another whilst connected to supply.

1.2.5 Approved—complying with the appropriate standard, or having the approval or acceptance of the Supply Authority, or where existing standards or Authority regulations do not apply, authorized by the Head of Laboratory.

1.2.6 Cable, flexible—a cable, the conductors, insulation and covering of which are such as to afford flexibility.

1.2.7 Conductor—a wire or other form of conducting material suitable for carrying current, but not including wire or other metallic parts directly employed in converting electrical energy into another form.

1.2.8 Cord, flexible—a flexible cable, no wire of which exceeds 0.30 mm diameter and no conductor of which exceeds 4 mm² cross-sectional area, and having not more than five cores.

1.2.9 Equipment—any component part of an electrical installation.

1.2.10 Flexible extension cord—a length of flexible cord one end of which is terminated in a plug, and the other end in a cord-extension socket.

1.2.11 Dead—an object at or about earth potential and disconnected from any live system.

1.2.12 Earthed—connected to the general mass of earth in accordance with the appropriate requirements of AS 3000, Part 1, in such a manner as will ensure, through the operation of protective apparatus, the electrical isolation of any defective equipment.

1.2.13 Hazardous location—an area in which explosive gas/air mixtures, or dusts, or fibres or flyings are present, or may be expected to be present in quantities such as to require special precautions for the installation and use of electrical equipment.

1.2.14 Live (alive)—an object is live (alive) when a difference of potential exists, or would exist, between it and earth under normal conditions of operation. With the exception of earthing conductors and neutral busbars or links in installations where the multiple-earthed neutral system is employed, all metal connected to the neutral conductor of the supply system, even if such neutral be earthed at the source of supply, is deemed to be live in accordance with AS 3000, Part 1.

1.2.15 Outlet, general purpose (GPO)—a plug socket having a rating of 10 A, and intended for the connection of low voltage appliances or luminaires through flexible cord or cable and a plug, but not including a plug socket or cord extension socket intended solely for the connection of one particular fixed or stationary appliance or luminaires or forming part of a special purpose final subcircuit in accordance with Rule 2.11 of AS 3000, Part 1.

1.2.16 Plug—a device intended for insertion in a plug socket, cord-extension socket or plug-socket adaptor to make a detachable connection between the contacts of any such accessory and the conductors of a flexible core or flexible cable.

NOTE: The commercial terms 'plug top' and 'appliance plug' are covered by this definition.

1.2.17 Cord extension socket—a device arranged for attachment to a flexible cord and having contacts whereby a detachable connection may be made with the contacts of a plug.

1.2.18 Plug socket—a device for fixing or suspension at a point and having contacts intended for making a detachable connection with the contacts of a plug. The term 'plug socket' is deemed to include a cord extension socket attached to a flexible cord which is permanently connected to fixed wiring. (See also outlet, general purpose).

1.2.19 Supply Authority—any person or organization engaged in the supply of electricity to the public or in the generation of electricity for supply directly or indirectly to the public, whether by virtue of any statute or any franchise agreement under any Act or otherwise.

1.2.20 Voltage—differences of potential normally existing between conductors, and between conductors and earth, as follows:

- (a) *Extra-low voltage*—not exceeding 32 V alternating current or 115 V direct current.
- (b) *Low voltage*—exceeding extra-low voltage, but not exceeding 250 V.
- (c) *Medium voltage*—exceeding low voltage but not exceeding 650 V.
- (d) *High voltage*—exceeding 650 V.