

Australian/New Zealand Standard™

Safety in laboratories

Part 4: Ionizing radiations



AS/NZS 2243.4:2018

This Joint Australian/New Zealand Standard was prepared by Joint Technical Committee CH-026, Safety in Laboratories. It was approved on behalf of the Council of Standards Australia on 29 March 2018 and by the New Zealand Standards Approval Board on 6 June 2018.

This Standard was published on 29 June 2018.

The following are represented on Committee CH-026:

Association of Biosafety for Australia and New Zealand
Australian Chamber of Commerce and Industry
Australian Industry Group
Australian Institute of Occupational Hygienists
Australian Nuclear Science and Technology Organisation
Bureau of Steel Manufacturers of Australia
CSIRO
Institute of Chemical Engineers Australia
Institute of Environmental Science and Research, New Zealand
Ministry for Primary Industries, New Zealand
National Measurement Institute, Australia
New Zealand Institute of Architects
New Zealand Microbiological Society
Responsible Care New Zealand
Royal Australian Chemical Institute
WorkSafe New Zealand
WorkSafe Victoria

Additional Interests:

Australian Radiation Protection and Nuclear Safety Agency
Bartolo Safety Management Service
NSW Ministry of Health
SGS Australian Radiation Services

Keeping Standards up-to-date

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about joint Australian/New Zealand Standards can be found by visiting the Standards Web Shop at www.saiglobal.com or Standards New Zealand web site at www.standards.govt.nz and looking up the relevant Standard in the on-line catalogue.

For more frequent listings or notification of revisions, amendments and withdrawals, Standards Australia and Standards New Zealand offer a number of update options. For information about these services, users should contact their respective national Standards organization.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Please address your comments to the Chief Executive of Standards Australia or the New Zealand Standards Executive at the address shown on the back cover.

This Standard was issued in draft form for comment as DR2 AS/NZS 2243.4:2017.

Australian/New Zealand Standard™

Safety in laboratories

Part 4: Ionizing radiations

Originated as AS 2243.4—1979.
Previous edition AS 2243.4—1998.
Revised and redesignated as AS/NZS 2243.4:2018.

COPYRIGHT

© Standards Australia Limited

© The Crown in right of New Zealand, administered by the New Zealand Standards Executive

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher, unless otherwise permitted under the Copyright Act 1968 (Australia) or the Copyright Act 1994 (New Zealand).

Jointly published by SAI Global Limited under licence from Standards Australia Limited, GPO Box 476, Sydney, NSW 2001 and by Standards New Zealand, PO Box 1473, Wellington 6140.

PREFACE

This Standard was prepared by the Joint Standards Australia/Standards New Zealand Committee CH-026, Safety in Laboratories, to supersede AS 2243.4—1998.

The objective of this Standard is to promote safe work practices and to prevent unnecessary exposure of persons working in laboratories containing sources of ionizing radiation.

This Standard does not include non-ionizing radiation, which is covered in AS 2243.5, *Safety in laboratories, Part 5: Non-ionizing radiations—Electromagnetic, sound and ultrasound*.

This Standard was prepared after consideration of the Australian Radiation Protection and Nuclear Safety Agency (ARPANSA), *Fundamentals for Protection Against Ionising Radiation* RPS F1 (2014) and *Code for Radiation Protection in Planned Exposure Situations* RPS C1 (2016), the most recent recommendations of the International Commission on Radiological Protection (ICRP), and the International Atomic Energy Agency (IAEA). Where necessary, future relevant publications of ARPANSA, the ICRP and the IAEA should be consulted as they become available.

The sections of this Standard dealing with dose limits, accident and emergency procedures and the requirements for radiological laboratories have been further refined since the 1998 edition. For more detailed information, the referenced source documents should be consulted.

This Standard is Part 4 in the AS/NZS 2243 series. The series is as follows:

AS

- 2243 Safety in laboratories
- 2243.1 Part 1: Planning and operational aspects
- 2243.2 Part 2: Chemical aspects
- 2243.3 Part 3: Microbiological safety and containment
- 2243.4 Part 4: Ionizing radiations (this Standard)
- 2243.5 Part 5: Non-ionizing radiations—Electromagnetic, sound and ultrasound
- 2243.6 Part 6: Plant and equipment aspects
- 2243.8 Part 8: Fume cupboards
- 2243.9 Part 9: Recirculating fume cabinets
- 2243.10 Part 10: Storage of chemicals

Statements expressed in mandatory terms in notes to tables are deemed to be requirements of this Standard.

The terms ‘normative’ and ‘informative’ have been used in this Standard to define the application of the appendix to which they apply. A ‘normative’ appendix is an integral part of a Standard, whereas an ‘informative’ appendix is only for information and guidance.

CONTENTS

	<i>Page</i>
FOREWORD.....	5
SECTION 1 SCOPE AND GENERAL	
1.1 SCOPE.....	6
1.2 APPLICATION	6
1.3 REFERENCED DOCUMENTS.....	6
1.4 DEFINITIONS.....	6
SECTION 2 HAZARDS AND THEIR CONTROL	
2.1 HAZARDS FROM IONIZING RADIATION	14
2.2 LEGISLATION	14
2.3 RISK ASSESSMENT	15
SECTION 3 DESIGN REQUIREMENTS FOR RADIATION, RADIOISOTOPE AND RADIOLOGICAL LABORATORIES	
3.1 GENERAL.....	16
3.2 PLANNING BRIEF	16
3.3 TYPES OF HAZARDS.....	17
3.4 DESIGN AND CONSTRUCTION REQUIREMENTS	17
3.5 RADIOISOTOPE LABORATORY GRADING	18
SECTION 4 IONIZING RADIATION PROTECTION	
4.1 GENERAL.....	24
4.2 DESIGNATED RADIATION AREA (DRA)	25
4.3 RADIATION WARNING SIGNS	25
4.4 LOCAL INSTRUCTIONS.....	26
4.5 SHIELDING	26
4.6 WORKING IN ISOLATION	26
4.7 SEALED SOURCES AND RADIATION-PRODUCING APPARATUS	27
4.8 UNSEALED RADIOACTIVE MATERIALS.....	31
SECTION 5 STORAGE AND TRANSPORT	
5.1 STORAGE OF RADIOACTIVE MATERIALS	36
5.2 TRANSPORT OF RADIOACTIVE MATERIALS.....	37
SECTION 6 MONITORING IONIZING RADIATION	
6.1 GENERAL.....	38
6.2 MEMBERS OF THE PUBLIC.....	39
6.3 CALIBRATION OF MONITORING INSTRUMENTS	39
6.4 PERSONAL MONITORING.....	39
6.5 RADIATION DOSE RECORD	40
6.6 AREA MONITORING	41
SECTION 7 RADIATION INCIDENTS	
7.1 GENERAL.....	42
7.2 RESPONSE PROCEDURES	43

SECTION 8 MANAGEMENT OF RADIOACTIVE WASTES

8.1	GENERAL.....	47
8.2	SOLID WASTE.....	47
8.3	LIQUID WASTE.....	47
8.4	AIRBORNE WASTE.....	49
8.5	OPTIONS FOR THE DISPOSAL OF SURPLUS SEALED SOURCES.....	49

APPENDICES

A	DOSE LIMITS.....	51
B	DERIVED RADIATION PROTECTION QUANTITIES.....	55
C	ROLES AND RESPONSIBILITIES.....	60
D	MEASUREMENT OF IONIZING RADIATIONS.....	66
E	IONIZING RADIATION SYMBOLS.....	73
F	MANAGEMENT OF RADIOACTIVE WASTE.....	75
G	REFERENCED AND RELATED DOCUMENTS.....	80

FOREWORD

Radiation is the emission of energy, in the form of rays, wave motion or particles, from a source. The various kinds of radiation can be subdivided broadly as follows:

- (a) *Ionizing radiations*—radiations that by have the potential to cause ionization when interacting with matter. This involves removing or adding electrons from or to the atoms of materials through which the radiations pass, including those that comprise the human organism. Sufficiently energetic radiation may also cause changes in the nuclei of the atoms in the material. Ionizing radiations comprises X-rays, gamma rays, alpha particles, beta particles (i.e. high energy electrons), neutrons, protons and other nuclear particles. Ultraviolet radiation of wavelength less than 100 nm will also cause ionization in air.
- (b) *Non-ionizing radiations*—radiations that do not cause ionization or nuclear changes, but can harm the human body in other ways. Potentially harmful non-ionizing radiations are propagated as electromagnetic waves (in particular, when the wavelength is between about 100 nm and 1 km), and acoustic noise (i.e. unwanted sound).

This Standard does not deal with non-ionizing radiation, which is covered in AS/NZS 2243.5, *Safety in laboratories, Part 5: Non-ionizing radiations*.

NOTES:

- 1 Although referred to in Item (a), ultraviolet radiation is generally classified as non-ionizing radiation and is therefore considered in AS 2243.5.
- 2 Radiation protection is concerned with the protection of individuals, their progeny, the human race as a whole and the environment, while still allowing necessary activities from which radiation exposure may result.
- 3 Various forms of high voltage apparatus may produce ionizing radiations.
- 4 Various chemicals and minerals may be radioactive by virtue of the fact that they contain traces of naturally occurring radioactive elements.

STANDARDS AUSTRALIA/STANDARDS NEW ZEALAND

Australian/New Zealand Standard
Safety in laboratories**Part 4: Ionizing radiations**

SECTION 1 SCOPE AND GENERAL

1.1 SCOPE

This Standard sets out requirements for the protection of people, non-human biota and the environment from the harmful effects of ionizing radiation used in, or as a result of any use within, or in connection with a laboratory.

This Standard also sets out the important characteristics of ionizing radiation-producing materials and apparatus, the nature of the hazards, laboratory design requirements, and other essential radiation protection information.

This Standard applies to the protection against ionizing radiation only, and does not specifically address the control of non-radiological aspects of health, safety and the environment, which are addressed in the relevant Parts of the AS/NZS 2243 series, even though these aspects also need to be considered.

NOTE: The protection from harmful effects of non-ionizing radiation is covered in AS/NZS 2243.5.

This Standard does not apply to exposures incurred during the course of medical or dental diagnosis or treatment and exposure to volunteers as part of a program of biomedical research.

NOTE: See Appendix G, Paragraph G3, [4] and [6].

Where requirements given in this Standard differ from requirements under jurisdictional legislation, the legislation takes precedence.

NOTE: For certain facilities, operations and activities (for example, nuclear installations, transport of radioactive material and radioactive waste management) other requirements apply.

1.2 APPLICATION

This Standard shall be implemented in all laboratories in which—

- (a) radioactive materials are used, processed or stored; or
- (b) ionizing radiation apparatus is used.

1.3 REFERENCED DOCUMENTS

A list of referenced documents and other related publications is provided in Appendix G.

1.4 DEFINITIONS

For the purpose of this Standard, the definitions below apply.

1.4.1 Absorbed dose

The energy absorbed by matter from ionizing radiation per unit mass of irradiated material.

NOTE: The SI unit of absorbed dose is the joule per kilogram (J/kg), termed the gray (Gy). For radiation protection purposes, the absorbed dose is averaged over a tissue or organ.