

Australian Standard™

**Quality systems—Medical devices—
Particular requirements for the
application of ISO 9003**

This Australian Standard was prepared by Committee QR-008, Quality Systems. It was approved on behalf of the Council of Standards Australia on 26 June 2002 and published on 28 June 2002.

The following are represented on Committee QR-008:

Australian Chamber of Commerce and Industry
Australian Electrical and Electronic Manufacturers Association
Australian Industry Group
Australian Information Industry Association
Australian Institute of Petroleum
Australian Organisation for Quality
Boiler and Pressure Vessel Manufacturers Association of Australia
Bureau of Steel Manufacturers of Australia
Certification Bodies (Australia)
Commonwealth Department of Transport and Regional Services
Department of Agriculture, Fisheries and Forestry (Commonwealth)
Department of Defence (Australia)
Department of Industry Science and Resources (Commonwealth)
Electricity Supply Association of Australia
Federal Chamber of Automotive Industries
Institute of Materials Engineering Australasia
Institution of Engineers Australia
Main Roads Department
Queensland Master Builders Australia
Plastics and Chemicals Industries Association Incorporated
Quality Society of Australasia
The Royal Australian Chemical Institute

Keeping Standards up-to-date

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about Standards can be found by visiting the Standards Australia web site at www.standards.com.au and looking up the relevant Standard in the on-line catalogue.

Alternatively, the printed Catalogue provides information current at 1 January each year, and the monthly magazine, *The Australian Standard*, has a full listing of revisions and amendments published each month.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at mail@standards.com.au, or write to the Chief Executive, Standards Australia International Ltd, GPO Box 5420, Sydney, NSW 2001.

This Standard was issued in draft form for comment as DR 02211.

Australian Standard™

**Quality systems— Medical devices—
Particular requirements for the
application of ISO 9003**

First published as AS EN 46003—2002.

COPYRIGHT

© Standards Australia International

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia International Ltd
GPO Box 5420, Sydney, NSW 2001, Australia

ISBN 0 7337 4703 5

PREFACE

This Standard has been developed to assist in the process of implementation of the Australian Medical Device legislation.

After consultation with stakeholders in both countries, Standards Australia and Standards New Zealand decided to develop this Standard as an Australian, rather than an Australian/New Zealand Standard, through the Joint Standards Australia/Standards New Zealand Committee QR-008 on Quality Systems.

This Standard is identical with and has been reproduced from EN 46003:1999, *Quality systems - Medical devices — Particular requirements for the application of EN ISO 9003*.

The objective of this Standard is to specify, for suppliers of medical devices, quality systems requirements that are more specific than those given in ISO 9003.

Users in Australia should be aware that, where reference is made to EN 29001, EN 29002 and/or EN 29003, they are identical with the 1994 editions of AS/NZS ISO 9001, AS/NZS ISO 9002 and AS/NZS ISO 9003 respectively. These Standards provide three quality assurance models that represent three distinct forms of quality system requirements suitable for the purpose of a supplier demonstrating its capability, and for the assessment of the capability of a supplier by external parties.

At the time of publication, the 1994 editions of AS/NZS ISO 9001, AS/NZS ISO 9002 and AS/NZS ISO 9003 have been superseded by AS/NZS ISO 9001:2000, *Quality management systems — Requirements*, but will remain available as superseded standards until December 2003. The use of the superseded standards and their EN equivalents beyond that date is endorsed for applications covered by the Australian Medical Device legislation.

As this Standard is reproduced from a European Standard, the following applies:

- (a) Its number does not appear on each page of text and its identity is shown only on the cover and title page.
- (b) In the source text 'this European Standard' should read 'this Australian Standard'.
- (c) A full point substitutes for a comma when referring to a decimal marker.

CONTENTS

Introduction	iv
1 Scope	1
2 Normative references	1
3 Definitions	1
4 Quality system requirements	2
4.1 Management responsibility	2
4.2 Quality system	2
4.3 Contract review	2
4.4 Design control	3
4.5 Document and data control	3
4.6 Purchasing	3
4.7 Control of customer-supplied product	3
4.8 Product identification and traceability	3
4.9 Process control	3
4.10 Inspection and testing	3
4.11 Control of inspection, measuring and test equipment	3
4.12 Inspection and test status	4
4.13 Control of nonconforming product	4
4.14 Corrective action	4
4.15 Handling, storage, packaging, preservation and delivery	4
4.16 Control of quality records	5
4.17 Internal quality audits	5
4.18 Training	5
4.19 Servicing	5
4.20 Statistical techniques	5
Annex A (informative) Bibliography	6
Annex ZA (informative) Clauses of this European Standard addressing essential requirements or other provisions of EU Directives	7

INTRODUCTION

EN ISO 9003 : 1994 is intended to be a general standard defining quality system requirements. EN 46003 provides particular requirements for **suppliers of medical devices** that are more specific than the general requirements specified in EN ISO 9003 : 1994.

In conjunction with EN ISO 9003 : 1994, this European Standard defines requirements for quality systems relating to final inspection and test of **medical devices**. It can only be used in combination with EN ISO 9003 : 1994 and is not a "stand alone" standard.

There is a wide variety of **medical devices** and some of the particular requirements of this standard only apply to named groups of **medical devices**. These groups are described in clause 3, Definitions.

Particular requirements in a number of clauses of this standard are covered in detail in other European Standards. **Suppliers** should review the requirements and consider using national standards implementing harmonized European Standards in these areas.

AUSTRALIAN STANDARD

Quality systems—Medical devices—Particular requirements for the application of ISO 9003

1 Scope

This European Standard specifies, in conjunction with EN ISO 9003 : 1994, the quality system requirements for the final inspection and test of **medical devices** excluding in vitro diagnostic medical devices and active implantable medical devices, and is applicable when a **medical device supplier's** quality system is assessed in accordance with regulatory requirements.

NOTE: For sterile medical devices the relevant particular clauses in EN 46002 : 1996 apply as this standard alone is not sufficient for manufacturers of sterile medical devices seeking to comply with regulatory requirements.

As part of an assessment by a third party for the purpose of regulatory requirements, the **supplier** may be required to provide access to confidential data in order to demonstrate compliance with this standard. The **supplier** may be required to exhibit these data but is not obliged to provide copies for retention.

2 Normative references

This European Standard incorporates, by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text and the publications are listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to this European Standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies.

EN ISO 8402 : 1995

Quality management and quality assurance – Vocabulary (ISO 8402 : 1994)

EN ISO 9003 : 1994

Quality systems – Model for quality assurance in final inspection and test (ISO 9003:1994)

3 Definitions

For the purposes of this standard, the following definitions apply.

3.1 Medical device: Any instrument, apparatus, appliance, material or other article, whether used alone or in combination, including the software necessary for its proper application, intended by the manufacturer to be used for human beings for the purpose of:

- diagnosis, prevention, monitoring, treatment or alleviation of disease;
- diagnosis, monitoring, treatment, alleviation of or compensation for an injury or handicap;
- investigation, replacement or modification of the anatomy or of a physiological process;
- control of conception;

and which does not achieve its principal intended action in or on the human body by pharmacological, immunological or metabolic means, but which may be assisted in its function by such means.

NOTE: Definition 3.1 reproduces the definition given in the Council Directive 93/42/EEC of 14 June 1993 concerning medical devices.

3.2 Implantable medical device: Any device which is intended:

- to be totally introduced into the human body or;
- to replace an epithelial surface or the surface of the eye

by surgical intervention which is intended to remain in place after the procedure.

Any device intended to be partially introduced into the human body through surgical intervention and intended to remain in place after the procedure for at least 30 days is also considered as implantable device.