

AS 2495

/H
Hypoc
Syrir

WITHDRAWN: 19920900

SUPERSEDED BY:
AS 1077-1992

~~REFERENCE COPY
INFORMATION CENTRE
STANDARDS AUSTRALIA~~

AS 2495—1981
UDC 615.473.3

Australian Standard 2495—1981

SINGLE-USE 0.5 mL SYRINGES (STERILE) FOR THE INJECTION OF 100 UNITS PER MILLILITRE INSULIN (U-100), SUITABLE FOR PAEDIATRIC USE

[Title allocated by the Defence Cataloguing Authority:
SYRINGE AND NEEDLE, HYPODERMIC (SINGLE-USE, 0.5 mL,
For 100 Units Per Millilitre Injections, Suitable for Paediatric Use)]



STANDARDS ASSOCIATION OF AUSTRALIA
Incorporated by Royal Charter



THE FOLLOWING SCIENTIFIC, INDUSTRIAL AND GOVERNMENTAL ORGANIZATIONS and departments were officially represented on the committee entrusted with the preparation of this standard:

Australian Chamber of Commerce
Australian Dental Standards Laboratory
Australian Diabetes Society
Australian Medical Association
Commonwealth and State Departments of Health
Confederation of Australian Industry
Diabetes Federation of Australia
Hospitals and Hospital Associations
New South Wales Government Stores Department
The Pharmacy Guild of Australia

This standard, prepared by Committee MD/2, Hypodermic Equipment for Insulin Injection, was approved on behalf of the Council of the Standards Association of Australia on 11 August 1981, and was published on 12 October 1981.

To keep abreast of progress in industry, Australian standards are subject to continuous review and are kept up-to-date by the issue of amendments or new editions as necessary. It is important therefore that standards users ensure that their standards are up-to-date. Full details of all SAA publications will be found in the Annual List of Australian Standards; these details are supplemented by listings in the SAA monthly journal 'The Australian Standard'. Information on the Annual List and 'The Australian Standard' may be obtained from any sales office of the Association, where details are also available of the current status of individual standards. Suggestions for improvements to published standards, addressed to the head office of the Association, are welcomed.

This standard was issued in draft form for public review as DR 80223.

AUSTRALIAN STANDARD

**SINGLE-USE 0.5 mL SYRINGES
(STERILE) FOR THE INJECTION
OF 100 UNITS PER MILLILITRE
INSULIN (U-100), SUITABLE
FOR PAEDIATRIC USE**

AS 2495—1981

First published1981

**PUBLISHED BY THE STANDARDS ASSOCIATION OF AUSTRALIA
STANDARDS HOUSE, 80 ARTHUR ST, NORTH SYDNEY, N.S.W.**

ISBN 0 7262 2369 7

22 OCT 1981



PREFACE

This standard was prepared by the Association's Committee on Hypodermic Equipment for Insulin Injection following the availability of U-100 strength insulin in Australia. The standard has been prepared in a manner intended to make it suitable for adoption as a statutory standard under the Federal Therapeutic Goods Act.

Following a submission by the College of Paediatricians who expressed the need for a smaller and more accurately marked insulin syringe than that provided by the 1 mL syringe which is marked in 2-unit intervals, this standard was prepared to provide for a lower volume syringe marked in 1-unit intervals and hence be suitable for paediatric use.

This standard makes reference to the following Australian standards:

AS 1077	Single-use 1 mL Syringes (Sterile) for the Injection of 100 Units per Millilitre Insulin (U-100)
AS 1386	Cleanrooms and Work-stations
AS 1600	Conical Fittings with 6 percent (Luer) Taper for Hypodermic and Other Surgical Equipment*
AS 1615	Single-use Needles (Sterile) for Insulin Injection*
AS 2070	Plastics Materials for Food Contact Use
AS 2134	Code of Practice for the Chemical Analysis of Materials by Flame Atomic Absorption Spectroscopy
AS 2193	Methods for Calibration and Grading of Force-measuring Systems of Testing Machines
AS K185	Colours for Specific Purposes

*In course of revision.

© Copyright — STANDARDS ASSOCIATION OF AUSTRALIA 1981

Users of standards are reminded that copyright subsists in all SAA publications. No part of this publication may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing of the Standards Association of Australia.

CONTENTS

	<i>Page</i>
SPECIFICATION	
1 Scope	4
2 Description of Syringe	4
3 Definitions	4
4 Dimensions of Syringe	4
5 Requirements for Barrel	4
6 Requirements for Plunger	7
7 Fit of Plunger in Barrel	7
8 Resistance to Heat	7
9 Requirements for Needle	7
10 Connection between Needle and Barrel	7
11 Dead Space	7
12 Particulate Contamination	7
13 Extractable Metals	7
14 Pyrogens	7
15 Toxicity	7
16 Sterilization	7
17 Sterility	8
18 Packaging	8
19 Labelling	8
APPENDICES	
A Method for Testing Durability of Scale	10
B Method for Determining Precision of Capacity Line and Dead Space	11
C Method for Testing Fit of Plunger in Barrel Under Pressure	12
D Method for Testing Fit of Plunger in Barrel Under Suction	13
E Method for Testing Movement of Plunger in Barrel	14
F Method for Testing Watertightness of Connection between Needle Hub and Nozzle	15
G Methods for Testing Strength of Joint between Needle and Barrel	16
H Method of Testing Syringes for Particulate Contamination	17
J Method for Determining Extractable Metals in Syringes	19
K Method of Testing Syringes for Pyrogens	20
L Method of Testing Syringes for Toxicity	22
M Method of Testing Syringes for Sterility....	23
N Method of Testing Unit Packs for Leakage	26
O Assessment of Compliance	27

STANDARDS ASSOCIATION OF AUSTRALIA

Australian Standard

for

SINGLE-USE 0.5 mL SYRINGES (STERILE) FOR THE INJECTION OF 100 UNITS PER MILLILITRE INSULIN (U-100), SUITABLE FOR PAEDIATRIC USE

1 SCOPE. This standard specifies requirements for sterilized 0.5 mL syringes individually packed with needles and intended for use once only, solely for the injection of 100 units per millilitre insulin (U-100). The syringes are graduated to 50 'units' in 1-unit intervals, and are suitable for paediatric use.

The needles are either attached or fixed on the syringes.

Syringes taking detachable needles have the 6 per cent (Luer) fitting.

NOTES:

1. Sterile hypodermic syringes specified herein are intended for use immediately after filling and are not intended for containing insulin over extended periods.
2. Advisory information on sampling for assessing compliance with this standard is given in Appendix O.

2 DESCRIPTION OF SYRINGE.

NOTE: The terms in italic type are illustrated in Fig. 1.

The components of the syringe are the *plunger* and the *barrel*, and the needle. The term *proximal* refers to the end of the plunger which protrudes from the barrel, and to the corresponding end of the barrel and syringe. The term *distal* refers to the end of the plunger, barrel or syringe opposite to the proximal end. The plunger is *fully inserted* when seated firmly but not under pressure against the distal end of the barrel. Outward projections at the proximal end of the barrel are termed the *flange*. The space enclosed by the inside walls of the barrel is termed the *lumen*.

Syringes taking a detachable needle have a *nozzle* for attaching the needle onto the distal end of the barrel. *Graduation lines* and numbering, which comprise the *scale*, are marked on the barrel. The graduation line nearest the distal end of the barrel is termed the *zero line*. The graduation line nearest the proximal end is termed the *capacity line*. The proximal end of the plunger is formed into a *push-button*. The distal end of the plunger (which is usually a resilient plug or ring, termed the *seal*) forms a zone of contact with the inside wall of the barrel. The distal edge of this zone of contact is termed the *fiducial line* and provides the reference position for setting the plunger against the scale.

3 DEFINITIONS. For the purpose of this standard, the following definitions apply:

3.1 Syringe—a syringe with detachable needle attached and a syringe with fixed needle, the plunger being inserted in the barrel of the syringe.

3.2 Needle—a detachable needle or a fixed needle. A detachable needle may include a design so as to minimize dead space.

3.3 Fixed (in relation to a needle)—a needle not designed to be detached from a syringe.

3.4 Detachable (in relation to a needle)—a needle not bonded or permanently fixed to the syringe.

3.5 Unit (in relation to packing)—a syringe with a needle complying with Clause 18.1.

3.6 Unit pack—a pack containing a single unit and providing the microbiological barrier.

3.7 Multiple pack—a pack, not being a store pack, containing two or more unit packs.

3.8 Store pack—a pack containing one or more multiple packs.

3.9 Unit (in relation to insulin)—the International Unit of Insulin.

3.10 Water—purified water of the British Pharmacopoeia.

4 DIMENSIONS OF SYRINGE.**4.1 Length.**

4.1.1 Syringe with detachable needle. If the syringe takes a detachable needle, the length of the syringe with the plunger fully inserted and any detachable needle removed (dimension A in Fig. 2) shall not exceed 100 mm.

4.1.2 Syringe with fixed needle. If the syringe has a fixed needle, the length of the syringe with the plunger fully inserted and the exposed length of the needle tube included (dimension F in Fig. 2) shall not exceed 120 mm.

4.2 Length of Projection of Plunger from Barrel. The length of the projection of the fully inserted plunger from the barrel of the syringe (dimension B in Fig. 2) shall not be less than 10 mm.

5 REQUIREMENTS FOR BARREL.

5.1 Composition. The barrel of the syringe shall be made of materials which—

- (a) are transparent or sufficiently translucent to afford good visibility of the fiducial line and any air/liquid interface in the barrel;
- (b) do not affect the therapeutic efficacy of insulin when the syringe is used to inject a preparation of insulin; and
- (c) if of plastics material comply with the appropriate Part(s) of AS 2070 for plastics materials for food contact use. The use of any plastics material not specified by the appropriate parts of AS 2070 for food contact use shall, in the first instance, be referred to the appropriate Regulatory Authority for evaluation. In addition, the plastics materials shall be able to withstand the sterilization procedure as stated on the pack by the manufacturer (see Clause 19.1).