

AS 1289.4.3.1:2021



Methods of testing soils for engineering purposes

Method 4.3.1: Soil chemical tests — Determination of the pH value of a soil — Electrometric method



AS 1289.4.3.1:2021

This Australian Standard® was prepared by CE-009, Testing Of Soils For Engineering Purposes. It was approved on behalf of the Council of Standards Australia on 25 March 2021.

This Standard was published on 7 May 2021.

The following are represented on Committee CE-009:

- Australian Building Codes Board
- Australian Chamber of Commerce and Industry
- Austroads
- Cement Concrete and Aggregates Australia — Aggregates
- Engineering and Construction Laboratories Association
- Engineers Australia/Australian Geomechanics Society
- National Association of Testing Authorities Australia
- University of Melbourne
- University of Sydney
- Victorian Construction Materials Laboratories Association

This Standard was issued in draft form for comment as DR AS 1289.4.3.1:2021.

Keeping Standards up-to-date

Ensure you have the latest versions of our publications and keep up-to-date about Amendments, Rulings, Withdrawals, and new projects by visiting:

www.standards.org.au

ISBN 978 1 76113 301 5

Methods of testing soils for engineering purposes

Method 4.3.1: Soil chemical tests — Determination of the pH value of a soil — Electrometric method

Originated as part of AS A89—1966.
Revised and redesignated AS 1289.D3.1—1977.
Revised and redesignated AS 1289.4.3.1—1997.
Second edition AS 1289.4.3.1:2021.

© Standards Australia Limited 2021

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher, unless otherwise permitted under the Copyright Act 1968 (Cth).

Preface

This Standard was prepared by the Standards Australia committee CE-009, Testing of Soils for Engineering Purposes, to supersede AS 1289.4.3.1:1997.

The objective of this document is to specify the electrometric determination of the pH value of a soil-suspension.

The major changes in this edition are as follows:

- (a) The accuracy requirements of the balance have decreased.
- (b) Removing the option of using deionised water instead of distilled water to make the soil-suspension.
- (c) Mixing and settling are better defined and measurement in the suspended soil is mandated.

Contents

Preface	ii
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Apparatus	1
5 Reagents	2
6 Procedure	2
7 Test report	3

NOTES

Australian Standard®

Methods of testing soils for engineering purposes

Method 4.3.1: Soil chemical tests — Determination of the pH value of a soil — Electrometric method

1 Scope

This Standard covers the electrometric determination of the pH value of a soil-suspension.

NOTE The pH value of a sample of groundwater can be measured in a similar manner to that used for a soil-suspension.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document.

NOTE Documents referenced for informative purposes are listed in the Bibliography.

AS 1289.1.1, *Methods of testing soils for engineering purposes, Method 1.1: Preparation of disturbed soil samples for testing*

ISO 3310-2, *Test sieves — Technical requirements and testing — Part 2: Test sieves of perforated metal plate*

3 Terms and definitions

No terms and definitions are listed in this document.

4 Apparatus

The following apparatus are required:

- (a) pH meter, covering at least the range pH 3.0 to pH 13.0, fitted with a scale that is readable and accurate to 0.05 pH units.

NOTE 1 Soils with high pH values, such as stabilized soils and naturally occurring soils with a pH in excess of 10, should use an electrode reading to pH 13.0 as other electrodes would be adversely affected.

NOTE 2 The pH meter manufacturer's instructions should be followed.

- (b) Balance of at least 100 g capacity, with a limit of performance not greater than ± 0.05 g.
- (c) 100 mL glass beakers, with cover glasses and stirring rods.
- (d) Wash-bottle, preferably plastic, containing distilled water.
- (e) 2.36 mm sieve, conforming to ISO 3310-2.
- (f) Thermometer, readable and accurate to 1 °C

NOTE 3 This thermometer may be incorporated into the pH meter.