

AS 3778.6.1:2022
ISO 2537:2007



Measurement of water flow in open channels

Part 6.1: Measuring devices, instruments and equipment — Rotating-element current meters (ISO 2537:2007, IDT)



AS 3778.6.1:2022

This Australian Standard ® was prepared by CE-024, Measurement Of Water Flow In Open Channels and Closed Conduits. It was approved on behalf of the Council of Standards Australia on 16 May 2022.

This Standard was published on 10 June 2022.

The following are represented on Committee CE-024:

- Australian Bureau of Meteorology
- Australian Hydrographers Association
- Australian Industry Group
- Department of Planning, Industry and Environment, NSW
- Engineers Australia
- Institute of Instrumentation, Control & Automation Australia
- Irrigation Australia
- Joint Accreditation System of Australia & New Zealand
- National Measurement Institute
- Water NSW

This Standard was issued in draft form for comment as DR AS 3778.6.1:2022.

Keeping Standards up-to-date

Ensure you have the latest versions of our publications and keep up-to-date about Amendments, Rulings, Withdrawals, and new projects by visiting:

www.standards.org.au

ISBN 978 1 76113 807 2

Measurement of water flow in open channels

Part 6.1: Measuring devices, instruments and equipment — Rotating-element current meters (ISO 2537:2007, IDT)

Originated as AS 3778.6.1—1992.
Second edition 2022.

COPYRIGHT

© ISO 2022 — All rights reserved
© Standards Australia Limited 2022

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher, unless otherwise permitted under the Copyright Act 1968 (Cth).

Preface

This Standard was prepared by the Standards Australia Committee CE-024, Measurement of water flow in open channels and closed conduits, to supersede AS 3778.6.1—1992, *Measurement of water flow in open channels, Part 6.1: Measuring devices, instruments and equipment — Rotating-element current meters*.

The objective of this document is to specify the operational requirements, construction, calibration, and maintenance of rotating-element devices for the measurement of flow velocities in open channels.

This document is identical with, and has been reproduced from, ISO 2537:2007, *Hydrometry — Rotating-element current-meters*.

As this document has been reproduced from an International document, the following apply:

- (a) In the source text “this International Standard” should read “this document”.
- (b) A full point substitutes for a comma when referring to a decimal marker.

Australian or Australian/New Zealand Standards that are identical adoptions of international normative references may be used interchangeably. Refer to the online catalogue for information on specific Standards.

The terms “normative” and “informative” are used in Standards to define the application of the appendices or annexes to which they apply. A “normative” appendix or annex is an integral part of a Standard, whereas an “informative” appendix or annex is only for information and guidance.

Contents

Preface	ii
Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle of operation	1
4.1 Proportionality	1
4.2 Flow velocity	1
5 Types of current-meters	2
5.1 General	2
5.2 Cup-type current-meter	2
5.3 Propeller-type current-meter	2
6 Operational requirements	2
6.1 Positioning	2
6.2 Resistance to flow	2
6.3 Limits of use	2
7 Characteristics of cup-type and propeller-type current-meter	3
7.1 General	3
7.2 Cup-type current-meter	3
7.3 Propeller-type current-meter	3
8 Construction features	3
9 Calibration (rating)	4
9.1 General	4
9.2 Conformity of calibration	4
9.3 Minimum speed of response	4
9.4 Calibration table	4
9.5 Types of calibration	5
9.5.1 General	5
9.5.2 Individual calibration	5
9.5.3 Group (standard) calibration	5
9.6 Recalibration	5
9.7 Type of suspension	5
9.8 Uncertainty in calibration	6
10 Maintenance of current-meter	6
11 Operational and servicing manual	6
Annex A (normative) Construction features of current-meters	7
Annex B (normative) Maintenance of current-meter	11
Bibliography	13

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 2537 was prepared by Technical Committee ISO/TC 113, *Hydrometry*, Subcommittee SC 5, *Instruments, equipment and data management*.

This fourth edition cancels and replaces the third edition (ISO 2537:1988), which has been technically revised.

Australian Standard®

Measurement of water flow in open channels

Part 6.1: Measuring devices, instruments and equipment — Rotating-element current meters (ISO 2537:2007, IDT)

1 Scope

This International Standard specifies the operational requirements, construction, calibration, and maintenance of rotating-element devices for the measurement of flow velocities in open channels.

ISO 748 gives information on the use of these devices.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 772, *Hydrometric determinations — Vocabulary and symbols*

ISO 3454, *Hydrometry — Direct depth sounding and suspension equipment*

ISO 3455, *Hydrometry — Calibration of current-meters in straight open tanks*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 772 and the following apply.

3.1

propeller pitch

distance the propeller current-meter relatively moves through the water during one revolution

4 Principle of operation

4.1 Proportionality

The rotating element of a current-meter is driven by the fluid at an angular velocity that is proportional to the local velocity of the fluid at the point of immersion when that velocity exceeds a critical value.

4.2 Flow velocity

In order to determine the velocity of the fluid, a current-meter is placed at a point in a stream and the number of revolutions of the rotor during a specified time interval is counted or the time required by the rotor to turn a given number of revolutions is observed. The velocity is obtained from the current-meter calibration table or calibration equation(s), established experimentally during its calibration ([Clause 9](#)). The number of current-meter revolutions (rotations) may be determined by sensing the signals emitted (such as electrical pulses) through the rotation of the rotor by using a suitable counting device. The velocity can be determined from a direct reading of the speed of rotation of the rotating element by means of equipment designed for this purpose.